

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA F\Data\BF110817\
 Data File : BF100291.D
 Acq On : 8 Nov 2017 14:57
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 08 17:11:40 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 08 14:33:31 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	93	0.00
2	1,4-Dioxane	40.000	39.331	1.7	91	0.00
3	Pyridine	40.000	39.517	1.2	89	0.00
4	n-Nitrosodimethylamine	40.000	41.188	-3.0	93	0.00
5 S	2-Fluorophenol	80.000	79.738	0.3	94	0.00
6	Aniline	40.000	40.471	-1.2	93	0.00
7 S	Phenol-d6	80.000	80.373	-0.5	94	0.00
8	2-Chlorophenol	40.000	40.561	-1.4	95	0.00
9	Benzaldehyde	40.000	39.197	2.0	92	0.00
10 C	Phenol	40.000	39.476	1.3	93	0.00
11	bis(2-Chloroethyl)ether	40.000	40.305	-0.8	94	0.00
12	1,3-Dichlorobenzene	40.000	40.423	-1.1	94	0.00
13 C	1,4-Dichlorobenzene	40.000	39.968	0.1	94	0.00
14	1,2-Dichlorobenzene	40.000	40.659	-1.6	95	0.00
15	Benzyl Alcohol	40.000	41.840	-4.6	95	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	40.415	-1.0	94	0.00
17	2-Methylphenol	40.000	41.462	-3.7	96	0.00
18	Hexachloroethane	40.000	41.156	-2.9	94	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.308	-0.8	95	0.00
20	3+4-Methylphenols	40.000	41.066	-2.7	95	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	94	0.00
22	Acetophenone	40.000	38.945	2.6	94	0.00
23 S	Nitrobenzene-d5	80.000	87.325	-9.2	99	0.00
24	Nitrobenzene	40.000	44.644	-11.6	98	0.00
25	Isophorone	40.000	40.467	-1.2	95	0.00
26 C	2-Nitrophenol	40.000	43.774	-9.4	104	0.00
27	2,4-Dimethylphenol	40.000	41.401	-3.5	96	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.931	0.2	95	0.00
29 C	2,4-Dichlorophenol	40.000	41.695	-4.2	95	0.00
30	1,2,4-Trichlorobenzene	40.000	40.468	-1.2	96	0.00
31	Naphthalene	40.000	40.143	-0.4	96	0.00
32	Benzoic acid	40.000	42.822	-7.1	110	0.00
33	4-Chloroaniline	40.000	40.739	-1.8	95	0.00
34 C	Hexachlorobutadiene	40.000	40.779	-1.9	96	0.00
35	Caprolactam	40.000	42.052	-5.1	99	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.727	-4.3	96	0.00
37	2-Methylnaphthalene	40.000	40.025	-0.1	96	0.00
38 I	Acenaphthene-d10	20.000	20.000	0.0	95	0.00
39	1,2,4,5-Tetrachlorobenzene	40.000	40.340	-0.9	97	0.00
40 P	Hexachlorocyclopentadiene	40.000	43.224	-8.1	97	0.00
41 S	2,4,6-Tribromophenol	80.000	85.553	-6.9	99	0.00
42 C	2,4,6-Trichlorophenol	40.000	42.455	-6.1	97	0.00
43	2,4,5-Trichlorophenol	40.000	42.156	-5.4	98	0.00
44 S	2-Fluorobiphenyl	80.000	77.639	3.0	96	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA F\Data\BF110817\
 Data File : BF100291.D
 Acq On : 8 Nov 2017 14:57
 Operator : SJ/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 08 17:11:40 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 08 14:33:31 2017
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45	1,1'-Biphenyl	40.000	39.424	1.4	97	0.00
46	2-Chloronaphthalene	40.000	39.835	0.4	96	0.00
47	2-Nitroaniline	40.000	43.513	-8.8	104	0.00
48	Acenaphthylene	40.000	40.031	-0.1	97	0.00
49	Dimethylphthalate	40.000	40.839	-2.1	100	0.00
50	2,6-Dinitrotoluene	40.000	43.369	-8.4	100	0.00
51 C	Acenaphthene	40.000	40.104	-0.3	98	0.00
52	3-Nitroaniline	40.000	43.248	-8.1	100	0.00
53 P	2,4-Dinitrophenol	40.000	45.276	-13.2	121	0.00
54	Dibenzofuran	40.000	39.932	0.2	97	0.00
55 P	4-Nitrophenol	40.000	44.464	-11.2	103	0.00
56	2,4-Dinitrotoluene	40.000	45.723	-14.3	103	0.00
57	Fluorene	40.000	40.380	-1.0	97	0.00
58	2,3,4,6-Tetrachlorophenol	40.000	41.676	-4.2	96	0.00
59	Diethylphthalate	40.000	40.508	-1.3	99	0.00
60	4-Chlorophenyl-phenylether	40.000	41.185	-3.0	99	0.00
61	4-Nitroaniline	40.000	43.655	-9.1	100	0.00
62	Azobenzene	40.000	39.786	0.5	97	0.00
63 I	Phenanthrene-d10	20.000	20.000	0.0	97	0.00
64	4,6-Dinitro-2-methylphenol	40.000	43.268	-8.2	112	0.00
65 c	n-Nitrosodiphenylamine	40.000	39.375	1.6	97	0.00
66	4-Bromophenyl-phenylether	40.000	39.991	0.0	98	0.00
67	Hexachlorobenzene	40.000	40.639	-1.6	99	0.00
68	Atrazine	40.000	41.081	-2.7	99	0.00
69 C	Pentachlorophenol	40.000	42.138	-5.3	100	0.00
70	Phenanthrene	40.000	39.202	2.0	100	0.00
71	Anthracene	40.000	39.313	1.7	99	0.00
72	Carbazole	40.000	38.955	2.6	98	0.00
73	Di-n-butylphthalate	40.000	40.365	-0.9	100	0.00
74 C	Fluoranthene	40.000	38.972	2.6	99	0.00
75 I	Chrysene-d12	20.000	20.000	0.0	97	0.00
76	Benzidine	40.000	40.258	-0.6	91	0.00
77	Pyrene	40.000	40.492	-1.2	99	0.00
78 S	Terphenyl-d14	80.000	77.496	3.1	99	0.00
79	Butylbenzylphthalate	40.000	42.089	-5.2	100	0.00
80	Benzo(a)anthracene	40.000	40.157	-0.4	98	0.00
81	3,3'-Dichlorobenzidine	40.000	41.479	-3.7	96	0.00
82	Chrysene	40.000	39.905	0.2	98	0.00
83	Bis(2-ethylhexyl)phthalate	40.000	42.729	-6.8	98	0.00
84 c	Di-n-octyl phthalate	40.000	42.421	-6.1	98	0.00
85	Indeno(1,2,3-cd)pyrene	40.000	38.190	4.5	95	0.00
86 I	Perylene-d12	20.000	20.000	0.0	96	0.00
87	Benzo(b)fluoranthene	40.000	40.278	-0.7	97	0.00

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA F\Data\BF110817\
Data File : BF100291.D
Acq On : 8 Nov 2017 14:57
Operator : SJ/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 08 17:11:40 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110817.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 08 14:33:31 2017
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(k)fluoranthene	40.000	43.191	-8.0	98	0.00
89 C	Benzo(a)pyrene	40.000	41.251	-3.1	96	0.00
90	Dibenzo(a,h)anthracene	40.000	39.578	1.1	96	0.00
91	Benzo(a,h,i)perylene	40.000	38.612	3.5	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0