

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110824\
 Data File : BF140293.D
 Acq On : 08 Nov 2024 12:47
 Operator : RC/JU
 Sample : P4720-01DL 2X
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 JC-701-COMP-01DL

Quant Time: Nov 08 13:24:37 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110524.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 05 16:47:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	132204	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	492322	20.000	ng	0.00
39) Acenaphthene-d10	9.916	164	291280	20.000	ng	0.00
64) Phenanthrene-d10	11.404	188	516504	20.000	ng	0.00
76) Chrysene-d12	14.051	240	280401	20.000	ng	-0.01
86) Perylene-d12	15.551	264	325761	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.487	112	41419	5.192	ng	0.00
7) Phenol-d6	6.492	99	55040	5.364	ng	-0.01
23) Nitrobenzene-d5	7.428	82	36577	3.694	ng	-0.02
42) 2,4,6-Tribromophenol	10.704	330	14487	5.032	ng	0.00
45) 2-Fluorobiphenyl	9.228	172	74165	4.066	ng	-0.01
79) Terphenyl-d14	12.986	244	71648	4.186	ng	-0.01
Target Compounds						
						Qvalue
75) Fluoranthene	12.621	202	51748	2.008	ng	98
78) Pyrene	12.851	202	47254	2.041	ng	99
84) Bis(2-ethylhexyl)phtha...	14.033	149	571651	48.925	ng	# 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF110824\
Data File : BF140293.D
Acq On : 08 Nov 2024 12:47
Operator : RC/JU
Sample : P4720-01DL 2X
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
JC-701-COMP-01DL

Quant Time: Nov 08 13:24:37 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110524.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 05 16:47:56 2024
Response via : Initial Calibration

