


Data Path : Z:\HPCHEM1\BNA F\DATA\BF110916\
 Data File : BF091122.D
 Acq On : 9 Nov 2016 13:30
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF110316.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 08 12:59:29 2016



AutoFind: Scans 853, 854, 855; Background Corrected with Scan 849

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.3	14490	PASS
68	69	0.00	2	1.1	168	PASS
69	198	0.00	100	45.9	15025	PASS
70	69	0.00	2	0.7	101	PASS
127	198	10	80	57.5	18826	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	32724	PASS
199	198	5	9	6.9	2248	PASS
275	198	10	60	25.0	8193	PASS
365	198	1	100	3.0	975	PASS
441	442	0.01	24	15.6	4433	PASS
442	198	50	100	86.7	28361	PASS
443	442	15	24	18.4	5215	PASS