

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF110917\
 Data File : BF100348.D
 Acq On : 9 Nov 2017 18:47
 Operator : SJ/JU
 Sample : PB103993BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB103993BL

Quant Time: Nov 10 00:51:41 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 09 16:04:38 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.90	152	157824	20.00	ng	0.00
21) Naphthalene-d8	8.18	136	633397	20.00	ng	0.00
38) Acenaphthene-d10	9.94	164	306056	20.00	ng	0.00
63) Phenanthrene-d10	11.42	188	620144	20.00	ng	0.00
75) Chrysene-d12	14.06	240	499050	20.00	ng	0.00
86) Perylene-d12	15.54	264	373247	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.53	112	1017917	103.39	ng	0.02
7) Phenol-d6	6.52	99	1245769	102.61	ng	0.00
23) Nitrobenzene-d5	7.46	82	780882	81.51	ng	0.00
41) 2,4,6-Tribromophenol	10.73	330	347782	111.51	ng	0.00
44) 2-Fluorobiphenyl	9.26	172	1490276	74.15	ng	0.00
78) Terphenyl-d14	13.02	244	1505359	69.31	ng	0.00


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SVOASRV\HPCHEM1\BNA F\DATA\BF110917\
Data File : BF100348.D
Acq On : 9 Nov 2017 18:47
Operator : SJ/JU
Sample : PB103993BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB103993BL

Quant Time: Nov 10 00:51:41 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110817.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 09 16:04:38 2017
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.90 min Scan# 819
Delta R.T. 0.00 min
Lab File: BF100348.D
Acq: 9 Nov 2017 18:47

Instrument :
BNA_F
ClientSampleId :
PB103993BL

Tot Ion:152 Resp: 157824
Ion Ratio Lower Upper
152 100
150 154.5 124.6 186.8
115 56.9 50.1 75.1




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 103.388 ng
RT: 5.53 min Scan# 585
Delta R.T. 0.02 min
Lab File: BF100348.D
Acq: 9 Nov 2017 18:47

Tgt Ion:112 Resp: 1017917
Ion Ratio Lower Upper
112 100
64 62.9 47.9 71.9
63 32.3 23.7 35.5



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF1
Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 102.609 ng

RT: 6.52 min Scan# 754

Delta R.T. 0.00 min

Lab File: BF100348.D

Acq: 9 Nov 2017 18:47

Instrument :

BNA_F

ClientSampled :

PB103993BL





Tot Ion: 99 Resp: 1245769

Ion Ratio Lower Upper

99 100

42 21.2 14.5 21.7

71 33.5 25.4 38.0



#21

Naphthalene-d8

Concen: 20.000 ng

RT: 8.18 min Scan# 1036

Delta R.T. -0.01 min

Lab File: BF100348.D

Acq: 9 Nov 2017 18:47

Tgt Ion: 136 Resp: 633397




Ion Ratio Lower Upper


136 100

137 11.5 8.9 13.3

54 9.1 6.6 9.8

68 6.5 5.0 7.4









#23
 Nitrobenzene-d5
 Concen: 81.508 ng
 RT: 7.46 min Scan# 914
 Delta R.T. -0.01 min
 Lab File: BF100348.D
 Acq: 9 Nov 2017 18:47

Instrument :
 BNA_F
 ClientSampleId :
 PB103993BL

Tot Ion: 82 Resp: 780882
 Ion Ratio Lower Upper
 82 100
 128 46.3 36.1 54.1
 54 53.7 41.4 62.2




Abundance Ion 82.00 (81.70 to 82.70): BF1
 Ion 128.00 (127.70 to 128.70): BF1
 Ion 54.00 (53.70 to 54.70): BF1






#38
 Acenaphthene-d10
 Concen: 20.000 ng
 RT: 9.94 min Scan# 1335
 Delta R.T. 0.00 min
 Lab File: BF100348.D
 Acq: 9 Nov 2017 18:47

Tgt Ion: 164 Resp: 306056
 Ion Ratio Lower Upper
 164 100
 162 103.1 86.1 129.1
 160 47.3 38.3 57.5



Abundance Ion 164.00 (163.70 to 164.70): E
 Ion 162.00 (161.70 to 162.70): E
 Ion 160.00 (159.70 to 160.70): E









#41
 2,4,6-Tribromophenol
 Concen: 111.515 ng
 RT: 10.73 min Scan# 1469
 Delta R.T. 0.00 min
 Lab File: BF100348.D
 Acq: 9 Nov 2017 18:47

Instrument :
 BNA_F
 ClientSampleId :
 PB103993BL

Tot Ion	Ratio	Lower	Upper
330	100		
332	95.5	77.0	115.6
141	30.7	29.9	44.9




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 74.145 ng
 RT: 9.26 min Scan# 1219
 Delta R.T. -0.01 min
 Lab File: BF100348.D
 Acq: 9 Nov 2017 18:47

Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.2	29.7	44.5
170	23.9	19.6	29.4



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#63

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.42 min Scan# 1587

Delta R.T. -0.01 min

Lab File: BF100348.D

Acq: 9 Nov 2017 18:47

Instrument :

BNA_F

ClientSampleId :

PB103993BL


Tot Ion:188 Resp: 620144

Ion Ratio Lower Upper

188 100

94 11.0 8.5 12.7




80 11.9 9.2 13.8



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1



#75

Chrysene-d12

Concen: 20.000 ng

RT: 14.06 min Scan# 2036

Delta R.T. -0.01 min

Lab File: BF100348.D

Acq: 9 Nov 2017 18:47


Tgt Ion:240 Resp: 499050

Ion Ratio Lower Upper

240 100

120 11.8 9.5 14.3



236 25.8 20.7 31.1




Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E





#78

Terphenyl-d14

Concen: 69.309 ng

RT: 13.02 min Scan# 1858

Delta R.T. 0.01 min

Lab File: BF100348.D

Acq: 9 Nov 2017 18:47

Instrument :

BNA_F

ClientSampleId :

PB103993BL


Tot Ion:244 Resp: 1505359

Ion Ratio Lower Upper

244 100

212 7.2 5.4 8.0




122 10.9 11.0 16.4#



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.54 min Scan# 2288

Delta R.T. 0.00 min

Lab File: BF100348.D

Acq: 9 Nov 2017 18:47


Tgt Ion:264 Resp: 373247

Ion Ratio Lower Upper

264 100

260 23.0 19.2 28.8

265 21.7 17.5 26.3



Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E

