

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF111015\
 Data File : BF082890.D
 Acq On : 11 Nov 2015 1:07
 Operator : UM/IZ
 Sample : G4253-07
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 1050-BOTTOM

Quant Time: Nov 13 17:39:20 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF110715.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 07 02:32:49 2015
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.82	152	219162	20.00	ng	-0.03
21) Naphthalene-d8	8.10	136	753080	20.00	ng	-0.05
38) Acenaphthene-d10	9.87	164	329515	20.00	ng	-0.02
63) Phenanthrene-d10	11.37	188	518281	20.00	ng	-0.01
75) Chrysene-d12	13.99	240	468467	20.00	ng	-0.03
86) Perylene-d12	15.40	264	495371	20.00	ng	-0.06
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	1524829	108.25	ng	-0.02
7) Phenol-d6	6.45	99	1839784	98.25	ng	-0.02
23) Nitrobenzene-d5	7.38	82	1255846	72.56	ng	-0.03
41) 2,4,6-Tribromophenol	10.67	330	366893	97.11	ng	-0.01
44) 2-Fluorobiphenyl	9.18	172	1643584	71.49	ng	-0.03
78) Terphenyl-d14	12.95	244	1578182	79.90	ng	-0.02
Target Compounds						Qvalue
25) Isophorone	7.38	82	1229355	38.39	ng	# 70
35) Caprolactam	8.50	113	12828	3.39	ng	# 1
49) Dimethylphthalate	9.58	163	383776	14.22	ng	# 91
52) 3-Nitroaniline	9.79	138	16604	2.62	ng	# 24
57) Fluorene	10.41	166	49323	2.06	ng	# 28
65) n-Nitrosodiphenylamine	10.54	169	612099	32.69	ng	# 37
68) Atrazine	11.09	200	36543	6.31	ng	# 27
70) Phenanthrene	11.39	178	512119	17.02	ng	# 89
71) Anthracene	11.44	178	137666	4.34	ng	# 68
72) Carbazole	11.59	167	61964	2.23	ng	# 46
74) Fluoranthene	12.58	202	1149509	35.59	ng	95
77) Pyrene	12.80	202	1118485	34.77	ng	98
80) Benzo(a)anthracene	13.97	228	514051	17.55	ng	95
82) Chrysene	14.01	228	583661	21.75	ng	98
83) Bis(2-ethylhexyl)phthalate	13.97	149	87778	4.51	ng	99
85) Indeno(1,2,3-cd)pyrene	16.77	276	376095	16.28	ng	93
87) Benzo(b)fluoranthene	14.99	252	1012452	31.84	ng	# 96
88) Benzo(k)fluoranthene	14.99	252	1012452	35.89	ng	# 96
89) Benzo(a)pyrene	15.33	252	472434	17.15	ng	# 96
90) Dibenzo(a,h)anthracene	16.77	278	97764	4.19	ng	# 94
91) Benzo(g,h,i)perylene	17.19	276	364115	16.30	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF111015\
Data File : BF082890.D
Acq On : 11 Nov 2015 1:07
Operator : UM/IZ
Sample : G4253-07
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
1050-BOTTOM

Quant Time: Nov 13 17:39:20 2015
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF110715.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Nov 07 02:32:49 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 6.82 min Scan# 414

Delta R.T. -0.03 min

Lab File: BF082890.D

Acq: 11 Nov 2015 1:07

Instrument :

BNA_F

ClientSampleId :

1050-BOTTOM


Tgt Ion:152 Resp: 219162

Ion Ratio Lower Upper

152 100

150 152.9 127.0 190.4




115 55.6 47.0 70.6



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#5

2-Fluorophenol

Concen: 108.25 ng

RT: 5.42 min Scan# 292

Delta R.T. -0.02 min

Lab File: BF082890.D

Acq: 11 Nov 2015 1:07


Tgt Ion:112 Resp: 1524829

Ion Ratio Lower Upper

112 100

64 59.0 48.9 73.3



63 30.1 28.1 42.1




Abundance Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF0

Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 98.25 ng

RT: 6.45 min Scan# 382

Delta R.T. -0.02 min

Lab File: BF082890.D

Acq: 11 Nov 2015 1:07

Instrument :

BNA_F

ClientSampleId :

1050-BOTTOM


Tgt Ion: 99 Resp: 1839784

Ion Ratio Lower Upper

99 100

42 20.1 19.2 28.8




71 36.9 35.0 52.4



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#21

Naphthalene-d8

Concen: 20.00 ng

RT: 8.10 min Scan# 526

Delta R.T. -0.05 min

Lab File: BF082890.D

Acq: 11 Nov 2015 1:07

Tgt Ion: 136 Resp: 753080


Ion Ratio Lower Upper

136 100

137 11.4 9.1 13.7

54 11.8 9.1 13.7

68 8.8 4.8 7.2#






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0









#23
 Nitrobenzene-d5
 Concen: 72.56 ng
 RT: 7.38 min Scan# 463
 Delta R.T. -0.03 min
 Lab File: BF082890.D
 Acq: 11 Nov 2015 1:07




Instrument :
 BNA_F
 ClientSampleId :
 1050-BOTTOM


Tgt Ion: 82 Resp: 1255846
 Ion Ratio Lower Upper
 82 100
 128 38.4 26.8 40.2
 54 52.0 45.7 68.5



#25
 Isophorone
 Concen: 38.39 ng
 RT: 7.38 min Scan# 463
 Delta R.T. -0.30 min
 Lab File: BF082890.D
 Acq: 11 Nov 2015 1:07

Tgt Ion: 82 Resp: 1229355
 Ion Ratio Lower Upper
 82 100
 95 0.3 6.6 9.8#
 138 0.0 11.4 17.0#









#35
 Caprolactam
 Concen: 3.39 ng
 RT: 8.50 min Scan# 561
 Delta R.T. -0.07 min
 Lab File: BF082890.D
 Acq: 11 Nov 2015 1:07




Instrument :
 BNA_F
 ClientSampleId :
 1050-BOTTOM


Tgt Ion:113 Resp: 12828
 Ion Ratio Lower Upper
 113 100
 55 1119.3 195.5 235.5#
 56 586.1 153.5 193.5#



#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 9.87 min Scan# 681
 Delta R.T. -0.02 min
 Lab File: BF082890.D
 Acq: 11 Nov 2015 1:07

Tgt Ion:164 Resp: 329515
 Ion Ratio Lower Upper
 164 100
 162 100.9 84.9 127.3
 160 48.8 37.5 56.3









#41
 2,4,6-Tribromophenol
 Concen: 97.11 ng
 RT: 10.67 min Scan# 751
 Delta R.T. -0.01 min
 Lab File: BF082890.D
 Acq: 11 Nov 2015 1:07




Instrument :
 BNA_F
 ClientSampleId :
 1050-BOTTOM


Tgt Ion:330 Resp: 366893
 Ion Ratio Lower Upper
 330 100
 332 96.6 76.6 114.8
 141 46.9 41.2 61.8



#44
 2-Fluorobiphenyl
 Concen: 71.49 ng
 RT: 9.18 min Scan# 621
 Delta R.T. -0.03 min
 Lab File: BF082890.D
 Acq: 11 Nov 2015 1:07

Tgt Ion:172 Resp: 1643584
 Ion Ratio Lower Upper
 172 100
 171 36.8 30.8 46.2
 170 25.2 21.2 31.8





#49

Dimethylphthalate

Concen: 14.22 ng

RT: 9.58 min Scan# 656

Delta R.T. -0.03 min

Lab File: BF082890.D

Acq: 11 Nov 2015 1:07

Instrument :

BNA_F

ClientSampleId :

1050-BOTTOM


Tgt Ion:163 Resp: 383776

Ion Ratio Lower Upper

163 100

194 14.3 2.6 4.0#




164 10.4 8.7 13.1



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#52

3-Nitroaniline

Concen: 2.62 ng

RT: 9.79 min Scan# 674

Delta R.T. -0.05 min

Lab File: BF082890.D

Acq: 11 Nov 2015 1:07


Tgt Ion:138 Resp: 16604

Ion Ratio Lower Upper

138 100

108 77.4 11.3 16.9#



92 64.2 122.4 183.6#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0





#57

Fluorene

Concen: 2.06 ng

RT: 10.41 min Scan# 728

Delta R.T. -0.03 min

Lab File: BF082890.D

Acq: 11 Nov 2015 1:07

Instrument :

BNA_F

ClientSampleId :

1050-BOTTOM





Tgt Ion:166 Resp: 49323

Ion Ratio Lower Upper

166 100

165 169.4 80.2 120.4#

167 50.7 11.2 16.8#



#63

Phenanthrene-d10

Concen: 20.00 ng

RT: 11.37 min Scan# 812

Delta R.T. -0.01 min

Lab File: BF082890.D

Acq: 11 Nov 2015 1:07




Tgt Ion:188 Resp: 518281


Ion Ratio Lower Upper

188 100

94 11.1 6.1 9.1#

80 11.6 7.0 10.6#









#65
n-Nitrosodiphenylamine
Concen: 32.69 ng
RT: 10.54 min Scan# 740
Delta R.T. -0.01 min
Lab File: BF082890.D
Acq: 11 Nov 2015 1:07




Instrument :
BNA_F
ClientSampleId :
1050-BOTTOM


Tgt Ion:169 Resp: 612099
Ion Ratio Lower Upper
169 100
168 9.9 55.0 82.6#
167 10.0 30.5 45.7#



#68
Atrazine
Concen: 6.31 ng
RT: 11.09 min Scan# 788
Delta R.T. 0.01 min
Lab File: BF082890.D
Acq: 11 Nov 2015 1:07

Tgt Ion:200 Resp: 36543
Ion Ratio Lower Upper
200 100
173 98.5 6.9 46.9#
215 63.4 22.9 62.9#





#70

Phenanthrene

Concen: 17.02 ng

RT: 11.39 min Scan# 814

Delta R.T. -0.01 min

Lab File: BF082890.D

Acq: 11 Nov 2015 1:07

Instrument :

BNA_F

ClientSampleId :

1050-BOTTOM


Tgt Ion:178 Resp: 512119

Ion Ratio Lower Upper

178 100

176 20.1 17.0 25.4




179 26.3 13.4 20.0#



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#71

Anthracene

Concen: 4.34 ng

RT: 11.44 min Scan# 818

Delta R.T. -0.01 min

Lab File: BF082890.D

Acq: 11 Nov 2015 1:07


Tgt Ion:178 Resp: 137666

Ion Ratio Lower Upper

178 100

176 22.4 16.5 24.7



179 45.1 13.2 19.8#




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E









#72
 Carbazole
 Concen: 2.23 ng
 RT: 11.59 min Scan# 831
 Delta R.T. -0.02 min
 Lab File: BF082890.D
 Acq: 11 Nov 2015 1:07




Instrument :
 BNA_F
 ClientSampleId :
 1050-BOTTOM


Tgt Ion:167 Resp: 61964
 Ion Ratio Lower Upper
 167 100
 166 44.6 18.7 28.1#
 139 43.6 11.8 17.6#



#74
 Fluoranthene
 Concen: 35.59 ng
 RT: 12.58 min Scan# 918
 Delta R.T. -0.01 min
 Lab File: BF082890.D
 Acq: 11 Nov 2015 1:07

Tgt Ion:202 Resp: 1149509
 Ion Ratio Lower Upper
 202 100
 101 5.9 0.0 30.5
 203 18.6 0.0 38.5





#75

Chrysene-d12

Concen: 20.00 ng

RT: 13.99 min Scan# 1041

Delta R.T. -0.03 min

Lab File: BF082890.D

Acq: 11 Nov 2015 1:07

Instrument :

BNA_F

ClientSampleId :

1050-BOTTOM


Tgt Ion:240 Resp: 468467

Ion Ratio Lower Upper

240 100

120 6.1 10.9 16.3#

236 26.1 21.6 32.4






Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E

Time-->



#77

Pyrene

Concen: 34.77 ng

RT: 12.80 min Scan# 937

Delta R.T. -0.01 min

Lab File: BF082890.D

Acq: 11 Nov 2015 1:07


Tgt Ion:202 Resp: 1118485

Ion Ratio Lower Upper

202 100

200 21.6 18.7 28.1

203 18.8 14.8 22.2






Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E

Time-->





#78

Terphenyl-d14

Concen: 79.90 ng

RT: 12.95 min Scan# 950

Delta R.T. -0.02 min

Lab File: BF082890.D

Acq: 11 Nov 2015 1:07

Instrument :

BNA_F

ClientSampleId :

1050-BOTTOM


Tgt Ion:244 Resp: 1578182

Ion Ratio Lower Upper

244 100

212 7.9 7.6 11.4




122 8.9 11.3 16.9#



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Benzo(a)anthracene

Concen: 17.55 ng

RT: 13.97 min Scan# 1040

Delta R.T. -0.03 min

Lab File: BF082890.D

Acq: 11 Nov 2015 1:07


Tgt Ion:228 Resp: 514051

Ion Ratio Lower Upper

228 100

226 32.7 23.4 35.2



229 21.9 16.3 24.5




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#82
 Chrysene
 Concen: 21.75 ng
 RT: 14.01 min Scan# 1043
 Delta R.T. -0.03 min
 Lab File: BF082890.D
 Acq: 11 Nov 2015 1:07




Instrument :
 BNA_F
 ClientSampleId :
 1050-BOTTOM


Tgt Ion: 228 Resp: 583661
 Ion Ratio Lower Upper
 228 100
 226 30.8 25.5 38.3
 229 21.1 16.3 24.5



#83
 Bis(2-ethylhexyl)phthalate
 Concen: 4.51 ng
 RT: 13.97 min Scan# 1040
 Delta R.T. -0.03 min
 Lab File: BF082890.D
 Acq: 11 Nov 2015 1:07

Tgt Ion: 149 Resp: 87778
 Ion Ratio Lower Upper
 149 100
 167 25.8 21.2 31.8
 279 2.9 2.2 3.2





#85

Indeno(1,2,3-cd)pyrene

Concen: 16.28 ng

RT: 16.77 min Scan# 1285

Delta R.T. -0.08 min

Lab File: BF082890.D

Acq: 11 Nov 2015 1:07

Instrument :

BNA_F

ClientSampleId :

1050-BOTTOM


Tgt Ion: 276 Resp: 376095

Ion Ratio Lower Upper

276 100

138 22.5 15.3 22.9




277 28.9 20.2 30.4



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#86

Perylene-d12

Concen: 20.00 ng

RT: 15.40 min Scan# 1165

Delta R.T. -0.06 min

Lab File: BF082890.D

Acq: 11 Nov 2015 1:07


Tgt Ion: 264 Resp: 495371

Ion Ratio Lower Upper

264 100

260 23.9 19.8 29.6



265 21.7 17.4 26.0




Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E









#87
Benzo(b)fluoranthene
Concen: 31.84 ng
RT: 14.99 min Scan# 1129
Delta R.T. -0.06 min
Lab File: BF082890.D
Acq: 11 Nov 2015 1:07




Instrument :
BNA_F
ClientSampleId :
1050-BOTTOM


Tgt Ion:252 Resp: 1012452
Ion Ratio Lower Upper
252 100
253 22.1 17.8 26.8
125 12.4 5.4 8.2#



#88
Benzo(k)fluoranthene
Concen: 35.89 ng
RT: 14.99 min Scan# 1129
Delta R.T. -0.09 min
Lab File: BF082890.D
Acq: 11 Nov 2015 1:07

Tgt Ion:252 Resp: 1012452
Ion Ratio Lower Upper
252 100
253 22.1 17.9 26.9
125 12.4 3.5 5.3#





#89

Benzo(a)pyrene

Concen: 17.15 ng

RT: 15.33 min Scan# 1159

Delta R.T. -0.07 min

Lab File: BF082890.D

Acq: 11 Nov 2015 1:07

Instrument :

BNA_F

ClientSampleId :

1050-BOTTOM


Tgt Ion:252 Resp: 472434

Ion Ratio Lower Upper

252 100

253 22.6 17.3 25.9




125 12.0 7.1 10.7#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 4.19 ng

RT: 16.77 min Scan# 1285

Delta R.T. -0.09 min

Lab File: BF082890.D

Acq: 11 Nov 2015 1:07


Tgt Ion:278 Resp: 97764

Ion Ratio Lower Upper

278 100

139 18.8 9.9 14.9#



279 24.8 19.6 29.4




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 16.30 ng

RT: 17.19 min Scan# 1321

Delta R.T. -0.08 min

Lab File: BF082890.D

Acq: 11 Nov 2015 1:07

Instrument :

BNA_F

ClientSampleId :

1050-BOTTOM


Tgt Ion: 276 Resp: 364115

Ion Ratio Lower Upper

276 100

277 23.8 19.2 28.8

138 19.1 13.1 19.7



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 277.00 (276.70 to 277.70): E

Ion 138.00 (137.70 to 138.70): E

