




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111021\
Data File : BP007920.D
Acq On : 10 Nov 2021 15:23
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 10 17:19:38 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 09 10:45:24 2021
Response via : Initial Calibration



TIC: BP007920.D\data.ms

(70) Pentachlorophenol (C)

16.904min (-0.006) 206108.65 ng

response 361903

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	61.60	66.53
264.00	60.90	63.48
0.00	0.00	0.00