




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP111021\
 Data File : BP007920.D
 Acq On : 10 Nov 2021 15:23
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 10 17:19:38 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP110221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 09 10:45:24 2021
 Response via : Initial Calibration



TIC: BP007920.D\data.ms

(77) Benzidine

19.445min (-0.000) 56799.72 ng

response 2035619

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.00	14.20
183.00	11.40	12.16
0.00	0.00	0.00