

Data Path : Z:\HPCHEM1\BNA F\DATA\BF111116\
 Data File : BF091139.D
 Acq On : 11 Nov 2016 14:49
 Operator : UM/SJ
 Sample : PB94644BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB94644BL

Quant Time: Nov 11 15:55:11 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 08 12:59:29 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.65	152	147610	20.00	ng	-0.01
21) Naphthalene-d8	7.93	136	643331	20.00	ng	-0.02
38) Acenaphthene-d10	9.68	164	320913	20.00	ng	-0.02
63) Phenanthrene-d10	11.15	188	555430	20.00	ng	-0.02
75) Chrysene-d12	13.78	240	533003	20.00	ng	-0.03
86) Perylene-d12	15.15	264	481021	20.00	ng	-0.03
System Monitoring Compounds						
5) 2-Fluorophenol	5.24	112	718061	80.34	ng	0.00
7) Phenol-d6	6.29	99	988570	81.49	ng	-0.02
23) Nitrobenzene-d5	7.21	82	1007474	85.43	ng	-0.03
41) 2,4,6-Tribromophenol	10.46	330	218081	75.17	ng	-0.02
44) 2-Fluorobiphenyl	9.01	172	1749180	93.84	ng	-0.02
78) Terphenyl-d14	12.74	244	1763941	82.59	ng	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF111116\
Data File : BF091139.D
Acq On : 11 Nov 2016 14:49
Operator : UM/SJ
Sample : PB94644BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB94644BL

Quant Time: Nov 11 15:55:11 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 08 12:59:29 2016
Response via : Initial Calibration

