


Data Path : Z:\HPCHEM1\BNA F\DATA\BF111116\
 Data File : BF091137.D
 Acq On : 11 Nov 2016 13:39
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF110316.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 08 12:59:29 2016



AutoFind: Scans 851, 852, 853; Background Corrected with Scan 848

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.9	12561	PASS
68	69	0.00	2	1.2	160	PASS
69	198	0.00	100	43.0	13557	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	56.8	17888	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	31518	PASS
199	198	5	9	6.4	2009	PASS
275	198	10	60	23.7	7476	PASS
365	198	1	100	2.7	851	PASS
441	442	0.01	24	15.3	4479	PASS
442	198	50	100	92.9	29267	PASS
443	442	15	24	19.2	5608	PASS