

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111122\
 Data File : BF131175.D
 Acq On : 12 Nov 2022 03:07
 Operator : CG\JU
 Sample : N5501-03 2X
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 HS-4-110722

Manual Integrations
 APPROVED

Reviewed By : Christian Giraldo 11/12/2022
 Supervised By : Jagrut Upadhyay 11/12/2022

Quant Time: Nov 12 03:49:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110922.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Nov 12 00:03:31 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	78532	20.000	ng	0.00
21) Naphthalene-d8	8.163	136	265457	20.000	ng	0.00
39) Acenaphthene-d10	9.922	164	130315	20.000	ng	0.00
64) Phenanthrene-d10	11.416	188	239640	20.000	ng	0.00
76) Chrysene-d12	14.069	240	134632	20.000	ng	# 0.01
86) Perylene-d12	15.568	264	101696	20.000	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.493	112	177740	38.490	ng	0.00
7) Phenol-d6	6.504	99	211959	35.108	ng	0.00
23) Nitrobenzene-d5	7.440	82	121157	20.593	ng	0.00
42) 2,4,6-Tribromophenol	10.716	330	42536	31.584	ng	0.00
45) 2-Fluorobiphenyl	9.234	172	220198	23.588	ng	0.00
79) Terphenyl-d14	13.010	244	208271	27.123	ng	0.01
Target Compounds						Qvalue
75) Fluoranthene	12.639	202	30355	2.301	ng	95
78) Pyrene	12.874	202	32745	2.835	ng	97
88) Benzo(b)fluoranthene	15.127	252	17256m	2.400	ng	
92) Benzo(g,h,i)perylene	17.551	276	9340	5.553	ng	# 67

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111122\
Data File : BF131175.D
Acq On : 12 Nov 2022 03:07
Operator : CG\JU
Sample : N5501-03 2X
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
HS-4-110722

Quant Time: Nov 12 03:49:12 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF110922.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Nov 12 00:03:31 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Christian Giraldo 11/12/2022
Supervised By : Jagrut Upadhyay 11/12/2022

