

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111124\
 Data File : BF140328.D
 Acq On : 11 Nov 2024 20:10
 Operator : RC/JU
 Sample : P4792-06
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 OILY-STONE-DRUM

Quant Time: Nov 12 00:45:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 12 00:05:37 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.875	152	140396	20.000	ng	0.00
21) Naphthalene-d8	8.157	136	538917	20.000	ng	0.00
39) Acenaphthene-d10	9.910	164	295572	20.000	ng	0.00
64) Phenanthrene-d10	11.398	188	558313	20.000	ng	0.00
76) Chrysene-d12	14.039	240	344447	20.000	ng	-0.02
86) Perylene-d12	15.533	264	281581	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.504	112	931938	114.681	ng	0.01
7) Phenol-d6	6.504	99	1150798	106.574	ng	0.00
23) Nitrobenzene-d5	7.434	82	872831	84.079	ng	-0.01
42) 2,4,6-Tribromophenol	10.704	330	380222	132.411	ng	0.00
45) 2-Fluorobiphenyl	9.228	172	1548950	82.737	ng	0.00
79) Terphenyl-d14	12.986	244	1833661	91.283	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111124\
Data File : BF140328.D
Acq On : 11 Nov 2024 20:10
Operator : RC/JU
Sample : P4792-06
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
OILY-STONE-DRUM

Quant Time: Nov 12 00:45:57 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 12 00:05:37 2024
Response via : Initial Calibration

