


Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA F\DATA\BF111317\
 Data File : BF100521.D
 Acq On : 14 Nov 2017 00:55
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF110817.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 14 02:37:14 2017



AutoFind: Scans 1626, 1627, 1628; Background Corrected with Scan 1622

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	40.3	23826	PASS
68	69	0.00	2	1.7	336	PASS
69	198	0.00	100	34.4	20297	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	44.5	26315	PASS
197	198	0.00	2	0.4	208	PASS
198	198	100	100	100.0	59069	PASS
199	198	5	9	6.5	3836	PASS
275	198	10	60	24.4	14388	PASS
365	198	1	100	2.6	1546	PASS
441	443	0.01	100	81.7	6980	PASS
442	198	50	100	78.2	46184	PASS
443	442	15	24	18.5	8544	PASS