

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111321\
 Data File : BF126167.D
 Acq On : 13 Nov 2021 14:18
 Operator : JU/CG
 Sample : PB140723BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB140723BL

Quant Time: Nov 15 08:13:16 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110321.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 15 08:09:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.834	152	92752	20.000	ng	0.00
21) Naphthalene-d8	8.116	136	349639	20.000	ng	0.00
39) Acenaphthene-d10	9.869	164	217252	20.000	ng	0.00
64) Phenanthrene-d10	11.357	188	418483	20.000	ng	# 0.00
76) Chrysene-d12	13.998	240	295042	20.000	ng	# 0.00
86) Perylene-d12	15.462	264	221554	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.457	112	765977	141.819	ng	0.01
7) Phenol-d6	6.463	99	972179	127.687	ng	0.00
23) Nitrobenzene-d5	7.392	82	714090	96.319	ng	0.00
42) 2,4,6-Tribromophenol	10.663	330	401040	161.078	ng	0.00
45) 2-Fluorobiphenyl	9.192	172	1457213	90.642	ng	0.00
79) Terphenyl-d14	12.951	244	1856095	101.445	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111321\
Data File : BF126167.D
Acq On : 13 Nov 2021 14:18
Operator : JU/CG
Sample : PB140723BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB140723BL

Quant Time: Nov 15 08:13:16 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110321.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 15 08:09:33 2021
Response via : Initial Calibration

