


Data Path : Z:\HPCHEM1\BNA_F\Data\BF111414\
 Data File : BF075571.D
 Acq On : 16 Nov 2014 5:24
 Operator : TP / JJ
 Sample : DFTTP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTTP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF111214.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Nov 17 01:17:49 2014



AutoFind: Scans 1087, 1088, 1089; Background Corrected with Scan 1084

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	47.4	10292	PASS
68	69	0.00	2	1.6	119	PASS
69	198	0.00	100	34.9	7565	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	47.1	10216	PASS
197	198	0.00	2	0.7	147	PASS
198	198	100	100	100.0	21695	PASS
199	198	5	9	6.3	1376	PASS
275	198	10	60	21.1	4579	PASS
365	198	1	100	1.9	421	PASS
441	442	0.01	24	16.3	3308	PASS
442	198	50	100	93.6	20307	PASS
443	442	15	24	19.8	4013	PASS