


Data Path : Z:\HPCHEM1\BNA F\DATA\BF111416\
 Data File : BF091161.D
 Acq On : 14 Nov 2016 13:17
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF110316.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 08 12:59:29 2016



AutoFind: Scans 851, 852, 853; Background Corrected with Scan 847

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.2	20469	PASS
68	69	0.00	2	1.6	357	PASS
69	198	0.00	100	37.5	22429	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	52.2	31202	PASS
197	198	0.00	2	0.2	128	PASS
198	198	100	100	100.0	59786	PASS
199	198	5	9	6.8	4041	PASS
275	198	10	60	25.5	15257	PASS
365	198	1	100	2.8	1647	PASS
441	442	0.01	24	15.0	6815	PASS
442	198	50	100	75.9	45348	PASS
443	442	15	24	19.2	8685	PASS