




Data File : BF100561.D
Acq On : 15 Nov 2017 00:04
Operator : SJ/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 15 07:53:37 2017
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF110817.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 14 14:17:16 2017
Response via : Initial Calibration



TIC: BF100561.D

(69) Pentachlorophenol (C)

11.204min (-0.006) 0.00ng

response 51374

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	62.13
264.00	61.90	66.06
0.00	0.00	0.00

Data File : BF100561.D

Acq On : 15 Nov 2017 00:04

Operator : SJ/JU

Sample : DFTPP

Misc :

ALS Vial : 1 Sample Multiplier: 1




Quant Time: Nov 15 07:53:37 2017

Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF110817.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Tue Nov 14 14:17:16 2017

Response via : Initial Calibration



TIC: BF100561.D

(76) Benzidine

12.739min (-0.000) 0.00ng

response 304795

Ion	Exp%	Act%
184.00	100	100
185.00	15.70	14.39
183.00	11.60	11.70
0.00	0.00	0.00