

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF111418\
 Data File : BF110778.D
 Acq On : 14 Nov 2018 21:34
 Operator : JU/SJ
 Sample : J5911-08 20X
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FDSB-06(21-23)

Quant Time: Nov 15 06:09:37 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 13 12:43:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.95	152	70173	20.00	ng	0.00
21) Naphthalene-d8	8.23	136	308906	20.00	ng	0.00
39) Acenaphthene-d10	9.98	164	143952	20.00	ng	-0.01
64) Phenanthrene-d10	11.48	188	248976	20.00	ng	0.00
76) Chrysene-d12	14.13	240	164933	20.00	ng	-0.01
87) Perylene-d12	15.66	264	157269	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.57	112	26637	6.17	ng	0.00
7) Phenol-d6	6.57	99	35728	6.47	ng	0.00
23) Nitrobenzene-d5	7.52	82	21141	3.94	ng	0.00
42) 2,4,6-Tribromophenol	10.79	330	6933	4.78	ng	0.00
45) 2-Fluorobiphenyl	9.30	172	47969	4.76	ng	0.00
79) Terphenyl-d14	13.06	244	35162	3.99	ng	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF111418\
Data File : BF110778.D
Acq On : 14 Nov 2018 21:34
Operator : JU/SJ
Sample : J5911-08 20X
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FDSB-06(21-23)

Quant Time: Nov 15 06:09:37 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 13 12:43:10 2018
Response via : Initial Calibration

