

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111422\
 Data File : BF131220.D
 Acq On : 14 Nov 2022 12:02
 Operator : CG\JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Quant Time: Nov 15 03:26:02 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 15 03:24:29 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.804	152	73520	20.000	ng	0.00
21) Naphthalene-d8	8.092	136	232735	20.000	ng	0.00
39) Acenaphthene-d10	9.851	164	133726	20.000	ng	0.00
64) Phenanthrene-d10	11.339	188	256235	20.000	ng	0.00
76) Chrysene-d12	13.986	240	208366	20.000	ng	0.00
86) Perylene-d12	15.451	264	171946	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.422	112	23609	4.913	ng	0.00
7) Phenol-d6	6.434	99	31847	5.324	ng	-0.01
23) Nitrobenzene-d5	7.369	82	25365	4.676	ng	-0.01
42) 2,4,6-Tribromophenol	10.639	330	5666	4.278	ng	0.00
45) 2-Fluorobiphenyl	9.163	172	51374	5.335	ng	-0.01
79) Terphenyl-d14	12.921	244	57880	4.766	ng	-0.01
Target Compounds						
19) n-Nitroso-di-n-propyla...	7.210	70	11626	2.824	ng	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111422\
Data File : BF131220.D
Acq On : 14 Nov 2022 12:02
Operator : CG\JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Quant Time: Nov 15 03:26:02 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 15 03:24:29 2022
Response via : Initial Calibration

