

Data Path : Z:\HPCHEM1\BNA F\DATA\BF111516\
 Data File : BF091168.D
 Acq On : 15 Nov 2016 9:45
 Operator : UM/SJ
 Sample : PB94688BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB94688BL

Quant Time: Nov 16 00:23:48 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 08 12:59:29 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.64	152	166769	20.00	ng	-0.02
21) Naphthalene-d8	7.92	136	708803	20.00	ng	-0.03
38) Acenaphthene-d10	9.66	164	375359	20.00	ng	-0.03
63) Phenanthrene-d10	11.15	188	690785	20.00	ng	-0.02
75) Chrysene-d12	13.78	240	619161	20.00	ng	-0.03
86) Perylene-d12	15.14	264	509303	20.00	ng	-0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.23	112	783750	77.62	ng	-0.01
7) Phenol-d6	6.28	99	959711	70.02	ng	-0.03
23) Nitrobenzene-d5	7.21	82	1181949	90.96	ng	-0.03
41) 2,4,6-Tribromophenol	10.45	330	231103	68.10	ng	-0.03
44) 2-Fluorobiphenyl	9.00	172	1865365	85.56	ng	-0.03
78) Terphenyl-d14	12.74	244	2179840	87.86	ng	-0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF111516\
Data File : BF091168.D
Acq On : 15 Nov 2016 9:45
Operator : UM/SJ
Sample : PB94688BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB94688BL

Quant Time: Nov 16 00:23:48 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 08 12:59:29 2016
Response via : Initial Calibration

