

Data Path : Z:\HPCHEM1\BNA F\DATA\BF111516\
 Data File : BF091171.D
 Acq On : 15 Nov 2016 11:13
 Operator : UM/SJ
 Sample : H5643-02
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 MW-2-111016

Quant Time: Nov 16 00:43:57 2016
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110316.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 08 12:59:29 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.64	152	156624	20.00	ng	-0.02
21) Naphthalene-d8	7.93	136	631051	20.00	ng	-0.02
38) Acenaphthene-d10	9.66	164	342606	20.00	ng	-0.03
63) Phenanthrene-d10	11.15	188	611201	20.00	ng	-0.02
75) Chrysene-d12	13.78	240	522725	20.00	ng	-0.03
86) Perylene-d12	15.14	264	400661	20.00	ng	-0.05
System Monitoring Compounds						
5) 2-Fluorophenol	5.22	112	694071	73.19	ng	-0.02
7) Phenol-d6	6.29	99	546926	42.49	ng	-0.02
23) Nitrobenzene-d5	7.21	82	975832	84.35	ng	-0.03
41) 2,4,6-Tribromophenol	10.46	330	589603	190.36	ng	-0.02
44) 2-Fluorobiphenyl	9.00	172	1907680	95.86	ng	-0.03
78) Terphenyl-d14	12.74	244	2148838	102.58	ng	-0.02
Target Compounds						Qvalue
31) Naphthalene	7.94	128	1283148	42.52	ng	99
37) 2-Methylnaphthalene	8.64	142	708225	36.50	ng	98
49) Dimethylphthalate	9.40	163	90290	4.02	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF111516\
Data File : BF091171.D
Acq On : 15 Nov 2016 11:13
Operator : UM/SJ
Sample : H5643-02
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
MW-2-111016

Quant Time: Nov 16 00:43:57 2016
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110316.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 08 12:59:29 2016
Response via : Initial Calibration

