


Data Path : Z:\HPCHEM1\BNA F\DATA\BF111516\  
 Data File : BF091166.D  
 Acq On : 15 Nov 2016 8:50  
 Operator : UM/SJ  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF110316.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Tue Nov 08 12:59:29 2016



AutoFind: Scans 851, 852, 853; Background Corrected with Scan 847

| Target Mass | Rel. to Mass | Lower Limit% | Upper Limit% | Rel. Abn% | Raw Abn | Result Pass/Fail |
|-------------|--------------|--------------|--------------|-----------|---------|------------------|
| 51          | 198          | 10           | 80           | 34.4      | 22794   | PASS             |
| 68          | 69           | 0.00         | 2            | 1.5       | 380     | PASS             |
| 69          | 198          | 0.00         | 100          | 37.2      | 24689   | PASS             |
| 70          | 69           | 0.00         | 2            | 0.0       | 0       | PASS             |
| 127         | 198          | 10           | 80           | 51.5      | 34190   | PASS             |
| 197         | 198          | 0.00         | 2            | 0.3       | 167     | PASS             |
| 198         | 198          | 100          | 100          | 100.0     | 66337   | PASS             |
| 199         | 198          | 5            | 9            | 6.5       | 4309    | PASS             |
| 275         | 198          | 10           | 60           | 24.9      | 16495   | PASS             |
| 365         | 198          | 1            | 100          | 2.6       | 1743    | PASS             |
| 441         | 442          | 0.01         | 24           | 15.5      | 7328    | PASS             |
| 442         | 198          | 50           | 100          | 71.0      | 47130   | PASS             |
| 443         | 442          | 15           | 24           | 19.4      | 9165    | PASS             |