

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111522\
 Data File : BF131241.D
 Acq On : 15 Nov 2022 19:43
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF111522

Quant Time: Nov 16 06:42:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 16 06:35:36 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	98	0.00
2	1,4-Dioxane	40.000	39.833	0.4	96	0.00
3	Pyridine	40.000	40.029	-0.1	97	0.00
4	n-Nitrosodimethylamine	40.000	41.106	-2.8	98	0.00
5 S	2-Fluorophenol	80.000	80.085	-0.1	98	0.00
6	Aniline	40.000	40.058	-0.1	97	0.00
7 S	Phenol-d6	80.000	79.984	0.0	98	0.00
8	2-Chlorophenol	40.000	40.728	-1.8	100	0.00
9	Benzaldehyde	40.000	36.117	9.7	99	0.00
10 C	Phenol	40.000	40.119	-0.3	99	0.00
11	bis(2-Chloroethyl)ether	40.000	39.751	0.6	98	0.00
12	1,3-Dichlorobenzene	40.000	39.759	0.6	98	0.00
13 C	1,4-Dichlorobenzene	40.000	39.536	1.2	97	0.00
14	1,2-Dichlorobenzene	40.000	39.936	0.2	99	0.00
15	Benzyl Alcohol	40.000	41.122	-2.8	98	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	40.040	-0.1	97	0.00
17	2-Methylphenol	40.000	40.561	-1.4	98	0.00
18	Hexachloroethane	40.000	41.795	-4.5	102	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	39.653	0.9	98	0.00
20	3+4-Methylphenols	40.000	40.551	-1.4	98	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	97	0.00
22	Acetophenone	40.000	39.154	2.1	98	0.00
23 S	Nitrobenzene-d5	80.000	89.336	-11.7	104	0.00
24	Nitrobenzene	40.000	43.783	-9.5	103	0.00
25	Isophorone	40.000	39.760	0.6	98	0.00
26 C	2-Nitrophenol	40.000	42.826	-7.1	110	0.00
27	2,4-Dimethylphenol	40.000	40.344	-0.9	98	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.274	1.8	97	0.00
29 C	2,4-Dichlorophenol	40.000	40.718	-1.8	99	0.00
30	1,2,4-Trichlorobenzene	40.000	39.566	1.1	98	0.00
31	Naphthalene	40.000	39.090	2.3	97	0.00
32	Benzoic acid	40.000	42.931	-7.3	112	0.00
33	4-Chloroaniline	40.000	39.506	1.2	99	0.00
34 C	Hexachlorobutadiene	40.000	40.357	-0.9	98	0.00
35	Caprolactam	40.000	41.175	-2.9	101	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.929	-2.3	101	0.00
37	2-Methylnaphthalene	40.000	39.252	1.9	98	0.00
38	1-Methylnaphthalene	40.000	39.125	2.2	97	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.680	3.3	99	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.262	-3.2	98	0.00
42 S	2,4,6-Tribromophenol	80.000	82.294	-2.9	101	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.776	-1.9	99	0.00
44	2,4,5-Trichlorophenol	40.000	40.992	-2.5	103	0.00
45 S	2-Fluorobiphenyl	80.000	76.071	4.9	100	0.00
46	1,1'-Biphenyl	40.000	38.526	3.7	99	0.00
47	2-Chloronaphthalene	40.000	38.750	3.1	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111522\
 Data File : BF131241.D
 Acq On : 15 Nov 2022 19:43
 Operator : CG\JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 ICVBF111522

Quant Time: Nov 16 06:42:14 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 16 06:35:36 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.636	-4.1	109	0.00
49	Acenaphthylene	40.000	38.907	2.7	99	0.00
50	Dimethylphthalate	40.000	38.821	2.9	99	0.00
51	2,6-Dinitrotoluene	40.000	41.651	-4.1	105	0.00
52 C	Acenaphthene	40.000	39.171	2.1	100	0.00
53	3-Nitroaniline	40.000	41.134	-2.8	105	0.00
54 P	2,4-Dinitrophenol	40.000	44.145	-10.4	120	0.00
55	Dibenzofuran	40.000	38.384	4.0	99	0.00
56 P	4-Nitrophenol	40.000	43.617	-9.0	108	0.00
57	2,4-Dinitrotoluene	40.000	42.315	-5.8	110	0.00
58	Fluorene	40.000	38.630	3.4	100	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.848	-2.1	99	0.00
60	Diethylphthalate	40.000	39.431	1.4	100	0.00
61	4-Chlorophenyl-phenylether	40.000	39.198	2.0	99	0.00
62	4-Nitroaniline	40.000	41.808	-4.5	107	0.00
63	Azobenzene	40.000	38.964	2.6	99	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	100	0.00
65	4,6-Dinitro-2-methylphenol	40.000	41.892	-4.7	112	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.114	2.2	100	0.00
67	4-Bromophenyl-phenylether	40.000	39.489	1.3	99	0.00
68	Hexachlorobenzene	40.000	38.947	2.6	98	0.00
69	Atrazine	40.000	39.572	1.1	100	0.00
70 C	Pentachlorophenol	40.000	43.829	-9.6	105	0.00
71	Phenanthrene	40.000	38.876	2.8	100	0.00
72	Anthracene	40.000	39.630	0.9	101	0.00
73	Carbazole	40.000	39.172	2.1	99	0.00
74	Di-n-butylphthalate	40.000	39.555	1.1	98	0.00
75 C	Fluoranthene	40.000	39.463	1.3	100	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzidine	40.000	41.827	-4.6	98	0.00
78	Pyrene	40.000	40.478	-1.2	101	0.00
79 S	Terphenyl-d14	80.000	79.998	0.0	100	0.00
80	Butylbenzylphthalate	40.000	42.428	-6.1	100	0.00
81	Benzo(a)anthracene	40.000	39.178	2.1	99	0.00
82	3,3'-Dichlorobenzidine	40.000	39.435	1.4	97	0.00
83	Chrysene	40.000	38.801	3.0	97	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.220	-3.0	100	0.00
85 c	Di-n-octyl phthalate	40.000	40.930	-2.3	102	0.00
86 I	Perylene-d12	20.000	20.000	0.0	99	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	42.632	-6.6	100	0.00
88	Benzo(b)fluoranthene	40.000	40.157	-0.4	99	0.00
89	Benzo(k)fluoranthene	40.000	39.116	2.2	99	0.00
90 C	Benzo(a)pyrene	40.000	40.127	-0.3	98	0.00
91	Dibenzo(a,h)anthracene	40.000	42.809	-7.0	100	0.00
92	Benzo(g,h,i)perylene	40.000	39.299	1.8	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111522\
Data File : BF131241.D
Acq On : 15 Nov 2022 19:43
Operator : CG\JU
Sample : SSTDICV040
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
ICVBF111522

Quant Time: Nov 16 06:42:14 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 16 06:35:36 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0