


Data Path : Z:\HPCHEM1\BNA F\DATA\BF111617\  
 Data File : BF100643.D  
 Acq On : 16 Nov 2017 21:50  
 Operator : SJ/JU  
 Sample : DFTPP  
 Misc :  
 ALS Vial : 1 Sample Multiplier: 1

Instrument :  
 BNA\_F  
 ClientSampleId :  
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA\_F\METHODS\8270-BF110817.M  
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 Last Update : Thu Nov 16 12:15:03 2017



AutoFind: Scans 1625, 1626, 1627; Background Corrected with Scan 1620

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 51             | 198             | 10              | 80              | 36.0         | 24957      | PASS                |
| 68             | 69              | 0.00            | 2               | 1.6          | 356        | PASS                |
| 69             | 198             | 0.00            | 100             | 31.9         | 22102      | PASS                |
| 70             | 69              | 0.00            | 2               | 0.0          | 0          | PASS                |
| 127            | 198             | 10              | 80              | 41.8         | 29032      | PASS                |
| 197            | 198             | 0.00            | 2               | 0.5          | 346        | PASS                |
| 198            | 198             | 100             | 100             | 100.0        | 69373      | PASS                |
| 199            | 198             | 5               | 9               | 6.6          | 4547       | PASS                |
| 275            | 198             | 10              | 60              | 24.6         | 17083      | PASS                |
| 365            | 198             | 1               | 100             | 3.0          | 2105       | PASS                |
| 441            | 443             | 0.01            | 100             | 84.6         | 9995       | PASS                |
| 442            | 198             | 50              | 100             | 92.4         | 64082      | PASS                |
| 443            | 442             | 15              | 24              | 18.4         | 11820      | PASS                |