

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF111618\
 Data File : BF110826.D
 Acq On : 16 Nov 2018 19:34
 Operator : JU/SJ
 Sample : J5951-03 10X
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FDSB-04(10-12)

Quant Time: Nov 16 23:59:55 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 13 12:43:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.94	152	48819	20.00	ng	0.00
21) Naphthalene-d8	8.23	136	209327	20.00	ng	0.00
39) Acenaphthene-d10	9.99	164	101426	20.00	ng	0.00
64) Phenanthrene-d10	11.47	188	166282	20.00	ng	-0.01
76) Chrysene-d12	14.13	240	144018	20.00	ng	-0.01
87) Perylene-d12	15.66	264	107650	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.56	112	30086	10.01	ng	0.00
7) Phenol-d6	6.57	99	43594	11.34	ng	0.00
23) Nitrobenzene-d5	7.51	82	25724	7.08	ng	0.00
42) 2,4,6-Tribromophenol	10.78	330	6745	6.60	ng	0.00
45) 2-Fluorobiphenyl	9.30	172	48556	6.84	ng	0.00
79) Terphenyl-d14	13.06	244	39695	5.16	ng	-0.01
Target Compounds						
37) 2-Methylnaphthalene	8.95	142	558302	86.667	ng	Qvalue 100
38) 1-Methylnaphthalene	9.05	142	428586	69.180	ng	# 99
58) Fluorene	10.53	166	26262	3.767	ng	# 83
71) Phenanthrene	11.50	178	19648	2.206	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF111618\
Data File : BF110826.D
Acq On : 16 Nov 2018 19:34
Operator : JU/SJ
Sample : J5951-03 10X
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FDSB-04(10-12)

Quant Time: Nov 16 23:59:55 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 13 12:43:10 2018
Response via : Initial Calibration

