

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF111618\
 Data File : BF110840.D
 Acq On : 17 Nov 2018 2:54
 Operator : JU/SJ
 Sample : J5967-01
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 FDSB-03

Manual Integrations
 APPROVED

Sohil
 11/19/2018 11:56:15 AM

Quant Time: Nov 17 04:09:10 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 13 12:43:10 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.96	152	37703	20.00	ng	0.01
21) Naphthalene-d8	8.22	136	47603	20.00	ng	-0.01
39) Acenaphthene-d10	10.03	164	77600	20.00	ng	0.04
64) Phenanthrene-d10	11.48	188	149510	20.00	ng	0.00
76) Chrysene-d12	14.13	240	144939	20.00	ng	-0.01
87) Perylene-d12	15.66	264	94304	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.56	112	350900	151.17	ng	0.00
7) Phenol-d6	6.57	99	420560	141.69	ng	0.00
23) Nitrobenzene-d5	7.53	82	379524	459.14	ng	0.02
42) 2,4,6-Tribromophenol	10.80	330	74348	95.08	ng	0.01
45) 2-Fluorobiphenyl	9.35	172	330800	60.88	ng	0.04
79) Terphenyl-d14	13.06	244	428400	55.35	ng	-0.01
Target Compounds						
31) Naphthalene	8.30	128	586435	262.422	ng	Qvalue # 55
37) 2-Methylnaphthalene	9.01	142	2241491	1530.074	ng	96
38) 1-Methylnaphthalene	9.10	142	1415825m	1004.941	ng	
50) Dimethylphthalate	9.73	163	61759	10.666	ng	# 73
58) Fluorene	10.56	166	117993	22.121	ng	# 68
71) Phenanthrene	11.50	178	85815	10.714	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF111618\
Data File : BF110840.D
Acq On : 17 Nov 2018 2:54
Operator : JU/SJ
Sample : J5967-01
Misc :
ALS Vial : 31 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
FDSB-03

Manual Integrations
APPROVED

Sohil
11/19/2018 11:56:15 AM

Quant Time: Nov 17 04:09:10 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 13 12:43:10 2018
Response via : Initial Calibration

