

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111624\
 Data File : BF140417.D
 Acq On : 16 Nov 2024 09:34
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 18 00:04:05 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 14:40:06 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	117	0.00
2	1,4-Dioxane	40.000	38.608	3.5	114	0.01
3	Pyridine	40.000	38.903	2.7	111	0.01
4	n-Nitrosodimethylamine	40.000	38.679	3.3	111	0.02
5 S	2-Fluorophenol	80.000	78.454	1.9	116	0.01
6	Aniline	40.000	34.173	14.6	95	0.00
7 S	Phenol-d6	80.000	77.144	3.6	111	0.02
8	2-Chlorophenol	40.000	39.489	1.3	115	0.00
9	Benzaldehyde	40.000	33.734	15.7	105	0.00
10 C	Phenol	40.000	37.755	5.6	107	0.01
11	bis(2-Chloroethyl)ether	40.000	39.048	2.4	114	0.00
12	1,3-Dichlorobenzene	40.000	40.148	-0.4	118	0.00
13 C	1,4-Dichlorobenzene	40.000	39.091	2.3	115	0.00
14	1,2-Dichlorobenzene	40.000	39.548	1.1	116	0.00
15	Benzyl Alcohol	40.000	39.455	1.4	109	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.798	8.0	103	0.00
17	2-Methylphenol	40.000	38.591	3.5	109	0.01
18	Hexachloroethane	40.000	39.154	2.1	115	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.372	6.6	106	0.00
20	3+4-Methylphenols	40.000	37.942	5.1	106	0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	109	0.00
22	Acetophenone	40.000	39.001	2.5	110	0.00
23 S	Nitrobenzene-d5	80.000	78.395	2.0	109	0.00
24	Nitrobenzene	40.000	39.081	2.3	108	0.00
25	Isophorone	40.000	38.928	2.7	107	0.00
26 C	2-Nitrophenol	40.000	41.139	-2.8	111	0.00
27	2,4-Dimethylphenol	40.000	40.141	-0.4	111	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.459	1.4	110	0.00
29 C	2,4-Dichlorophenol	40.000	39.458	1.4	109	0.00
30	1,2,4-Trichlorobenzene	40.000	39.752	0.6	112	0.00
31	Naphthalene	40.000	39.453	1.4	111	0.00
32	Benzoic acid	40.000	45.304	-13.3	116	0.02
33	4-Chloroaniline	40.000	34.870	12.8	98	0.00
34 C	Hexachlorobutadiene	40.000	39.687	0.8	111	0.00
35	Caprolactam	40.000	41.008	-2.5	113	0.00
36 C	4-Chloro-3-methylphenol	40.000	39.981	0.0	109	0.01
37	2-Methylnaphthalene	40.000	39.301	1.7	110	0.00
38	1-Methylnaphthalene	40.000	39.279	1.8	109	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	109	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	39.350	1.6	112	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.835	7.9	102	0.00
42 S	2,4,6-Tribromophenol	80.000	79.764	0.3	112	0.00
43 C	2,4,6-Trichlorophenol	40.000	39.230	1.9	108	0.00
44	2,4,5-Trichlorophenol	40.000	39.696	0.8	111	0.01
45 S	2-Fluorobiphenyl	80.000	76.179	4.8	110	0.00
46	1,1'-Biphenyl	40.000	38.433	3.9	109	0.00
47	2-Chloronaphthalene	40.000	38.553	3.6	109	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111624\
 Data File : BF140417.D
 Acq On : 16 Nov 2024 09:34
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 18 00:04:05 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 14:40:06 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	39.329	1.7	108	0.00
49	Acenaphthylene	40.000	38.643	3.4	108	0.00
50	Dimethylphthalate	40.000	39.001	2.5	110	0.00
51	2,6-Dinitrotoluene	40.000	39.596	1.0	108	0.00
52 C	Acenaphthene	40.000	38.745	3.1	109	0.00
53	3-Nitroaniline	40.000	38.075	4.8	104	0.00
54 P	2,4-Dinitrophenol	40.000	42.441	-6.1	120	0.00
55	Dibenzofuran	40.000	38.569	3.6	108	0.00
56 P	4-Nitrophenol	40.000	38.806	3.0	102	0.02
57	2,4-Dinitrotoluene	40.000	40.012	-0.0	110	0.00
58	Fluorene	40.000	38.581	3.5	111	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.529	-1.3	109	0.00
60	Diethylphthalate	40.000	39.351	1.6	111	0.00
61	4-Chlorophenyl-phenylether	40.000	38.886	2.8	110	0.00
62	4-Nitroaniline	40.000	40.025	-0.1	109	0.00
63	Azobenzene	40.000	38.024	4.9	106	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	109	0.00
65	4,6-Dinitro-2-methylphenol	40.000	43.912	-9.8	119	0.00
66 c	n-Nitrosodiphenylamine	40.000	38.802	3.0	110	0.00
67	4-Bromophenyl-phenylether	40.000	39.654	0.9	112	0.00
68	Hexachlorobenzene	40.000	39.812	0.5	112	0.00
69	Atrazine	40.000	35.023	12.4	116	0.00
70 C	Pentachlorophenol	40.000	40.271	-0.7	105	0.00
71	Phenanthrene	40.000	38.224	4.4	109	0.00
72	Anthracene	40.000	38.679	3.3	110	0.00
73	Carbazole	40.000	38.618	3.5	110	0.00
74	Di-n-butylphthalate	40.000	39.934	0.2	111	0.00
75 C	Fluoranthene	40.000	38.456	3.9	109	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	100	0.00
77	Benzidine	40.000	27.139	32.2#	80	0.00
78	Pyrene	40.000	42.188	-5.5	108	0.00
79 S	Terphenyl-d14	80.000	82.901	-3.6	107	0.00
80	Butylbenzylphthalate	40.000	44.155	-10.4	106	0.00
81	Benzo(a)anthracene	40.000	40.455	-1.1	106	0.00
82	3,3'-Dichlorobenzidine	40.000	37.880	5.3	99	0.00
83	Chrysene	40.000	37.339	6.7	96	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	43.706	-9.3	104	0.00
85 c	Di-n-octyl phthalate	40.000	41.261	-3.2	99	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	104	-0.01
87	Indeno(1,2,3-cd)pyrene	40.000	40.946	-2.4	105	0.00
88	Benzo(b)fluoranthene	40.000	36.971	7.6	100	0.00
89	Benzo(k)fluoranthene	40.000	39.773	0.6	105	-0.01
90 C	Benzo(a)pyrene	40.000	39.340	1.6	104	-0.01
91	Dibenzo(a,h)anthracene	40.000	41.141	-2.9	106	0.00
92	Benzo(g,h,i)perylene	40.000	40.399	-1.0	104	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111624\
Data File : BF140417.D
Acq On : 16 Nov 2024 09:34
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 18 00:04:05 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 13 14:40:06 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0