

Data Path : Z:\HPCHEM1\BNA F\DATA\BF111717\
 Data File : BF100668.D
 Acq On : 17 Nov 2017 11:14
 Operator : SJ/JU
 Sample : I6479-01
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 T-20

Manual Integrations
 APPROVED

SOHIL
 11/17/2017 5:17:13 PM

Quant Time: Nov 17 16:10:27 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 17 15:22:08 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.88	152	86951	20.00	ng	0.00
21) Naphthalene-d8	8.16	136	307736	20.00	ng	0.00
38) Acenaphthene-d10	9.95	164	69385	20.00	ng	0.02
63) Phenanthrene-d10	11.49	188	37972	20.00	ng	0.08
75) Chrysene-d12	14.09	240	123110	20.00	ng	0.05
86) Perylene-d12	15.54	264	218228	20.00	ng	0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	424257	78.21	ng	0.00
7) Phenol-d6	6.50	99	513182	76.72	ng	0.00
23) Nitrobenzene-d5	7.44	82	303151	65.13	ng	-0.01
41) 2,4,6-Tribromophenol	10.76	330	41082	58.10	ng	0.05
44) 2-Fluorobiphenyl	9.24	172	509488	111.81	ng	0.00
78) Terphenyl-d14	13.07	244	110394	20.60	ng	0.09
Target Compounds						
49) Dimethylphthalate	9.64	163	65098m	12.288	ng	Qvalue
80) Benzo(a)anthracene	14.08	228	23793	3.209	ng	# 66
82) Chrysene	14.12	228	19543	2.722	ng	# 64
83) Bis(2-ethylhexyl)phthalate	14.04	149	21701	4.610	ng	# 69
85) Indeno(1,2,3-cd)pyrene	17.00	276	13815	2.379	ng	# 88
87) Benzo(b)fluoranthene	15.11	252	48703m	3.767	ng	#
89) Benzo(a)pyrene	15.47	252	25582	2.232	ng	# 56

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF111717\
Data File : BF100668.D
Acq On : 17 Nov 2017 11:14
Operator : SJ/JU
Sample : I6479-01
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampled :
T-20

Manual Integrations
APPROVED

SOHIL
11/17/2017 5:17:13 PM

Quant Time: Nov 17 16:10:27 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110817.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 17 15:22:08 2017
Response via : Initial Calibration

