


Data Path : Z:\HPCHEM1\BNA F\Data\BF111717\
 Data File : BF100705.D
 Acq On : 18 Nov 2017 6:24
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF110817.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Nov 17 15:22:08 2017



AutoFind: Scans 1625, 1626, 1627; Background Corrected with Scan 1620

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.4	13505	PASS
68	69	0.00	2	1.8	214	PASS
69	198	0.00	100	29.1	11764	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	41.3	16718	PASS
197	198	0.00	2	0.6	250	PASS
198	198	100	100	100.0	40439	PASS
199	198	5	9	6.7	2727	PASS
275	198	10	60	25.3	10211	PASS
365	198	1	100	2.7	1099	PASS
441	443	0.01	100	88.7	6585	PASS
442	198	50	100	97.8	39546	PASS
443	442	15	24	18.8	7420	PASS