


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF111815\
 Data File : BF083005.D
 Acq On : 18 Nov 2015 12:28
 Operator : UM/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF110715.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Nov 18 13:32:32 2015



AutoFind: Scans 825, 826, 827; Background Corrected with Scan 822

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.5	52887	PASS
68	69	0.00	2	1.6	865	PASS
69	198	0.00	100	38.4	52746	PASS
70	69	0.00	2	0.4	191	PASS
127	198	10	80	47.5	65187	PASS
197	198	0.00	2	0.4	540	PASS
198	198	100	100	100.0	137352	PASS
199	198	5	9	6.5	8900	PASS
275	198	10	60	22.6	31083	PASS
365	198	1	100	2.2	3042	PASS
441	442	0.01	24	15.4	16101	PASS
442	198	50	100	76.2	104702	PASS
443	442	15	24	19.4	20287	PASS