

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111822\
 Data File : BF131306.D
 Acq On : 18 Nov 2022 19:49
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 21 00:41:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 18 01:41:28 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	132	0.00
2	1,4-Dioxane	40.000	40.492	-1.2	132	0.00
3	Pyridine	40.000	40.011	-0.0	130	0.00
4	n-Nitrosodimethylamine	40.000	38.460	3.8	124	0.00
5 S	2-Fluorophenol	80.000	79.578	0.5	128	0.00
6	Aniline	40.000	39.002	2.5	128	0.00
7 S	Phenol-d6	80.000	77.960	2.6	127	0.00
8	2-Chlorophenol	40.000	40.984	-2.5	131	0.00
9	Benzaldehyde	40.000	35.964	10.1	126	0.00
10 C	Phenol	40.000	40.166	-0.4	131	0.00
11	bis(2-Chloroethyl)ether	40.000	39.402	1.5	130	0.00
12	1,3-Dichlorobenzene	40.000	38.922	2.7	128	0.00
13 C	1,4-Dichlorobenzene	40.000	38.737	3.2	129	0.00
14	1,2-Dichlorobenzene	40.000	38.068	4.8	126	0.00
15	Benzyl Alcohol	40.000	38.979	2.6	123	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.793	0.5	131	0.00
17	2-Methylphenol	40.000	39.227	1.9	129	0.00
18	Hexachloroethane	40.000	40.808	-2.0	131	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.191	4.5	126	0.00
20	3+4-Methylphenols	40.000	39.637	0.9	128	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	128	0.00
22	Acetophenone	40.000	38.533	3.7	127	0.00
23 S	Nitrobenzene-d5	80.000	86.757	-8.4	135	0.00
24	Nitrobenzene	40.000	45.047	-12.6	133	0.00
25	Isophorone	40.000	40.162	-0.4	129	0.00
26 C	2-Nitrophenol	40.000	43.932	-9.8	157	0.00
27	2,4-Dimethylphenol	40.000	42.767	-6.9	136	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.252	-0.6	131	0.00
29 C	2,4-Dichlorophenol	40.000	42.394	-6.0	129	0.00
30	1,2,4-Trichlorobenzene	40.000	39.238	1.9	126	0.00
31	Naphthalene	40.000	38.937	2.7	127	0.00
32	Benzoic acid	40.000	44.201	-10.5	146	0.00
33	4-Chloroaniline	40.000	39.606	1.0	126	0.00
34 C	Hexachlorobutadiene	40.000	38.739	3.2	124	0.00
35	Caprolactam	40.000	42.304	-5.8	128	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.758	-1.9	126	0.00
37	2-Methylnaphthalene	40.000	38.810	3.0	126	0.00
38	1-Methylnaphthalene	40.000	38.657	3.4	127	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	128	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.787	3.0	126	0.00
41 P	Hexachlorocyclopentadiene	40.000	44.827	-12.1	137	0.00
42 S	2,4,6-Tribromophenol	80.000	84.280	-5.4	122	0.00
43 C	2,4,6-Trichlorophenol	40.000	41.254	-3.1	123	0.00
44	2,4,5-Trichlorophenol	40.000	41.773	-4.4	126	0.00
45 S	2-Fluorobiphenyl	80.000	74.517	6.9	123	0.00
46	1,1'-Biphenyl	40.000	38.534	3.7	127	0.00
47	2-Chloronaphthalene	40.000	38.716	3.2	125	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111822\
 Data File : BF131306.D
 Acq On : 18 Nov 2022 19:49
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 21 00:41:39 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111722.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 18 01:41:28 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.205	-5.5	138	0.00
49	Acenaphthylene	40.000	38.524	3.7	125	0.00
50	Dimethylphthalate	40.000	39.451	1.4	128	0.00
51	2,6-Dinitrotoluene	40.000	42.342	-5.9	137	0.00
52 C	Acenaphthene	40.000	38.701	3.2	126	0.00
53	3-Nitroaniline	40.000	41.953	-4.9	135	0.00
54 P	2,4-Dinitrophenol	40.000	45.914	-14.8	160	0.00
55	Dibenzofuran	40.000	38.113	4.7	124	0.00
56 P	4-Nitrophenol	40.000	42.053	-5.1	139	0.00
57	2,4-Dinitrotoluene	40.000	42.871	-7.2	144	0.00
58	Fluorene	40.000	37.236	6.9	123	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	43.076	-7.7	126	0.00
60	Diethylphthalate	40.000	39.487	1.3	127	0.00
61	4-Chlorophenyl-phenylether	40.000	38.260	4.4	125	0.00
62	4-Nitroaniline	40.000	42.126	-5.3	134	0.00
63	Azobenzene	40.000	38.653	3.4	125	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	124	0.00
65	4,6-Dinitro-2-methylphenol	40.000	46.354	-15.9	151	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.672	0.8	125	0.00
67	4-Bromophenyl-phenylether	40.000	41.375	-3.4	126	0.00
68	Hexachlorobenzene	40.000	40.389	-1.0	124	0.00
69	Atrazine	40.000	41.165	-2.9	126	0.00
70 C	Pentachlorophenol	40.000	44.006	-10.0	127	0.00
71	Phenanthrene	40.000	38.770	3.1	121	0.00
72	Anthracene	40.000	38.805	3.0	120	0.00
73	Carbazole	40.000	39.444	1.4	123	0.00
74	Di-n-butylphthalate	40.000	42.222	-5.6	129	0.00
75 C	Fluoranthene	40.000	38.414	4.0	122	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	128	0.00
77	Benzydine	40.000	43.304	-8.3	117	0.00
78	Pyrene	40.000	39.578	1.1	121	0.00
79 S	Terphenyl-d14	80.000	78.253	2.2	121	0.00
80	Butylbenzylphthalate	40.000	41.963	-4.9	136	0.00
81	Benzo(a)anthracene	40.000	39.414	1.5	129	0.00
82	3,3'-Dichlorobenzidine	40.000	44.888	-12.2	139	0.00
83	Chrysene	40.000	39.573	1.1	128	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	44.303	-10.8	149	0.00
85 c	Di-n-octyl phthalate	40.000	50.162	-25.4#	174	0.00
86 I	Perylene-d12	20.000	20.000	0.0	144	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.015	-0.0	141	0.00
88	Benzo(b)fluoranthene	40.000	39.640	0.9	143	0.00
89	Benzo(k)fluoranthene	40.000	38.422	3.9	140	0.00
90 C	Benzo(a)pyrene	40.000	40.986	-2.5	144	0.00
91	Dibenzo(a,h)anthracene	40.000	40.483	-1.2	143	0.00
92	Benzo(g,h,i)perylene	40.000	39.917	0.2	140	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF111822\
Data File : BF131306.D
Acq On : 18 Nov 2022 19:49
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 21 00:41:39 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF111722.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 18 01:41:28 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 1