


Data Path : Z:\HPCHEM1\BNA_F\Data\BF112014\
 Data File : BF075701.D
 Acq On : 20 Nov 2014 9:49
 Operator : TP / JJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF111914.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Nov 19 18:09:59 2014



AutoFind: Scans 1081, 1082, 1083; Background Corrected with Scan 1078

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	57.1	10549	PASS
68	69	0.00	2	0.0	0	PASS
69	198	0.00	100	39.3	7256	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.1	8893	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	18485	PASS
199	198	5	9	7.4	1360	PASS
275	198	10	60	20.7	3834	PASS
365	198	1	100	2.4	447	PASS
441	442	0.01	24	16.9	2828	PASS
442	198	50	100	90.5	16723	PASS
443	442	15	24	18.8	3146	PASS