

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF112014\
 Data File : BF075738.D
 Acq On : 21 Nov 2014 6:33
 Operator : TP / JJ
 Sample : F4761-05
 Misc :
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 P001-WOODBE001-111314-001

Quant Time: Nov 21 07:35:41 2014
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF111914.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 19 18:09:59 2014
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.48	152	64104	20.00	ng	-0.01
21) Naphthalene-d8	9.05	136	281947	20.00	ng	-0.01
38) Acenaphthene-d10	11.23	164	138331	20.00	ng	-0.02
63) Phenanthrene-d10	13.08	188	241430	20.00	ng	-0.01
75) Chrysene-d12	16.37	240	252012	20.00	ng	-0.02
86) Perylene-d12	18.14	264	251033	20.00	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.77	112	337987	81.91	ng	-0.01
7) Phenol-d6	7.03	99	349295	69.61	ng	-0.01
23) Nitrobenzene-d5	8.16	82	259065	56.40	ng	-0.01
41) 2,4,6-Tribromophenol	12.22	330	113188	100.36	ng	-0.01
44) 2-Fluorobiphenyl	10.40	172	518239	58.65	ng	-0.01
78) Terphenyl-d14	15.07	244	566640	56.24	ng	-0.02
Target Compounds						
9) Benzaldehyde	6.92	77	9929	2.57	ng	Qvalue 87
20) 3+4-Methylphenols	7.99	107	12944	2.56	ng	# 53
31) Naphthalene	9.08	128	29627	2.06	ng	98
32) Benzoic acid	8.71	122	8601	3.48	ng	# 77
83) Bis(2-ethylhexyl)phthalate	16.41	149	277521	28.02	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF112014\
Data File : BF075738.D
Acq On : 21 Nov 2014 6:33
Operator : TP / JJ
Sample : F4761-05
Misc :
ALS Vial : 35 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
P001-WOODBE001-111314-001

Quant Time: Nov 21 07:35:41 2014
Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF111914.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 19 18:09:59 2014
Response via : Initial Calibration

