

Data Path : V:\HPCHEM1\BNA F\DATA\BF112015\
 Data File : BF083076.D
 Acq On : 20 Nov 2015 12:21
 Operator : UM/IZ
 Sample : G4436-09
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 OS-1601

Quant Time: Nov 20 16:08:01 2015
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF111915.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 20 01:33:39 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.70	152	147656	20.00	ng	0.00
21) Naphthalene-d8	7.98	136	549200	20.00	ng	-0.01
38) Acenaphthene-d10	9.74	164	289490	20.00	ng	0.00
63) Phenanthrene-d10	11.22	188	458913	20.00	ng	0.00
75) Chrysene-d12	13.86	240	372361	20.00	ng	0.00
86) Perylene-d12	15.25	264	333524	20.00	ng	0.01

System Monitoring Compounds						
5) 2-Fluorophenol	5.30	112	857339	90.18	ng	0.01
7) Phenol-d6	6.36	99	1440106	115.02	ng	0.01
23) Nitrobenzene-d5	7.26	82	786932	65.60	ng	-0.01
41) 2,4,6-Tribromophenol	10.53	330	218694	71.15	ng	0.00
44) 2-Fluorobiphenyl	9.07	172	1210800	56.33	ng	-0.01
78) Terphenyl-d14	12.81	244	898817	56.98	ng	0.00


Target Compounds						Qvalue
31) Naphthalene	8.01	128	137691	4.81	ng	98
37) 2-Methylnaphthalene	8.70	142	67372	3.54	ng	96
48) Acenaphthylene	9.61	152	213725	6.70	ng	99
49) Dimethylphthalate	9.47	163	434633	18.24	ng	100
51) Acenaphthene	9.78	154	85857	4.26	ng	99
54) Dibenzofuran	9.95	168	113561	4.20	ng	90
57) Fluorene	10.28	166	117590	5.41	ng	# 97
70) Phenanthrene	11.25	178	1527959	55.19	ng	98
71) Anthracene	11.30	178	485751	17.42	ng	97
72) Carbazole	11.46	167	208798	8.56	ng	97
74) Fluoranthene	12.44	202	2252585	78.06	ng	97
77) Pyrene	12.67	202	2001040	76.33	ng	100
80) Benzo(a)anthracene	13.85	228	1238541m	52.34	ng	
82) Chrysene	13.89	228	922583	42.52	ng	97
85) Indeno(1,2,3-cd)pyrene	16.56	276	401467	22.74	ng	98
87) Benzo(b)fluoranthene	14.88	252	1008731m	43.25	ng	
88) Benzo(k)fluoranthene	14.88	252	497524m	28.01	ng	
89) Benzo(a)pyrene	15.20	252	685555	36.63	ng	97
90) Dibenzo(a,h)anthracene	16.56	278	117280m	7.20	ng	
91) Benzo(a,h,i)perylene	16.95	276	348559	21.94	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : V:\HPCHEM1\BNA F\DATA\BF112015\
Data File : BF083076.D
Acq On : 20 Nov 2015 12:21
Operator : UM/IZ
Sample : G4436-09
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
OS-1601

Quant Time: Nov 20 16:08:01 2015
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF111915.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 20 01:33:39 2015
Response via : Initial Calibration






#1




1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.70 min Scan# 404
Delta R.T. 0.00 min
Lab File: BF083076.D
Acq: 20 Nov 2015 12:21

Instrument :
BNA_F
ClientSampleId :
OS-1601

Tot Ion	Ratio	Lower	Upper
152	100		
150	150.3	126.4	189.6
115	55.8	49.0	73.6




Abundance Ion 152.00 (151.70 to 152.70): E
Ion 150.00 (149.70 to 150.70): E
Ion 115.00 (114.70 to 115.70): E





#5


2-Fluorophenol
Concen: 90.18 ng
RT: 5.30 min Scan# 281
Delta R.T. 0.01 min
Lab File: BF083076.D
Acq: 20 Nov 2015 12:21

Tgt Ion	Ratio	Lower	Upper
112	100		
64	70.2	54.8	82.2
63	36.1	29.0	43.4



Abundance Ion 112.00 (111.70 to 112.70): E
Ion 64.00 (63.70 to 64.70): BF0
Ion 63.00 (62.70 to 63.70): BF0





#7

Phenol-d6

Concen: 115.02 ng

RT: 6.36 min Scan# 374

Delta R.T. 0.01 min

Lab File: BF083076.D

Acq: 20 Nov 2015 12:21

Instrument :

BNA_F

ClientSampleId :

OS-1601


Tot Ion: 99 Resp: 1440106

Ion Ratio Lower Upper

99 100

42 15.9 16.6 25.0#




71 33.5 29.2 43.8



Abundance Ion 99.00 (98.70 to 99.70): BF0

Ion 42.00 (41.70 to 42.70): BF0

Ion 71.00 (70.70 to 71.70): BF0



#21

Naphthalene-d8

Concen: 20.00 ng

RT: 7.98 min Scan# 516

Delta R.T. -0.01 min

Lab File: BF083076.D

Acq: 20 Nov 2015 12:21

Tgt Ion: 136 Resp: 549200


Ion Ratio Lower Upper

136 100

137 11.2 8.6 12.8

54 11.2 6.2 9.2#

68 7.5 4.2 6.2#






Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BF0

Ion 68.00 (67.70 to 68.70): BF0





#23

Nitrobenzene-d5

Concen: 65.60 ng

RT: 7.26 min Scan# 453

Delta R.T. -0.01 min

Lab File: BF083076.D

Acq: 20 Nov 2015 12:21

Instrument :

BNA_F

ClientSampleId :

OS-1601


Tot Ion: 82 Resp: 786932

Ion Ratio Lower Upper

82 100

128 34.8 33.1 49.7




54 52.3 39.8 59.6



Abundance Ion 82.00 (81.70 to 82.70): BF0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BF0



#31

Naphthalene

Concen: 4.81 ng

RT: 8.01 min Scan# 518

Delta R.T. -0.01 min

Lab File: BF083076.D

Acq: 20 Nov 2015 12:21


Tgt Ion: 128 Resp: 137691

Ion Ratio Lower Upper

128 100

129 10.9 9.0 13.4



127 12.7 10.9 16.3




Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E









#37
 2-Methylnaphthalene
 Concen: 3.54 ng
 RT: 8.70 min Scan# 579
 Delta R.T. 0.00 min
 Lab File: BF083076.D
 Acq: 20 Nov 2015 12:21




Instrument :
 BNA_F
 ClientSampled :
 OS-1601


Tot Ion:142 Resp: 67372
 Ion Ratio Lower Upper
 142 100
 141 85.9 70.7 106.1
 115 37.2 33.1 49.7



#38
 Acenaphthene-d10
 Concen: 20.00 ng
 RT: 9.74 min Scan# 670
 Delta R.T. 0.00 min
 Lab File: BF083076.D
 Acq: 20 Nov 2015 12:21

Tgt Ion:164 Resp: 289490
 Ion Ratio Lower Upper
 164 100
 162 102.3 80.5 120.7
 160 45.5 35.0 52.4









#41
 2,4,6-Tribromophenol
 Concen: 71.15 ng
 RT: 10.53 min Scan# 739
 Delta R.T. 0.00 min
 Lab File: BF083076.D
 Acq: 20 Nov 2015 12:21

Instrument :
 BNA_F
 ClientSampleId :
 OS-1601

Tot Ion:330 Resp: 218694
 Ion Ratio Lower Upper
 330 100
 332 97.3 77.6 116.4
 141 40.0 34.6 52.0




Abundance Ion 329.80 (329.50 to 330.50): E
 Ion 331.80 (331.50 to 332.50): E
 Ion 141.00 (140.70 to 141.70): E






#44
 2-Fluorobiphenyl
 Concen: 56.33 ng
 RT: 9.07 min Scan# 611
 Delta R.T. -0.01 min
 Lab File: BF083076.D
 Acq: 20 Nov 2015 12:21

Tgt Ion:172 Resp: 1210800
 Ion Ratio Lower Upper
 172 100
 171 35.7 28.2 42.4
 170 24.1 19.4 29.0



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#48

Acenaphthylene

Concen: 6.70 ng

RT: 9.61 min Scan# 658

Delta R.T. 0.00 min

Lab File: BF083076.D

Acq: 20 Nov 2015 12:21

Instrument :

BNA_F

ClientSampleId :

OS-1601


Tot Ion:152 Resp: 213725

Ion Ratio Lower Upper

152 100

151 19.4 16.2 24.2




153 13.4 10.6 15.8



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 151.00 (150.70 to 151.70): E

Ion 153.00 (152.70 to 153.70): E



#49

Dimethylphthalate

Concen: 18.24 ng

RT: 9.47 min Scan# 646

Delta R.T. -0.01 min

Lab File: BF083076.D

Acq: 20 Nov 2015 12:21


Tgt Ion:163 Resp: 434633

Ion Ratio Lower Upper

163 100

194 3.3 2.7 4.1



164 10.4 8.5 12.7




Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E





#51

Acenaphthene

Concen: 4.26 ng

RT: 9.78 min Scan# 673

Delta R.T. 0.00 min

Lab File: BF083076.D

Acq: 20 Nov 2015 12:21

Instrument :

BNA_F

ClientSampleId :

OS-1601


Tot Ion:154 Resp: 85857

Ion Ratio Lower Upper

154 100

153 113.7 90.1 135.1




152 54.2 43.3 64.9



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#54

Dibenzofuran

Concen: 4.20 ng

RT: 9.95 min Scan# 688

Delta R.T. 0.00 min

Lab File: BF083076.D

Acq: 20 Nov 2015 12:21


Tgt Ion:168 Resp: 113561

Ion Ratio Lower Upper

168 100

139 34.2 33.5 50.3



169 16.0 11.2 16.8




Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 169.00 (168.70 to 169.70): E









#57
 Fluorene
 Concen: 5.41 ng
 RT: 10.28 min Scan# 717
 Delta R.T. -0.01 min
 Lab File: BF083076.D
 Acq: 20 Nov 2015 12:21




Instrument :
 BNA_F
 ClientSampleId :
 OS-1601


Tot Ion:166 Resp: 117590
 Ion Ratio Lower Upper
 166 100
 165 94.5 77.1 115.7
 167 18.8 11.0 16.4#



#63
 Phenanthrene-d10
 Concen: 20.00 ng
 RT: 11.22 min Scan# 799
 Delta R.T. 0.00 min
 Lab File: BF083076.D
 Acq: 20 Nov 2015 12:21

Tgt Ion:188 Resp: 458913
 Ion Ratio Lower Upper
 188 100
 94 10.9 9.0 13.4
 80 11.7 10.2 15.4









#70
 Phenanthrene
 Concen: 55.19 ng
 RT: 11.25 min Scan# 802
 Delta R.T. 0.00 min
 Lab File: BF083076.D
 Acq: 20 Nov 2015 12:21




Instrument :
 BNA_F
 ClientSampleId :
 OS-1601


Tot Ion:178 Resp: 1527959
 Ion Ratio Lower Upper
 178 100
 176 20.2 15.6 23.4
 179 16.0 12.3 18.5



#71
 Anthracene
 Concen: 17.42 ng
 RT: 11.30 min Scan# 806
 Delta R.T. 0.00 min
 Lab File: BF083076.D
 Acq: 20 Nov 2015 12:21

Tgt Ion:178 Resp: 485751
 Ion Ratio Lower Upper
 178 100
 176 18.3 15.8 23.6
 179 15.3 13.1 19.7









#72
 Carbazole
 Concen: 8.56 ng
 RT: 11.46 min Scan# 820
 Delta R.T. 0.00 min
 Lab File: BF083076.D
 Acq: 20 Nov 2015 12:21




Instrument :
 BNA_F
 ClientSampleId :
 OS-1601


Tot Ion	Ratio	Lower	Upper
167	100		
166	21.1	17.8	26.6
139	12.2	11.0	16.4



#74
 Fluoranthene
 Concen: 78.06 ng
 RT: 12.44 min Scan# 906
 Delta R.T. 0.01 min
 Lab File: BF083076.D
 Acq: 20 Nov 2015 12:21

Tgt Ion	Ratio	Lower	Upper
202	100		
101	9.7	0.0	31.1
203	19.2	0.0	38.4





#75

Chrysene-d12

Concen: 20.00 ng

RT: 13.86 min Scan# 1030

Delta R.T. 0.00 min

Lab File: BF083076.D

Acq: 20 Nov 2015 12:21

Instrument :

BNA_F

ClientSampleId :

OS-1601


Tot Ion:240 Resp: 372361

Ion Ratio Lower Upper

240 100

120 11.1 5.3 7.9#




236 25.8 20.4 30.6



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#77

Pyrene

Concen: 76.33 ng

RT: 12.67 min Scan# 926

Delta R.T. 0.01 min

Lab File: BF083076.D

Acq: 20 Nov 2015 12:21


Tgt Ion:202 Resp: 2001040

Ion Ratio Lower Upper

202 100

200 21.8 17.6 26.4



203 18.3 14.6 22.0




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E





#78

Terphenyl-d14

Concen: 56.98 ng

RT: 12.81 min Scan# 938

Delta R.T. 0.00 min

Lab File: BF083076.D

Acq: 20 Nov 2015 12:21

Instrument :

BNA_F

ClientSampleId :

OS-1601


Tot Ion:244 Resp: 898817

Ion Ratio Lower Upper

244 100

212 8.1 6.4 9.6




122 13.7 10.9 16.3



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Benzo(a)anthracene

Concen: 52.34 ng m

RT: 13.85 min Scan# 1029

Delta R.T. 0.00 min

Lab File: BF083076.D

Acq: 20 Nov 2015 12:21


Tgt Ion:228 Resp: 1238541

Ion Ratio Lower Upper

228 100

226 34.6 22.6 34.0#



229 22.1 16.1 24.1




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E





#82

Chrysene

Concen: 42.52 ng

RT: 13.89 min Scan# 1033

Delta R.T. 0.01 min

Lab File: BF083076.D

Acq: 20 Nov 2015 12:21

Instrument :

BNA_F

ClientSampleId :

OS-1601


Tot Ion:228 Resp: 922583

Ion Ratio Lower Upper

228 100

226 31.0 25.3 37.9




229 22.8 16.4 24.6



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#85

Indeno(1,2,3-cd)pyrene

Concen: 22.74 ng

RT: 16.56 min Scan# 1266

Delta R.T. 0.01 min

Lab File: BF083076.D

Acq: 20 Nov 2015 12:21


Tgt Ion:276 Resp: 401467

Ion Ratio Lower Upper

276 100

138 18.5 16.3 24.5



277 25.3 19.8 29.8




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#86
Pervlene-d12
Concen: 20.00 ng
RT: 15.25 min Scan# 1152
Delta R.T. 0.01 min
Lab File: BF083076.D
Acq: 20 Nov 2015 12:21

Instrument :
BNA_F
ClientSampleId :
OS-1601

Tot Ion:264 Resp: 333524
Ion Ratio Lower Upper
264 100
260 24.2 18.6 27.8
265 25.5 17.0 25.6




Abundance Ion 264.00 (263.70 to 264.70): E
Ion 260.00 (259.70 to 260.70): E
Ion 265.00 (264.70 to 265.70): E






#87
Benzo(b)fluoranthene
Concen: 43.25 ng m
RT: 14.88 min Scan# 1119
Delta R.T. 0.02 min
Lab File: BF083076.D
Acq: 20 Nov 2015 12:21

Tgt Ion:252 Resp: 1008731
Ion Ratio Lower Upper
252 100
253 22.8 17.6 26.4
125 9.0 7.7 11.5



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E





#88

Benzo(k)fluoranthene

Concen: 28.01 ng/mL

RT: 14.88 min Scan# 1119

Delta R.T. -0.01 min

Lab File: BF083076.D

Acq: 20 Nov 2015 12:21

Instrument :

BNA_F

ClientSampleId :

OS-1601


Tot Ion:252 Resp: 497524

Ion Ratio Lower Upper

252 100

253 22.8 17.8 26.6




125 9.0 3.8 5.6#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#89

Benzo(a)pyrene

Concen: 36.63 ng/mL

RT: 15.20 min Scan# 1147

Delta R.T. 0.01 min

Lab File: BF083076.D

Acq: 20 Nov 2015 12:21


Tgt Ion:252 Resp: 685555

Ion Ratio Lower Upper

252 100

253 23.6 17.5 26.3



125 9.0 6.4 9.6




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#90
Dibenzo(a,h)anthracene
Concen: 7.20 ng/mL
RT: 16.56 min Scan# 1266
Delta R.T. 0.00 min
Lab File: BF083076.D
Acq: 20 Nov 2015 12:21

Instrument :
BNA_F
ClientSampleId :
OS-1601

Tot Ion:278 Resp: 117280
Ion Ratio Lower Upper
278 100
139 15.8 9.8 14.8#
279 28.9 19.6 29.4



#91
Benzo(g,h,i)perylene
Concen: 21.94 ng/mL
RT: 16.95 min Scan# 1300
Delta R.T. 0.01 min
Lab File: BF083076.D
Acq: 20 Nov 2015 12:21

Tgt Ion:276 Resp: 348559
Ion Ratio Lower Upper
276 100
277 24.6 18.6 27.8
138 16.9 12.5 18.7

