

Data Path : Z:\HPCHEM1\BNA F\DATA\BF112017\
 Data File : BF100742.D
 Acq On : 20 Nov 2017 20:46
 Operator : SJ/JU
 Sample : I6521-05
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 TP-1-(0-5)

Manual Integrations
 APPROVED

SOHIL
 11/21/2017 2:15:36 PM

Quant Time: Nov 21 04:45:26 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 17 15:22:08 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	96209	20.00	ng	-0.01
21) Naphthalene-d8	8.15	136	381925	20.00	ng	-0.02
38) Acenaphthene-d10	9.90	164	171306	20.00	ng	-0.02
63) Phenanthrene-d10	11.39	188	298266	20.00	ng	-0.01
75) Chrysene-d12	14.03	240	191233	20.00	ng	-0.02
86) Perylene-d12	15.50	264	220402	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.49	112	702870	117.11	ng	0.00
7) Phenol-d6	6.50	99	864004	116.74	ng	0.00
23) Nitrobenzene-d5	7.43	82	547615	94.80	ng	-0.02
41) 2,4,6-Tribromophenol	10.70	330	208379	119.37	ng	-0.01
44) 2-Fluorobiphenyl	9.23	172	974680	86.64	ng	-0.01
78) Terphenyl-d14	12.97	244	689364	82.83	ng	-0.01
Target Compounds						
10) Phenol	6.51	94	28669	3.690	ng	Qvalue 96
49) Dimethylphthalate	9.62	163	72471	5.541	ng	99
70) Phenanthrene	11.42	178	42536	2.709	ng	98
74) Fluoranthene	12.60	202	96755	5.813	ng	93
77) Pyrene	12.83	202	95092	6.976	ng	98
80) Benzo(a)anthracene	14.02	228	51992	4.515	ng	99
82) Chrysene	14.05	228	48156	4.318	ng	97
83) Bis(2-ethylhexyl)phthalate	14.00	149	18500	2.530	ng	# 93
85) Indeno(1,2,3-cd)pyrene	16.97	276	38956	4.319	ng	92
87) Benzo(b)fluoranthene	15.07	252	69483m	5.321	ng	
89) Benzo(a)pyrene	15.43	252	55586	4.802	ng	96
91) Benzo(a,h,i)perylene	17.42	276	41399	4.467	ng	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF112017\
Data File : BF100742.D
Acq On : 20 Nov 2017 20:46
Operator : SJ/JU
Sample : I6521-05
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
TP-1-(0-5)

Manual Integrations
APPROVED

SOHIL
11/21/2017 2:15:36 PM

Quant Time: Nov 21 04:45:26 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110817.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 17 15:22:08 2017
Response via : Initial Calibration

