

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF112018\
 Data File : BF110886.D
 Acq On : 20 Nov 2018 11:01
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 20 11:32:53 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 08 14:14:47 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	1,4-Dioxane	0.613	0.548	10.6	84	-0.02
3	Pyridine	1.324	1.233	6.9	82	-0.01
4	n-Nitrosodimethylamine	0.568	0.597	-5.1	94	-0.01
5 S	2-Fluorophenol	1.231	1.222	0.7	91	0.00
6	Aniline	2.053	2.026	1.3	94	0.00
7 S	Phenol-d6	1.575	1.591	-1.0	95	0.00
8	2-Chlorophenol	1.443	1.476	-2.3	96	0.00
9	Benzaldehyde	0.893	0.898	-0.6	98	0.00
10 C	Phenol	1.826	1.835	-0.5	94	0.00
11	bis(2-Chloroethyl)ether	1.368	1.353	1.1	94	0.00
12	1,3-Dichlorobenzene	1.579	1.586	-0.4	94	0.00
13 C	1,4-Dichlorobenzene	1.604	1.605	-0.1	95	0.00
14	1,2-Dichlorobenzene	1.492	1.492	0.0	96	-0.01
15	Benzyl Alcohol	1.232	1.265	-2.7	95	0.00
16	2,2'-oxybis(1-Chloropropane	2.149	2.160	-0.5	95	0.00
17	2-Methylphenol	1.149	1.152	-0.3	95	0.00
18	Hexachloroethane	0.592	0.598	-1.0	95	-0.01
19 P	n-Nitroso-di-n-propylamine	1.005	1.042	-3.7	99	0.00
20	3+4-Methylphenols	1.475	1.491	-1.1	96	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
22	Acetophenone	0.466	0.466	0.0	96	0.00
23 S	Nitrobenzene-d5	0.347	0.351	-1.2	97	0.00
24	Nitrobenzene	0.356	0.359	-0.8	97	0.00
25	Isophorone	0.569	0.568	0.2	97	0.00
26 C	2-Nitrophenol	0.178	0.183	-2.8	96	0.00
27	2,4-Dimethylphenol	0.270	0.272	-0.7	97	0.00
28	bis(2-Chloroethoxy)methane	0.385	0.385	0.0	96	0.00
29 C	2,4-Dichlorophenol	0.278	0.288	-3.6	98	0.00
30	1,2,4-Trichlorobenzene	0.314	0.313	0.3	95	0.00
31	Naphthalene	0.939	0.933	0.6	96	0.00
32	Benzoic acid	0.191	0.186	2.6	87	0.00
33	4-Chloroaniline	0.425	0.415	2.4	97	0.00
34 C	Hexachlorobutadiene	0.179	0.181	-1.1	98	0.00
35	Caprolactam	0.093	0.096	-3.2	99	0.00
36 C	4-Chloro-3-methylphenol	0.300	0.308	-2.7	98	0.00
37	2-Methylnaphthalene	0.615	0.606	1.5	95	0.00
38	1-Methylnaphthalene	0.592	0.591	0.2	97	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	96	0.00
40	1,2,4,5-Tetrachlorobenzene	0.633	0.646	-2.1	97	0.00
41 P	Hexachlorocyclopentadiene	0.203	0.182	10.3	81	0.00
42 S	2,4,6-Tribromophenol	0.202	0.202	0.0	95	0.00
43 C	2,4,6-Trichlorophenol	0.398	0.401	-0.8	94	0.00
44	2,4,5-Trichlorophenol	0.424	0.417	1.7	93	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF112018\
 Data File : BF110886.D
 Acq On : 20 Nov 2018 11:01
 Operator : JU/SJ
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 20 11:32:53 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 08 14:14:47 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	1.400	1.383	1.2	96	0.00
46	1,1'-Biphenyl	1.866	1.879	-0.7	97	0.00
47	2-Chloronaphthalene	1.344	1.350	-0.4	97	0.00
48	2-Nitroaniline	0.414	0.437	-5.6	99	0.00
49	Acenaphthylene	1.936	1.907	1.5	94	0.00
50	Dimethylphthalate	1.492	1.478	0.9	96	0.00
51	2,6-Dinitrotoluene	0.328	0.333	-1.5	96	0.00
52 C	Acenaphthene	1.223	1.223	0.0	97	0.00
53	3-Nitroaniline	0.380	0.394	-3.7	99	0.00
54 P	2,4-Dinitrophenol	0.116	0.110	5.2	85	0.00
55	Dibenzofuran	1.790	1.738	2.9	94	0.00
56 P	4-Nitrophenol	0.252	0.207	17.9	75	0.00
57	2,4-Dinitrotoluene	0.427	0.426	0.2	94	0.00
58	Fluorene	1.375	1.380	-0.4	98	0.00
59	2,3,4,6-Tetrachlorophenol	0.334	0.320	4.2	91	0.00
60	Diethylphthalate	1.476	1.497	-1.4	99	0.00
61	4-Chlorophenyl-phenylether	0.712	0.723	-1.5	98	0.00
62	4-Nitroaniline	0.383	0.360	6.0	89	0.00
63	Azobenzene	1.440	1.460	-1.4	97	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	96	0.00
65	4,6-Dinitro-2-methylphenol	0.101	0.093	7.9	86	0.00
66 c	n-Nitrosodiphenylamine	0.674	0.682	-1.2	98	0.00
67	4-Bromophenyl-phenylether	0.221	0.221	0.0	97	0.00
68	Hexachlorobenzene	0.233	0.233	0.0	98	0.00
69	Atrazine	0.206	0.199	3.4	94	0.00
70 C	Pentachlorophenol	0.124	0.101	18.5	76	0.00
71	Phenanthrene	1.071	1.065	0.6	99	0.00
72	Anthracene	1.081	1.075	0.6	97	0.00
73	Carbazole	1.038	1.045	-0.7	98	0.00
74	Di-n-butylphthalate	1.217	1.239	-1.8	98	0.00
75 C	Fluoranthene	1.074	1.063	1.0	97	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
77	Benzidine	0.550	0.486	11.6	91	0.00
78	Pyrene	1.540	1.500	2.6	98	0.00
79 S	Terphenyl-d14	1.068	1.026	3.9	99	0.00
80	Butylbenzylphthalate	0.663	0.688	-3.8	101	0.00
81	Benzo(a)anthracene	1.195	1.185	0.8	99	0.00
82	3,3'-Dichlorobenzidine	0.400	0.410	-2.5	104	0.00
83	Chrysene	1.139	1.137	0.2	103	0.00
84	Bis(2-ethylhexyl)phthalate	0.822	0.876	-6.6	107	0.00
85 c	Di-n-octyl phthalate	1.209	1.334	-10.3	115	0.00
86	Indeno(1,2,3-cd)pyrene	0.817	0.809	1.0	101	0.02
87 I	Perylene-d12	1.000	1.000	0.0	101	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF112018\
Data File : BF110886.D
Acq On : 20 Nov 2018 11:01
Operator : JU/SJ
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 20 11:32:53 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF110818.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 08 14:14:47 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.196	1.166	2.5	96	0.00
89	Benzo(k)fluoranthene	1.182	1.261	-6.7	112	0.00
90 C	Benzo(a)pyrene	1.087	1.102	-1.4	103	0.00
91	Dibenzo(a,h)anthracene	0.920	0.965	-4.9	104	0.01
92	Benzo(q,h,i)perylene	0.913	0.960	-5.1	103	0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0