


Data Path : Z:\SVOASRV\HPCHEM1\BNA_F\DATA\BF112018\
 Data File : BF110885.D
 Acq On : 20 Nov 2018 10:33
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110818.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 13 12:43:10 2018



AutoFind: Scans 1634, 1635, 1636; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	39.8	18615	PASS
68	69	0.00	2	0.8	131	PASS
69	198	0.00	100	36.5	17061	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	49.8	23258	PASS
197	198	0.00	2	0.6	266	PASS
198	198	100	100	100.0	46722	PASS
199	198	5	9	6.7	3120	PASS
275	198	10	60	24.5	11426	PASS
365	198	1	100	2.8	1328	PASS
441	198	0.01	100	14.1	6586	PASS
442	198	50	100	88.9	41551	PASS
443	442	15	24	18.8	7793	PASS