



Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112020\
 Data File : BF122397.D
 Acq On : 20 Nov 2020 11:54
 Operator : JU/CG
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_F\METHODS\8270-BF110920.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Nov 24 14:38:56 2020



AutoFind: Scans 1626, 1627, 1628; Background Corrected with Scan 1621

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.2	54707	PASS
68	69	0.00	2	1.7	933	PASS
69	198	0.00	100	36.4	54925	PASS
70	69	0.00	2	0.6	320	PASS
127	198	10	80	48.8	73659	PASS
197	198	0.00	2	0.1	151	PASS
198	198	100	100	100.0	150963	PASS
199	198	5	9	6.6	10022	PASS
275	198	10	60	28.3	42704	PASS
365	198	1	100	3.2	4827	PASS
441	198	0.01	100	15.8	23907	PASS
442	442	100	100	100.0	153709	PASS
443	442	15	24	19.2	29570	PASS