

Data Path : Z:\HPCHEM1\BNA_F\DATA\BF112115\
 Data File : BF083134.D
 Acq On : 22 Nov 2015 3:03
 Operator : UM/IZ
 Sample : PB86810BS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB86810BS

Quant Time: Nov 22 03:45:46 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF112115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Nov 22 01:05:26 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.65	152	199218	20.00	ng	-0.01
21) Naphthalene-d8	7.94	136	746622	20.00	ng	-0.01
38) Acenaphthene-d10	9.69	164	378911	20.00	ng	-0.01
63) Phenanthrene-d10	11.16	188	739141	20.00	ng	-0.01
75) Chrysene-d12	13.78	240	657332	20.00	ng	-0.02
86) Perylene-d12	15.15	264	572123	20.00	ng	-0.01

System Monitoring Compounds

5) 2-Fluorophenol	5.27	112	204512	7.76	ng	-0.01
7) Phenol-d6	6.32	99	265439	7.78	ng	-0.02
23) Nitrobenzene-d5	7.22	82	163875	5.01	ng	-0.01
41) 2,4,6-Tribromophenol	10.47	330	57642	7.43	ng	-0.02
44) 2-Fluorobiphenyl	9.02	172	316637	5.83	ng	-0.01
78) Terphenyl-d14	12.74	244	322887	5.78	ng	-0.02

Target Compounds

						Qvalue
2) 1,4-Dioxane	2.19	88	32965	5.05	ng	# 90
3) Pyridine	3.05	79	54854	3.18	ng	# 85
4) n-Nitrosodimethylamine	2.86	42	35032	4.41	ng	# 93
6) Aniline	6.33	93	65967	3.17	ng	# 43
8) 2-Chlorophenol	6.44	128	70915	4.98	ng	94
9) Benzaldehyde	6.20	77	43419	3.10	ng	96
10) Phenol	6.33	94	104335	5.06	ng	97
11) bis(2-Chloroethyl)ether	6.39	93	79242	5.38	ng	92
12) 1,3-Dichlorobenzene	6.82	146	74071	4.56	ng	95
13) 1,4-Dichlorobenzene	6.67	146	76630	4.66	ng	93
14) 1,2-Dichlorobenzene	6.82	146	79101	5.29	ng	97
15) Benzyl Alcohol	6.82	79	64571	4.53	ng	97
16) 2,2'-oxybis(1-Chloropropan	6.95	45	131293	5.72	ng	# 66
17) 2-Methylphenol	6.94	107	60190	5.11	ng	95
18) Hexachloroethane	7.16	117	27419	4.73	ng	91
19) n-Nitroso-di-n-propylamine	7.07	70	53074	4.47	ng	# 93
20) 3+4-Methylphenols	7.11	107	77339	5.19	ng	97
22) Acetophenone	7.07	105	105068	4.90	ng	# 98
24) Nitrobenzene	7.24	77	81186	4.65	ng	94
25) Isophorone	7.48	82	136037	4.40	ng	# 93
26) 2-Nitrophenol	7.56	139	31828	4.13	ng	97
27) 2,4-Dimethylphenol	7.62	122	62049	5.04	ng	92
28) bis(2-Chloroethoxy)methane	7.70	93	91741	5.10	ng	99
29) 2,4-Dichlorophenol	7.82	162	59036	5.05	ng	100
30) 1,2,4-Trichlorobenzene	7.88	180	63905	4.97	ng	98
31) Naphthalene	7.96	128	203559	5.05	ng	98
32) Benzoic acid	7.71	122	32163	3.36	ng	91
33) 4-Chloroaniline	8.03	127	53852	3.44	ng	# 89
34) Hexachlorobutadiene	8.09	225	36946	4.86	ng	96
35) Caprolactam	8.39	113	14151	3.77	ng	94
36) 4-Chloro-3-methylphenol	8.52	107	57630	4.25	ng	95
37) 2-Methylnaphthalene	8.65	142	134816	5.16	ng	97
39) 1,2,4,5-Tetrachlorobenzene	8.82	216	60797	4.95	ng	99
40) Hexachlorocyclopentadiene	8.81	237	57132	8.18	ng	99


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF112115\
 Data File : BF083134.D
 Acq On : 22 Nov 2015 3:03
 Operator : UM/IZ
 Sample : PB86810BS
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB86810BS

Quant Time: Nov 22 03:45:46 2015
 Quant Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF112115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Nov 22 01:05:26 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
42) 2,4,6-Trichlorophenol	8.95	196	37279	4.24	ng	94
43) 2,4,5-Trichlorophenol	8.98	196	44133	4.91	ng	# 84
45) 1,1'-Biphenyl	9.12	154	167247	5.19	ng	98
46) 2-Chloronaphthalene	9.14	162	124406	4.91	ng	97
47) 2-Nitroaniline	9.24	65	40139	4.48	ng	88
48) Acenaphthylene	9.54	152	199566	5.08	ng	97
49) Dimethylphthalate	9.43	163	141683	4.86	ng	99
50) 2,6-Dinitrotoluene	9.48	165	28638	4.38	ng	# 64
51) Acenaphthene	9.71	154	119887	5.02	ng	97
52) 3-Nitroaniline	9.66	138	27585	3.95	ng	# 81
53) 2,4-Dinitrophenol	9.76	184	22579	5.99	ng	94
54) Dibenzofuran	9.90	168	187261	5.54	ng	96
55) 4-Nitrophenol	9.84	139	40027	7.48	ng	95
56) 2,4-Dinitrotoluene	9.88	165	38303	4.62	ng	# 80
57) Fluorene	10.23	166	143616	5.14	ng	97
58) 2,3,4,6-Tetrachlorophenol	9.98	232	35898	4.87	ng	90
59) Diethylphthalate	10.12	149	135702	4.67	ng	99
60) 4-Chlorophenyl-phenylether	10.23	204	69922	5.47	ng	96
61) 4-Nitroaniline	10.26	138	28717	4.18	ng	98
62) Azobenzene	10.39	77	178933	5.40	ng	96
64) 4,6-Dinitro-2-methylphenol	10.28	198	18910	3.71	ng	93
65) n-Nitrosodiphenylamine	10.35	169	120927	4.81	ng	97
66) 4-Bromophenyl-phenylether	10.72	248	38463	4.69	ng	95
67) Hexachlorobenzene	10.78	284	47349	5.24	ng	93
68) Atrazine	10.88	200	36919	5.06	ng	94
69) Pentachlorophenol	10.98	266	45045	7.66	ng	94
70) Phenanthrene	11.19	178	227184	5.33	ng	97
71) Anthracene	11.23	178	231530	5.33	ng	96
72) Carbazole	11.40	167	192821	4.99	ng	# 95
73) Di-n-butylphthalate	11.75	149	208426	4.42	ng	# 96
74) Fluoranthene	12.37	202	228240	5.24	ng	99
76) Benzidine	12.50	184	89478	5.19	ng	99
77) Pyrene	12.59	202	239277	5.33	ng	96
79) Butylbenzylphthalate	13.23	149	86678	4.08	ng	# 82
80) Benzo(a)anthracene	13.77	228	206919	5.09	ng	97
81) 3,3'-Dichlorobenzidine	13.75	252	46732	3.30	ng	# 96
82) Chrysene	13.82	228	190097	5.04	ng	95
83) Bis(2-ethylhexyl)phthalate	13.79	149	133758	4.82	ng	100
84) Di-n-octyl phthalate	14.40	149	216050	4.52	ng	98
85) Indeno(1,2,3-cd)pyrene	16.41	276	155208	4.53	ng	96
87) Benzo(b)fluoranthene	14.78	252	349894	8.95	ng	# 98
88) Benzo(k)fluoranthene	14.78	252	349894	12.01	ng	98
89) Benzo(a)pyrene	15.10	252	162115	5.02	ng	# 97
90) Dibenzo(a,h)anthracene	16.42	278	134736	4.60	ng	# 96
91) Benzo(g,h,i)perylene	16.79	276	126234	4.41	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

1,4-Dichlorobenzene-d4

Concen: 20.00 ng

RT: 6.65 min Scan# 408

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS


Tgt Ion:152 Resp: 199218

Ion Ratio Lower Upper

152 100

150 154.0 125.8 188.8




115 62.2 46.9 70.3



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

1,4-Dioxane

Concen: 5.05 ng

RT: 2.19 min Scan# 18

Delta R.T. -0.00 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03


Tgt Ion: 88 Resp: 32965

Ion Ratio Lower Upper

88 100

58 89.2 65.0 97.6



43 38.6 24.7 37.1#




Abundance Ion 88.00 (87.70 to 88.70): BF0

Ion 58.00 (57.70 to 58.70): BF0

Ion 43.00 (42.70 to 43.70): BF0





#3

Pyridine

Concen: 3.18 ng

RT: 3.05 min Scan# 93

Delta R.T. -0.05 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS





Tgt Ion: 79 Resp: 54854

Ion Ratio Lower Upper

79 100

52 71.4 60.7 91.1

51 56.7 27.9 41.9#



#4

n-Nitrosodimethylamine

Concen: 4.41 ng

RT: 2.86 min Scan# 76

Delta R.T. 0.03 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03




Tgt Ion: 42 Resp: 35032




Ion Ratio Lower Upper

42 100

74 123.4 104.8 157.2

44 12.4 6.2 9.4#





#5

2-Fluorophenol

Concen: 7.76 ng

RT: 5.27 min Scan# 287

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS





Tgt Ion: 112 Resp: 204512

Ion Ratio Lower Upper

112 100

64 57.7 58.4 87.6#

63 28.5 30.0 45.0#



#6

Aniline

Concen: 3.17 ng

RT: 6.33 min Scan# 380

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03


Tgt Ion: 93 Resp: 65967


Ion Ratio Lower Upper

93 100

66 84.7 38.7 58.1#

65 53.8 18.6 27.8#





#7

Phenol-d6

Concen: 7.78 ng

RT: 6.32 min Scan# 379

Delta R.T. -0.02 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS





Tgt Ion: 99 Resp: 265439

Ion Ratio Lower Upper

99 100

42 15.7 17.7 26.5#

71 31.1 30.2 45.4



#8

2-Chlorophenol

Concen: 4.98 ng

RT: 6.44 min Scan# 390

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03




Tgt Ion: 128 Resp: 70915


Ion Ratio Lower Upper

128 100

130 32.6 12.6 52.6

64 55.3 41.9 81.9





#9

Benzaldehyde

Concen: 3.10 ng

RT: 6.20 min Scan# 369

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS





Tgt Ion: 77 Resp: 43419

Ion Ratio Lower Upper

77 100

105 88.4 73.4 113.4

106 90.7 67.5 107.5



#10

Phenol

Concen: 5.06 ng

RT: 6.33 min Scan# 380

Delta R.T. -0.02 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03




Tgt Ion: 94 Resp: 104335


Ion Ratio Lower Upper

94 100

65 38.8 21.1 61.1

66 61.1 43.1 83.1









#11
bis(2-Chloroethyl)ether
Concen: 5.38 ng
RT: 6.39 min Scan# 385
Delta R.T. -0.02 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03




Instrument :
BNA_F
ClientSampleId :
PB86810BS


Tgt Ion: 93 Resp: 79242
Ion Ratio Lower Upper
93 100
63 89.4 59.8 99.8
95 32.3 13.0 53.0



#12
1,3-Dichlorobenzene
Concen: 4.56 ng
RT: 6.82 min Scan# 423
Delta R.T. 0.23 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03

Tgt Ion: 146 Resp: 74071
Ion Ratio Lower Upper
146 100
148 63.1 50.2 75.2
75 33.2 32.7 49.1









#13
 1,4-Dichlorobenzene
 Concen: 4.66 ng
 RT: 6.67 min Scan# 410
 Delta R.T. -0.00 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03




Instrument :
 BNA_F
 ClientSampleId :
 PB86810BS


Tgt Ion:146 Resp: 76630
 Ion Ratio Lower Upper
 146 100
 148 65.8 50.2 75.2
 111 40.1 37.5 56.3



#14
 1,2-Dichlorobenzene
 Concen: 5.29 ng
 RT: 6.82 min Scan# 423
 Delta R.T. -0.01 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03

Tgt Ion:146 Resp: 79101
 Ion Ratio Lower Upper
 146 100
 148 63.1 51.8 77.6
 111 42.7 36.2 54.4





#15

Benzyl Alcohol

Concen: 4.53 ng

RT: 6.82 min Scan# 423

Delta R.T. -0.02 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS





Tgt Ion: 79 Resp: 64571

Ion Ratio Lower Upper

79 100

108 72.6 61.4 92.2

77 63.5 50.5 75.7



#16

2,2'-oxybis(1-chloropropane)

Concen: 5.72 ng

RT: 6.95 min Scan# 434

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03




Tgt Ion: 45 Resp: 131293


Ion Ratio Lower Upper

45 100

77 24.3 26.7 66.7#

79 21.5 24.3 64.3#









#17
2-Methylphenol
Concen: 5.11 ng
RT: 6.94 min Scan# 433
Delta R.T. -0.02 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03




Instrument :
BNA_F
ClientSampleId :
PB86810BS







Tgt Ion:107 Resp: 60190
Ion Ratio Lower Upper
107 100
108 116.1 93.8 140.6
77 66.6 57.4 86.0
79 60.8 54.4 81.6



#18
Hexachloroethane
Concen: 4.73 ng
RT: 7.16 min Scan# 453
Delta R.T. -0.01 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03

Tgt Ion:117 Resp: 27419
Ion Ratio Lower Upper
117 100
119 96.1 76.8 115.2
201 93.2 88.6 133.0






#19
n-Nitroso-di-n-propylamine
Concen: 4.47 ng
RT: 7.07 min Scan# 445
Delta R.T. -0.03 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03


Instrument :
BNA_F
ClientSampleId :
PB86810BS


Tgt Ion: 70 Resp: 53074
Ion Ratio Lower Upper
70 100
42 74.6 54.5 81.7
101 7.0 6.6 9.8
130 11.5 11.5 17.3#



#20
3+4-Methylphenols
Concen: 5.19 ng
RT: 7.11 min Scan# 448
Delta R.T. -0.01 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03

Tgt Ion: 107 Resp: 77339
Ion Ratio Lower Upper
107 100
108 91.1 73.8 113.8
77 38.5 19.1 59.1
79 31.6 13.7 53.7









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 7.94 min Scan# 521
Delta R.T. -0.01 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03

Instrument :
BNA_F
ClientSampleId :
PB86810BS

Tgt Ion:	136	Resp:	746622
Ion	Ratio	Lower	Upper
136	100		
137	11.2	8.7	13.1
54	7.9	5.8	8.6
68	6.0	3.8	5.8#




Abundance Ion 136.00 (135.70 to 136.70): E
Ion 137.00 (136.70 to 137.70): E
Ion 54.00 (53.70 to 54.70): BF0
Ion 68.00 (67.70 to 68.70): BF0






#22
Acetophenone
Concen: 4.90 ng
RT: 7.07 min Scan# 445
Delta R.T. -0.02 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03

Tgt Ion:	105	Resp:	105068
Ion	Ratio	Lower	Upper
105	100		
71	7.0	1.0	1.4#
51	30.5	25.6	38.4
120	19.2	15.8	23.6



Abundance Ion 105.00 (104.70 to 105.70): E
Ion 71.00 (70.70 to 71.70): BF0
Ion 51.00 (50.70 to 51.70): BF0
Ion 120.00 (119.70 to 120.70): E









#23
 Nitrobenzene-d5
 Concen: 5.01 ng
 RT: 7.22 min Scan# 458
 Delta R.T. -0.01 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03




Instrument :
 BNA_F
 ClientSampleId :
 PB86810BS


Tgt Ion	Ratio	Lower	Upper
82	100		
128	41.1	28.5	42.7
54	49.1	44.6	67.0



#24
 Nitrobenzene
 Concen: 4.65 ng
 RT: 7.24 min Scan# 460
 Delta R.T. -0.01 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03

Tgt Ion	Ratio	Lower	Upper
77	100		
123	41.6	30.1	45.1
65	12.3	11.6	17.4





#25

Isophorone

Concen: 4.40 ng

RT: 7.48 min Scan# 481

Delta R.T. -0.05 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS





Tgt Ion: 82 Resp: 136037

Ion Ratio Lower Upper

82 100

95 6.2 5.8 8.8

138 10.9 11.6 17.4#



#26

2-Nitrophenol

Concen: 4.13 ng

RT: 7.56 min Scan# 488

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03




Tgt Ion: 139 Resp: 31828


Ion Ratio Lower Upper

139 100

109 29.7 23.9 35.9

65 56.9 42.7 64.1









#27
 2,4-Dimethylphenol
 Concen: 5.04 ng
 RT: 7.62 min Scan# 493
 Delta R.T. -0.01 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03

Instrument :
 BNA_F
 ClientSampleId :
 PB86810BS

Tgt Ion	Ratio	Lower	Upper
122	100		
107	113.5	101.1	151.7
121	57.3	44.9	67.3




Abundance Ion 122.00 (121.70 to 122.70): E
 Ion 107.00 (106.70 to 107.70): E
 Ion 121.00 (120.70 to 121.70): E






#28
 bis(2-Chloroethoxy)methane
 Concen: 5.10 ng
 RT: 7.70 min Scan# 500
 Delta R.T. -0.02 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03

Tgt Ion	Ratio	Lower	Upper
93	100		
95	32.7	26.5	39.7
123	9.7	8.1	12.1



Abundance Ion 93.00 (92.70 to 93.70): BF0
 Ion 95.00 (94.70 to 95.70): BF0
 Ion 123.00 (122.70 to 123.70): E





#29

2,4-Dichlorophenol

Concen: 5.05 ng

RT: 7.82 min Scan# 510

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS


Tgt Ion:162 Resp: 59036

Ion Ratio Lower Upper

162 100

164 64.4 44.8 84.8




98 43.0 23.1 63.1



Abundance Ion 162.00 (161.70 to 162.70): E

Ion 164.00 (163.70 to 164.70): E

Ion 98.00 (97.70 to 98.70): BF0



#30

1,2,4-Trichlorobenzene

Concen: 4.97 ng

RT: 7.88 min Scan# 516

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03


Tgt Ion:180 Resp: 63905

Ion Ratio Lower Upper

180 100

182 100.4 79.0 118.6



145 27.9 24.2 36.2




Abundance Ion 180.00 (179.70 to 180.70): E

Ion 182.00 (181.70 to 182.70): E

Ion 145.00 (144.70 to 145.70): E









#31
Naphthalene
Concen: 5.05 ng
RT: 7.96 min Scan# 523
Delta R.T. -0.01 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03

Instrument :
BNA_F
ClientSampleId :
PB86810BS

Tgt Ion	Ratio	Lower	Upper
128	100		
129	10.1	9.0	13.4
127	12.6	10.6	15.8




Abundance Ion 128.00 (127.70 to 128.70): E
Ion 129.00 (128.70 to 129.70): E
Ion 127.00 (126.70 to 127.70): E






#32
Benzoic acid
Concen: 3.36 ng
RT: 7.71 min Scan# 501
Delta R.T. -0.10 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03

Tgt Ion	Ratio	Lower	Upper
122	100		
105	130.7	109.7	149.7
77	119.6	79.7	119.7



Abundance Ion 122.00 (121.70 to 122.70): E
Ion 105.00 (104.70 to 105.70): E
Ion 77.00 (76.70 to 77.70): BF0









#33
4-Chloroaniline
Concen: 3.44 ng
RT: 8.03 min Scan# 529
Delta R.T. -0.00 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03




Instrument :
BNA_F
ClientSampleId :
PB86810BS


Tgt Ion:	127	Resp:	53852
Ion Ratio		Lower	Upper
127	100		
129	32.0	25.7	38.5
65	33.1	35.5	53.3#
92	17.4	19.4	29.2#



#34
Hexachlorobutadiene
Concen: 4.86 ng
RT: 8.09 min Scan# 534
Delta R.T. -0.00 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03

Tgt Ion:	225	Resp:	36946
Ion Ratio		Lower	Upper
225	100		
223	59.1	51.3	76.9
227	60.8	49.8	74.8





#35

Caprolactam

Concen: 3.77 ng

RT: 8.39 min Scan# 560

Delta R.T. -0.10 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS


Tgt Ion:113 Resp: 14151

Ion Ratio Lower Upper

113 100

55 209.3 192.6 232.6




56 153.8 148.6 188.6



Abundance Ion 113.00 (112.70 to 113.70): E

Ion 55.00 (54.70 to 55.70): BF0

Ion 56.00 (55.70 to 56.70): BF0



#36

4-Chloro-3-methylphenol

Concen: 4.25 ng

RT: 8.52 min Scan# 572

Delta R.T. -0.02 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03


Tgt Ion:107 Resp: 57630

Ion Ratio Lower Upper

107 100

144 24.9 20.8 31.2



142 72.3 62.1 93.1




Abundance Ion 107.00 (106.70 to 107.70): E

Ion 144.00 (143.70 to 144.70): E

Ion 142.00 (141.70 to 142.70): E





#37

2-Methylnaphthalene

Concen: 5.16 ng

RT: 8.65 min Scan# 583

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS


Tgt Ion:142 Resp: 134816

Ion Ratio Lower Upper

142 100

141 87.3 70.3 105.5




115 37.4 33.4 50.2



Abundance Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E



#38

Acenaphthene-d10

Concen: 20.00 ng

RT: 9.69 min Scan# 674

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03


Tgt Ion:164 Resp: 378911

Ion Ratio Lower Upper

164 100

162 102.1 81.2 121.8



160 44.0 35.6 53.4




Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E









#39
 1,2,4,5-Tetrachlorobenzene
 Concen: 4.95 ng
 RT: 8.82 min Scan# 598
 Delta R.T. -0.01 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03

Instrument :
 BNA_F
 ClientSampleId :
 PB86810BS

Tgt Ion:	216	Resp:	60797
Ion	Ratio	Lower	Upper
216	100		
214	80.0	63.8	95.8
179	20.8	17.0	25.6
108	20.1	15.2	22.8




Abundance Ion 216.00 (215.70 to 216.70): E
 Ion 214.00 (213.70 to 214.70): E
 Ion 179.00 (178.70 to 179.70): E
 Ion 108.00 (107.70 to 108.70): E






#40
 Hexachlorocyclopentadiene
 Concen: 8.18 ng
 RT: 8.81 min Scan# 597
 Delta R.T. -0.01 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03

Tgt Ion:	237	Resp:	57132
Ion	Ratio	Lower	Upper
237	100		
235	62.5	43.5	83.5
272	13.1	0.0	33.1



Abundance Ion 236.80 (236.50 to 237.50): E
 Ion 234.90 (234.60 to 235.60): E
 Ion 271.75 (271.45 to 272.45): E









#41
 2,4,6-Tribromophenol
 Concen: 7.43 ng
 RT: 10.47 min Scan# 742
 Delta R.T. -0.02 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03




Instrument :
 BNA_F
 ClientSampleId :
 PB86810BS


Tgt Ion:330 Resp: 57642
 Ion Ratio Lower Upper
 330 100
 332 95.3 77.7 116.5
 141 43.3 29.7 44.5



#42
 2,4,6-Trichlorophenol
 Concen: 4.24 ng
 RT: 8.95 min Scan# 609
 Delta R.T. -0.01 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03

Tgt Ion:196 Resp: 37279
 Ion Ratio Lower Upper
 196 100
 198 100.8 75.9 113.9
 200 34.0 24.5 36.7









#43
 2,4,5-Trichlorophenol
 Concen: 4.91 ng
 RT: 8.98 min Scan# 612
 Delta R.T. -0.02 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03

Instrument :
 BNA_F
 ClientSampleId :
 PB86810BS

Tgt Ion:196 Resp: 44133
 Ion Ratio Lower Upper
 196 100
 198 93.4 77.6 116.4
 97 69.1 32.6 48.8#
 132 33.0 19.8 29.8#




Abundance Ion 196.00 (195.70 to 196.70): E
 Ion 198.00 (197.70 to 198.70): E
 Ion 97.00 (96.70 to 97.70): BF0
 Ion 132.00 (131.70 to 132.70): E






#44
 2-Fluorobiphenyl
 Concen: 5.83 ng
 RT: 9.02 min Scan# 615
 Delta R.T. -0.01 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03

Tgt Ion:172 Resp: 316637
 Ion Ratio Lower Upper
 172 100
 171 35.6 29.5 44.3
 170 22.6 19.9 29.9



Abundance Ion 172.00 (171.70 to 172.70): E
 Ion 171.00 (170.70 to 171.70): E
 Ion 170.00 (169.70 to 170.70): E





#45

1,1'-Biphenyl

Concen: 5.19 ng

RT: 9.12 min Scan# 624

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS


Tgt Ion:154 Resp: 167247

Ion Ratio Lower Upper

154 100

153 38.7 20.5 60.5




76 11.1 0.0 31.2



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 76.00 (75.70 to 76.70): BF0



#46

2-Chloronaphthalene

Concen: 4.91 ng

RT: 9.14 min Scan# 626

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03


Tgt Ion:162 Resp: 124406

Ion Ratio Lower Upper

162 100

127 34.0 29.4 44.2



164 32.7 26.2 39.4




Abundance Ion 162.00 (161.70 to 162.70): E

Ion 127.00 (126.70 to 127.70): E

Ion 164.00 (163.70 to 164.70): E





#47

2-Nitroaniline

Concen: 4.48 ng

RT: 9.24 min Scan# 635

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS


Tgt Ion: 65 Resp: 40139

Ion Ratio Lower Upper




65 100

92 70.7 49.5 74.3

138 87.2 60.7 91.1



Abundance Ion 65.00 (64.70 to 65.70): BF0
Ion 92.00 (91.70 to 92.70): BF0
Ion 138.00 (137.70 to 138.70): BF0



#48

Acenaphthylene

Concen: 5.08 ng

RT: 9.54 min Scan# 661

Delta R.T. -0.02 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03


Tgt Ion: 152 Resp: 199566

Ion Ratio Lower Upper



152 100


151 18.2 16.3 24.5

153 14.0 10.8 16.2



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E





#49

Dimethylphthalate

Concen: 4.86 ng

RT: 9.43 min Scan# 651

Delta R.T. -0.02 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS


Tgt Ion:163 Resp: 141683

Ion Ratio Lower Upper

163 100

194 3.9 2.6 4.0




164 10.8 8.6 12.8



Abundance Ion 163.00 (162.70 to 163.70): E

Ion 194.00 (193.70 to 194.70): E

Ion 164.00 (163.70 to 164.70): E



#50

2,6-Dinitrotoluene

Concen: 4.38 ng

RT: 9.48 min Scan# 656

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03


Tgt Ion:165 Resp: 28638

Ion Ratio Lower Upper

165 100

63 44.1 62.5 93.7#



89 46.1 59.1 88.7#




Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BF0

Ion 89.00 (88.70 to 89.70): BF0





#51

Acenaphthene

Concen: 5.02 ng

RT: 9.71 min Scan# 676

Delta R.T. -0.02 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS


Tgt Ion:154 Resp: 119887

Ion Ratio Lower Upper

154 100

153 107.3 89.6 134.4




152 53.3 43.3 64.9



Abundance Ion 154.00 (153.70 to 154.70): E

Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E



#52

3-Nitroaniline

Concen: 3.95 ng

RT: 9.66 min Scan# 671

Delta R.T. -0.02 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03


Tgt Ion:138 Resp: 27585

Ion Ratio Lower Upper

138 100

108 10.2 7.8 11.6



92 121.9 81.1 121.7#




Abundance Ion 138.00 (137.70 to 138.70): E

Ion 108.00 (107.70 to 108.70): E

Ion 92.00 (91.70 to 92.70): BF0









#53
 2,4-Dinitrophenol
 Concen: 5.99 ng
 RT: 9.76 min Scan# 680
 Delta R.T. -0.01 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03

Instrument :
 BNA_F
 ClientSampleId :
 PB86810BS

Tgt Ion:184 Resp: 22579
 Ion Ratio Lower Upper
 184 100
 63 70.9 60.5 90.7
 154 62.9 54.3 81.5




Abundance Ion 184.00 (183.70 to 184.70): E
 Ion 63.00 (62.70 to 63.70): BF0
 Ion 154.00 (153.70 to 154.70): E






#54
 Dibenzofuran
 Concen: 5.54 ng
 RT: 9.90 min Scan# 692
 Delta R.T. -0.01 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03

Tgt Ion:168 Resp: 187261
 Ion Ratio Lower Upper
 168 100
 139 34.9 30.2 45.2
 169 12.7 10.9 16.3



Abundance Ion 168.00 (167.70 to 168.70): E
 Ion 139.00 (138.70 to 139.70): E
 Ion 169.00 (168.70 to 169.70): E









#55
4-Nitrophenol
Concen: 7.48 ng
RT: 9.84 min Scan# 687
Delta R.T. -0.02 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03

Instrument :
BNA_F
ClientSampleId :
PB86810BS

Tgt Ion:139 Resp: 40027
Ion Ratio Lower Upper
139 100
109 73.2 64.0 104.0
65 128.8 109.5 149.5






Abundance Ion 139.00 (138.70 to 139.70): E
Ion 109.00 (108.70 to 109.70): E
Ion 65.00 (64.70 to 65.70): BF0




#56
2,4-Dinitrotoluene
Concen: 4.62 ng
RT: 9.88 min Scan# 691
Delta R.T. -0.01 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03

Tgt Ion:165 Resp: 38303
Ion Ratio Lower Upper
165 100
63 57.4 54.5 81.7
89 71.2 77.3 115.9#
182 2.2 1.8 2.6

Abundance Ion 165.00 (164.70 to 165.70): E
Ion 63.00 (62.70 to 63.70): BF0
Ion 89.00 (88.70 to 89.70): BF0
Ion 182.00 (181.70 to 182.70): E









#57
 Fluorene
 Concen: 5.14 ng
 RT: 10.23 min Scan# 721
 Delta R.T. -0.01 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03




Instrument :
 BNA_F
 ClientSampleId :
 PB86810BS


Tgt Ion:166 Resp: 143616
 Ion Ratio Lower Upper
 166 100
 165 100.3 78.0 117.0
 167 13.8 11.3 16.9



#58
 2,3,4,6-Tetrachlorophenol
 Concen: 4.87 ng
 RT: 9.98 min Scan# 699
 Delta R.T. -0.06 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03

Tgt Ion:232 Resp: 35898
 Ion Ratio Lower Upper
 232 100
 131 41.3 38.6 57.8
 130 2.4 1.9 2.9
 166 25.7 25.4 38.0









#59
Diethylphthalate
Concen: 4.67 ng
RT: 10.12 min Scan# 712
Delta R.T. -0.02 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03

Instrument :
BNA_F
ClientSampleId :
PB86810BS

Tgt Ion:149 Resp: 135702
Ion Ratio Lower Upper
149 100
177 24.1 19.4 29.2
150 12.0 10.2 15.2




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 177.00 (176.70 to 177.70): E
Ion 150.00 (149.70 to 150.70): E






#60
4-Chlorophenyl-phenylether
Concen: 5.47 ng
RT: 10.23 min Scan# 721
Delta R.T. -0.01 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03

Tgt Ion:204 Resp: 69922
Ion Ratio Lower Upper
204 100
206 33.8 27.4 41.0
141 67.4 57.4 86.2



Abundance Ion 204.00 (203.70 to 204.70): E
Ion 206.00 (205.70 to 206.70): E
Ion 141.00 (140.70 to 141.70): E









#61
4-Nitroaniline
Concen: 4.18 ng
RT: 10.26 min Scan# 724
Delta R.T. -0.03 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03




Instrument :
BNA_F
ClientSampleId :
PB86810BS


Tgt Ion:138 Resp: 28717
Ion Ratio Lower Upper
138 100
92 50.9 28.4 68.4
108 74.1 53.5 93.5



#62
Azobenzene
Concen: 5.40 ng
RT: 10.39 min Scan# 735
Delta R.T. -0.01 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03

Tgt Ion: 77 Resp: 178933
Ion Ratio Lower Upper
77 100
182 20.5 0.0 38.3
105 21.5 1.2 41.2
51 27.9 11.1 51.1









#63
Phenanthrene-d10
Concen: 20.00 ng
RT: 11.16 min Scan# 803
Delta R.T. -0.01 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03




Instrument :
BNA_F
ClientSampleId :
PB86810BS


Tgt Ion:188 Resp: 739141
Ion Ratio Lower Upper
188 100
94 8.3 6.0 9.0
80 9.3 6.7 10.1



#64
4,6-Dinitro-2-methylphenol
Concen: 3.71 ng
RT: 10.28 min Scan# 726
Delta R.T. -0.02 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03

Tgt Ion:198 Resp: 18910
Ion Ratio Lower Upper
198 100
51 51.9 27.0 67.0
105 49.9 26.1 66.1





#65

n-Nitrosodiphenylamine

Concen: 4.81 ng

RT: 10.35 min Scan# 732

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS


Tgt Ion:169 Resp: 120927

Ion Ratio Lower Upper

169 100

168 65.3 53.8 80.6




167 34.9 29.8 44.8



Abundance Ion 169.00 (168.70 to 169.70): E

Ion 168.00 (167.70 to 168.70): E

Ion 167.00 (166.70 to 167.70): E



#66

4-Bromophenyl-phenylether

Concen: 4.69 ng

RT: 10.72 min Scan# 764

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03


Tgt Ion:248 Resp: 38463

Ion Ratio Lower Upper

248 100

250 100.7 75.4 113.2



141 57.8 48.1 72.1




Abundance Ion 248.00 (247.70 to 248.70): E

Ion 250.00 (249.70 to 250.70): E

Ion 141.00 (140.70 to 141.70): E









#67
Hexachlorobenzene
Concen: 5.24 ng
RT: 10.78 min Scan# 769
Delta R.T. -0.01 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03

Instrument :
BNA_F
ClientSampleId :
PB86810BS

Tgt Ion:284 Resp: 47349
Ion Ratio Lower Upper
284 100
142 28.6 27.6 41.4
249 29.7 25.5 38.3




Abundance Ion 283.80 (283.50 to 284.50): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E






#68
Atrazine
Concen: 5.06 ng
RT: 10.88 min Scan# 778
Delta R.T. -0.03 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03

Tgt Ion:200 Resp: 36919
Ion Ratio Lower Upper
200 100
173 32.4 7.2 47.2
215 43.8 25.7 65.7



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 7.66 ng
 RT: 10.98 min Scan# 787
 Delta R.T. -0.01 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03

Instrument :
 BNA_F
 ClientSampleId :
 PB86810BS

Tgt Ion: 266 Resp: 45045
 Ion Ratio Lower Upper
 266 100
 268 60.5 51.4 77.0
 264 59.6 51.7 77.5




Abundance Ion 265.70 (265.40 to 266.40): E
 Ion 268.00 (267.70 to 268.70): E
 Ion 264.00 (263.70 to 264.70): E






#70
 Phenanthrene
 Concen: 5.33 ng
 RT: 11.19 min Scan# 805
 Delta R.T. -0.01 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03

Tgt Ion: 178 Resp: 227184
 Ion Ratio Lower Upper
 178 100
 176 18.8 16.2 24.2
 179 14.7 12.8 19.2



Abundance Ion 178.00 (177.70 to 178.70): E
 Ion 176.00 (175.70 to 176.70): E
 Ion 179.00 (178.70 to 179.70): E





#71

Anthracene

Concen: 5.33 ng

RT: 11.23 min Scan# 809

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS






Tgt Ion:178 Resp: 231530

Ion Ratio Lower Upper

178 100

176 17.4 15.6 23.4

179 14.7 13.2 19.8



#72

Carbazole

Concen: 4.99 ng

RT: 11.40 min Scan# 824

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03





Tgt Ion:167 Resp: 192821


Ion Ratio Lower Upper

167 100

166 20.9 18.0 27.0

139 10.4 10.5 15.7#





#73

Di-n-butylphthalate

Concen: 4.42 ng

RT: 11.75 min Scan# 854

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS


Tgt Ion:149 Resp: 208426

Ion Ratio Lower Upper

149 100

150 8.6 8.0 12.0




104 4.0 4.4 6.6#



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#74

Fluoranthene

Concen: 5.24 ng

RT: 12.37 min Scan# 908

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03


Tgt Ion:202 Resp: 228240

Ion Ratio Lower Upper

202 100

101 12.4 0.0 32.3



203 18.0 0.0 38.4




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 203.00 (202.70 to 203.70): E





#75

Chrysene-d12

Concen: 20.00 ng

RT: 13.78 min Scan# 1032

Delta R.T. -0.02 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS


Tgt Ion:240 Resp: 657332

Ion Ratio Lower Upper

240 100

120 13.8 5.4 8.2#




236 25.9 20.6 31.0



Abundance Ion 240.00 (239.70 to 240.70): E

Ion 120.00 (119.70 to 120.70): E

Ion 236.00 (235.70 to 236.70): E



#76

Benzidine

Concen: 5.19 ng

RT: 12.50 min Scan# 920

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03


Tgt Ion:184 Resp: 89478

Ion Ratio Lower Upper

184 100

185 14.7 11.7 17.5



183 11.9 8.8 13.2




Abundance Ion 184.00 (183.70 to 184.70): E

Ion 185.00 (184.70 to 185.70): E

Ion 183.00 (182.70 to 183.70): E





#77

Pyrene

Concen: 5.33 ng

RT: 12.59 min Scan# 928

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS


Tgt Ion: 202 Resp: 239277

Ion Ratio Lower Upper

202 100

200 21.0 17.8 26.6




203 16.5 15.0 22.4



Abundance Ion 202.00 (201.70 to 202.70): E

Ion 200.00 (199.70 to 200.70): E

Ion 203.00 (202.70 to 203.70): E



#78

Terphenyl-d14

Concen: 5.78 ng

RT: 12.74 min Scan# 941

Delta R.T. -0.02 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03


Tgt Ion: 244 Resp: 322887

Ion Ratio Lower Upper

244 100

212 7.4 5.4 8.2



122 11.0 4.3 6.5#




Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E









#79
Butylbenzylphthalate
Concen: 4.08 ng
RT: 13.23 min Scan# 984
Delta R.T. -0.01 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03

Instrument :
BNA_F
ClientSampleId :
PB86810BS

Tgt Ion:149 Resp: 86678
Ion Ratio Lower Upper
149 100
91 60.2 61.5 92.3#
206 24.1 15.3 22.9#




Abundance Ion 149.00 (148.70 to 149.70): E
Ion 91.00 (90.70 to 91.70): BF0
Ion 206.00 (205.70 to 206.70): E






#80
Benzo(a)anthracene
Concen: 5.09 ng
RT: 13.77 min Scan# 1031
Delta R.T. -0.02 min
Lab File: BF083134.D
Acq: 22 Nov 2015 3:03

Tgt Ion:228 Resp: 206919
Ion Ratio Lower Upper
228 100
226 26.5 22.6 33.8
229 19.5 16.6 24.8



Abundance Ion 228.00 (227.70 to 228.70): E
Ion 226.00 (225.70 to 226.70): E
Ion 229.00 (228.70 to 229.70): E









#81
 3,3'-Dichlorobenzidine
 Concen: 3.30 ng
 RT: 13.75 min Scan# 1029
 Delta R.T. -0.02 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03




Instrument :
 BNA_F
 ClientSampleId :
 PB86810BS


Tgt Ion:252 Resp: 46732
 Ion Ratio Lower Upper
 252 100
 254 62.6 51.5 77.3
 126 13.7 5.2 7.8#



#82
 Chrysene
 Concen: 5.04 ng
 RT: 13.82 min Scan# 1035
 Delta R.T. -0.01 min
 Lab File: BF083134.D
 Acq: 22 Nov 2015 3:03

Tgt Ion:228 Resp: 190097
 Ion Ratio Lower Upper
 228 100
 226 29.1 25.5 38.3
 229 18.9 16.6 25.0





#83

Bis(2-ethylhexyl)phthalate

Concen: 4.82 ng

RT: 13.79 min Scan# 1033

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS


Tgt Ion:149 Resp: 133758

Ion Ratio Lower Upper

149 100

167 25.3 20.1 30.1




279 2.1 1.7 2.5



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 279.00 (278.70 to 279.70): E



#84

Di-n-octyl phthalate

Concen: 4.52 ng

RT: 14.40 min Scan# 1086

Delta R.T. -0.02 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03


Tgt Ion:149 Resp: 216050

Ion Ratio Lower Upper

149 100

167 1.4 1.2 1.8



43 12.6 9.5 14.3




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 167.00 (166.70 to 167.70): E

Ion 43.00 (42.70 to 43.70): BF0





#85

Indeno(1,2,3-cd)pyrene

Concen: 4.53 ng

RT: 16.41 min Scan# 1262

Delta R.T. -0.03 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS


Tgt Ion:276 Resp: 155208

Ion Ratio Lower Upper

276 100

138 24.9 18.0 27.0




277 24.0 20.2 30.4



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E



#86

Perylene-d12

Concen: 20.00 ng

RT: 15.15 min Scan# 1152

Delta R.T. -0.01 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03


Tgt Ion:264 Resp: 572123

Ion Ratio Lower Upper

264 100

260 24.7 19.1 28.7



265 21.9 18.2 27.4




Abundance Ion 264.00 (263.70 to 264.70): E

Ion 260.00 (259.70 to 260.70): E

Ion 265.00 (264.70 to 265.70): E





#87

Benzo(b)fluoranthene

Concen: 8.95 ng

RT: 14.78 min Scan# 1119

Delta R.T. -0.02 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS


Tgt Ion:252 Resp: 349894

Ion Ratio Lower Upper

252 100

253 21.9 17.7 26.5




125 9.9 5.6 8.4#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#88

Benzo(k)fluoranthene

Concen: 12.01 ng

RT: 14.78 min Scan# 1119

Delta R.T. -0.05 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03


Tgt Ion:252 Resp: 349894

Ion Ratio Lower Upper

252 100

253 21.9 18.2 27.2



125 9.9 6.8 10.2




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#89

Benzo(a)pyrene

Concen: 5.02 ng

RT: 15.10 min Scan# 1147

Delta R.T. -0.02 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS


Tgt Ion: 252 Resp: 162115

Ion Ratio Lower Upper

252 100

253 21.5 17.5 26.3




125 10.1 5.4 8.0#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#90

Dibenzo(a,h)anthracene

Concen: 4.60 ng

RT: 16.42 min Scan# 1263

Delta R.T. -0.03 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03


Tgt Ion: 278 Resp: 134736

Ion Ratio Lower Upper

278 100

139 16.4 10.7 16.1#



279 23.3 19.3 28.9




Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E





#91

Benzo(g,h,i)perylene

Concen: 4.41 ng

RT: 16.79 min Scan# 1295

Delta R.T. -0.03 min

Lab File: BF083134.D

Acq: 22 Nov 2015 3:03

Instrument :

BNA_F

ClientSampleId :

PB86810BS

Tgt Ion: 276 Resp: 126234

Ion Ratio Lower Upper

276 100

277 22.3 18.7 28.1

138 19.1 13.9 20.9

