

Data Path : Z:\HPCHEM1\BNA F\DATA\BF112117\
 Data File : BF100774.D
 Acq On : 21 Nov 2017 12:29
 Operator : SJ/JU
 Sample : PB104316BL
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB104316BL

Quant Time: Nov 22 05:33:38 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 17 15:22:08 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	82752	20.00	ng	-0.01
21) Naphthalene-d8	8.15	136	329209	20.00	ng	-0.02
38) Acenaphthene-d10	9.91	164	158693	20.00	ng	-0.01
63) Phenanthrene-d10	11.39	188	327273	20.00	ng	-0.01
75) Chrysene-d12	14.03	240	274179	20.00	ng	-0.01
86) Perylene-d12	15.50	264	199197	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.50	112	629836	122.01	ng	0.00
7) Phenol-d6	6.50	99	756448	118.83	ng	0.00
23) Nitrobenzene-d5	7.43	82	477508	95.90	ng	-0.02
41) 2,4,6-Tribromophenol	10.70	330	226543	140.09	ng	-0.01
44) 2-Fluorobiphenyl	9.23	172	990498	95.04	ng	-0.01
78) Terphenyl-d14	12.98	244	1081527	90.64	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA F\DATA\BF112117\
Data File : BF100774.D
Acq On : 21 Nov 2017 12:29
Operator : SJ/JU
Sample : PB104316BL
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB104316BL

Quant Time: Nov 22 05:33:38 2017
Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110817.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 17 15:22:08 2017
Response via : Initial Calibration

