


Data Path : Z:\HPCHEM1\BNA F\Data\BF112117\
 Data File : BF100797.D
 Acq On : 21 Nov 2017 23:41
 Operator : SJ/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF110817.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Nov 17 15:22:08 2017



AutoFind: Scans 1623, 1624, 1625; Background Corrected with Scan 1619

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	36.1	14634	PASS
68	69	0.00	2	1.0	130	PASS
69	198	0.00	100	33.7	13633	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	43.0	17395	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	40498	PASS
199	198	5	9	6.8	2766	PASS
275	198	10	60	24.9	10073	PASS
365	198	1	100	2.4	989	PASS
441	443	0.01	100	88.3	5826	PASS
442	198	50	100	92.3	37385	PASS
443	442	15	24	17.6	6595	PASS