

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF112119\
 Data File : BF117917.D
 Acq On : 21 Nov 2019 17:27
 Operator : JU
 Sample : K5904-18RX
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 14-(0-2)RX

Manual Integrations
 APPROVED

mohammad
 11/22/2019 1:21:02 PM

Quant Time: Nov 22 00:42:02 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF111319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 18:36:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.92	152	150563	20.00	ng	-0.01
21) Naphthalene-d8	8.20	136	596793	20.00	ng	-0.01
39) Acenaphthene-d10	9.96	164	307419	20.00	ng	-0.01
64) Phenanthrene-d10	11.46	188	541586	20.00	ng	-0.01
76) Chrysene-d12	14.10	240	386429	20.00	ng	-0.01
87) Perylene-d12	15.62	264	472264	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.52	112	247916	29.15	ng	-0.01
7) Phenol-d6	6.53	99	911745	88.87	ng	-0.01
23) Nitrobenzene-d5	7.47	82	633343	63.09	ng	-0.02
42) 2,4,6-Tribromophenol	10.76	330	43894	13.49	ng	-0.01
45) 2-Fluorobiphenyl	9.28	172	1113602	64.99	ng	-0.01
79) Terphenyl-d14	13.04	244	1269533	60.04	ng	-0.01
Target Compounds						
50) Dimethylphthalate	9.67	163	238419	11.060	ng	98
78) Pyrene	12.90	202	125653	4.024	ng	98
81) Benzo(a)anthracene	14.09	228	55286	2.165	ng	98
83) Chrysene	14.13	228	56781	2.295	ng	98
84) Bis(2-ethylhexyl)phthalate	14.08	149	34720	2.135	ng	# 98
88) Benzo(b)fluoranthene	15.16	252	72847m	2.374	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF112119\
Data File : BF117917.D
Acq On : 21 Nov 2019 17:27
Operator : JU
Sample : K5904-18RX
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
14-(0-2)RX

Quant Time: Nov 22 00:42:02 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF111319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 13 18:36:07 2019
Response via : Initial Calibration

Manual Integrations
APPROVED

mohammad
11/22/2019 1:21:02 PM

