

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF112119\
 Data File : BF117900.D
 Acq On : 21 Nov 2019 9:46
 Operator : JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 21 15:14:43 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF111319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 18:36:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	88	0.00
2	1,4-Dioxane	40.000	45.386	-13.5	102	0.00
3	Pyridine	40.000	33.809	15.5	76	0.00
4	n-Nitrosodimethylamine	40.000	34.753	13.1	79	0.00
5 S	2-Fluorophenol	80.000	77.116	3.6	89	0.00
6	Aniline	40.000	39.813	0.5	89	0.00
7 S	Phenol-d6	80.000	78.891	1.4	91	0.00
8	2-Chlorophenol	40.000	41.008	-2.5	92	0.00
9	Benzaldehyde	40.000	37.641	5.9	84	0.00
10 C	Phenol	40.000	43.329	-8.3	98	0.00
11	bis(2-Chloroethyl)ether	40.000	39.699	0.8	89	0.00
12	1,3-Dichlorobenzene	40.000	40.427	-1.1	91	0.00
13 C	1,4-Dichlorobenzene	40.000	39.959	0.1	89	0.00
14	1,2-Dichlorobenzene	40.000	40.225	-0.6	92	0.00
15	Benzyl Alcohol	40.000	42.885	-7.2	95	0.00
16	2,2'-oxybis(1-Chloropropane)	40.000	37.236	6.9	81	0.00
17	2-Methylphenol	40.000	40.528	-1.3	90	0.00
18	Hexachloroethane	40.000	41.431	-3.6	92	-0.01
19 P	n-Nitroso-di-n-propylamine	40.000	40.421	-1.1	92	0.00
20	3+4-Methylphenols	40.000	39.338	1.7	89	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	90	0.00
22	Acetophenone	40.000	38.751	3.1	91	0.00
23 S	Nitrobenzene-d5	80.000	80.290	-0.4	93	0.00
24	Nitrobenzene	40.000	40.256	-0.6	93	0.00
25	Isophorone	40.000	38.162	4.6	91	0.00
26 C	2-Nitrophenol	40.000	43.074	-7.7	99	0.00
27	2,4-Dimethylphenol	40.000	37.788	5.5	87	0.00
28	bis(2-Chloroethoxy)methane	40.000	38.716	3.2	91	0.00
29 C	2,4-Dichlorophenol	40.000	38.946	2.6	91	0.00
30	1,2,4-Trichlorobenzene	40.000	39.048	2.4	91	0.00
31	Naphthalene	40.000	40.192	-0.5	93	0.00
32	Benzoic acid	40.000	40.495	-1.2	95	0.02
33	4-Chloroaniline	40.000	39.614	1.0	93	0.00
34 C	Hexachlorobutadiene	40.000	40.217	-0.5	93	0.00
35	Caprolactam	40.000	39.755	0.6	94	0.01
36 C	4-Chloro-3-methylphenol	40.000	40.249	-0.6	95	0.00
37	2-Methylnaphthalene	40.000	39.729	0.7	92	0.00
38	1-Methylnaphthalene	40.000	40.029	-0.1	93	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	93	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.849	2.9	93	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.709	8.2	88	-0.01
42 S	2,4,6-Tribromophenol	80.000	81.543	-1.9	99	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.716	3.2	92	0.00
44	2,4,5-Trichlorophenol	40.000	38.736	3.2	92	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF112119\
 Data File : BF117900.D
 Acq On : 21 Nov 2019 9:46
 Operator : JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 21 15:14:43 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF111319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 18:36:07 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
45 S	2-Fluorobiphenyl	80.000	82.575	-3.2	93	0.00
46	1,1'-Biphenyl	40.000	39.389	1.5	93	0.00
47	2-Chloronaphthalene	40.000	38.965	2.6	93	0.00
48	2-Nitroaniline	40.000	40.435	-1.1	96	0.00
49	Acenaphthylene	40.000	39.794	0.5	94	0.00
50	Dimethylphthalate	40.000	39.711	0.7	96	0.00
51	2,6-Dinitrotoluene	40.000	39.925	0.2	95	0.00
52 C	Acenaphthene	40.000	39.125	2.2	94	0.00
53	3-Nitroaniline	40.000	40.472	-1.2	96	0.00
54 P	2,4-Dinitrophenol	40.000	42.067	-5.2	110	0.00
55	Dibenzofuran	40.000	39.042	2.4	93	0.00
56 P	4-Nitrophenol	40.000	40.383	-1.0	96	0.01
57	2,4-Dinitrotoluene	40.000	42.370	-5.9	102	0.00
58	Fluorene	40.000	39.189	2.0	97	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	38.339	4.2	92	0.00
60	Diethylphthalate	40.000	40.230	-0.6	96	0.00
61	4-Chlorophenyl-phenylether	40.000	39.686	0.8	95	0.00
62	4-Nitroaniline	40.000	42.008	-5.0	105	0.00
63	Azobenzene	40.000	38.905	2.7	93	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	96	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.763	-11.9	112	0.00
66 c	n-Nitrosodiphenylamine	40.000	35.869	10.3	89	0.00
67	4-Bromophenyl-phenylether	40.000	39.111	2.2	97	0.00
68	Hexachlorobenzene	40.000	37.538	6.2	92	0.00
69	Atrazine	40.000	39.357	1.6	98	0.00
70 C	Pentachlorophenol	40.000	37.266	6.8	92	0.00
71	Phenanthrene	40.000	37.232	6.9	95	0.00
72	Anthracene	40.000	37.870	5.3	97	0.00
73	Carbazole	40.000	38.762	3.1	96	0.00
74	Di-n-butylphthalate	40.000	39.705	0.7	102	0.00
75 C	Fluoranthene	40.000	42.468	-6.2	101	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	106	0.00
77	Benzidine	40.000	34.824	12.9	92	0.00
78	Pyrene	40.000	37.203	7.0	102	0.00
79 S	Terphenyl-d14	80.000	75.618	5.5	104	0.00
80	Butylbenzylphthalate	40.000	40.516	-1.3	113	0.00
81	Benzo(a)anthracene	40.000	39.068	2.3	108	0.00
82	3,3'-Dichlorobenzidine	40.000	39.919	0.2	109	0.00
83	Chrysene	40.000	38.444	3.9	109	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	42.752	-6.9	122	0.00
85 c	Di-n-octyl phthalate	40.000	43.615	-9.0	125	0.00
86	Indeno(1,2,3-cd)pyrene	40.000	36.662	8.3	113	0.01
87 I	Perylene-d12	20.000	20.000	0.0	115	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF112119\
Data File : BF117900.D
Acq On : 21 Nov 2019 9:46
Operator : JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 21 15:14:43 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF111319.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 13 18:36:07 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	40.000	39.392	1.5	114	0.00
89	Benzo(k)fluoranthene	40.000	36.933	7.7	112	0.00
90 C	Benzo(a)pyrene	40.000	38.063	4.8	113	0.00
91	Dibenzo(a,h)anthracene	40.000	37.346	6.6	115	0.01
92	Benzo(q,h,i)perylene	40.000	36.262	9.3	112	0.02

(#) = Out of Range

SPCC's out = 0 CCC's out = 0