

Data Path : Z:\SVOASRV\HPCHEM1\BNA F\DATA\BF112119\
 Data File : BF117923.D
 Acq On : 21 Nov 2019 20:10
 Operator : JU
 Sample : K5676-21
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SU-02-112019

Manual Integrations
 APPROVED

mohammad
 11/22/2019 1:21:12 PM


Quant Time: Nov 22 01:00:55 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA F\METHODS\8270-BF111319.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 18:36:07 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.92	152	136575	20.00	ng	-0.01
21) Naphthalene-d8	8.20	136	559294	20.00	ng	-0.01
39) Acenaphthene-d10	9.97	164	276199	20.00	ng	0.00
64) Phenanthrene-d10	11.46	188	504781	20.00	ng	0.00
76) Chrysene-d12	14.12	240	358326	20.00	ng	0.00
87) Perylene-d12	15.63	264	395768	20.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
5) 2-Fluorophenol	5.53	112	883157	114.46	ng	0.00
7) Phenol-d6	6.54	99	1151099	123.69	ng	0.00
23) Nitrobenzene-d5	7.48	82	708982	75.36	ng	-0.01
42) 2,4,6-Tribromophenol	10.76	330	310006	106.02	ng	-0.01
45) 2-Fluorobiphenyl	9.28	172	1373543	89.22	ng	-0.01
79) Terphenyl-d14	13.05	244	1181740	60.27	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
10) Phenol	6.55	94	23833	2.464	ng	94
31) Naphthalene	8.23	128	280314	10.751	ng	97
37) 2-Methylnaphthalene	8.92	142	152665	8.666	ng	98
38) 1-Methylnaphthalene	9.02	142	135239	8.120	ng	99
46) 1,1'-Biphenyl	9.38	154	41175	2.009	ng	99
49) Acenaphthylene	9.83	152	144036	5.861	ng	98
50) Dimethylphthalate	9.67	163	250210	12.919	ng	98
52) Acenaphthene	10.00	154	166942	10.603	ng	99
55) Dibenzofuran	10.17	168	204774	9.393	ng	98
58) Fluorene	10.52	166	346262	20.495	ng	99
71) Phenanthrene	11.49	178	2096175	82.596	ng	99
72) Anthracene	11.54	178	626495	24.898	ng	98
73) Carbazole	11.69	167	150979	6.866	ng	97
75) Fluoranthene	12.69	202	2287374	109.813	ng	98
78) Pyrene	12.92	202	2012530	69.498	ng	99
81) Benzo(a)anthracene	14.10	228	1437343	60.702	ng	96
83) Chrysene	14.14	228	1184128	51.608	ng	98
86) Indeno(1,2,3-cd)pyrene	17.17	276	523186m	26.791	ng	
88) Benzo(b)fluoranthene	15.19	252	1726542m	67.146	ng	
89) Benzo(k)fluoranthene	15.21	252	471092m	19.444	ng	
90) Benzo(a)pyrene	15.57	252	1188813	52.578	ng	98
91) Dibenz(a,h)anthracene	17.17	278	135507	6.963	ng	96
92) Benzo(g,h,i)perylene	17.63	276	474710	24.490	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed




#1
 1,4-Dichlorobenzene-d4
 Concen: 20.000 ng
 RT: 6.92 min Scan# 821
 Delta R.T. -0.01 min
 Lab File: BF117923.D
 Acq: 21 Nov 2019 20:10

Instrument :
 BNA_F
 ClientSampleId :
 SU-02-112019

Tot Ion	Ratio	Lower	Upper
152	100		
150	158.7	125.3	187.9
115	62.4	47.0	70.4

Manual Integrations
 APPROVED

mohammad
 11/22/2019 1:21:12 PM






Abundance

Ion 152.00 (151.70 to 152.70): E


Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#5
 2-Fluorophenol
 Concen: 114.463 ng
 RT: 5.53 min Scan# 585
 Delta R.T. -0.01 min
 Lab File: BF117923.D
 Acq: 21 Nov 2019 20:10

Tgt Ion	Ratio	Lower	Upper
112	100		
64	56.9	45.6	68.4
63	31.1	24.1	36.1






Abundance

Ion 112.00 (111.70 to 112.70): E

Ion 64.00 (63.70 to 64.70): BF1

Ion 63.00 (62.70 to 63.70): BF1





#7

Phenol-d6

Concen: 123.690 ng

RT: 6.54 min Scan# 757

Delta R.T. -0.01 min

Lab File: BF117923.D

Acq: 21 Nov 2019 20:10

Instrument :

BNA_F

ClientSampleId :

SU-02-112019

Tot Ion: 99 Resp: 1151099

Ion Ratio Lower Upper

99 100

42 17.6 14.9 22.3


71 36.2 28.6 43.0

Manual Integrations




APPROVED

mohammad

11/22/2019 1:21:12 PM



Abundance Ion 99.00 (98.70 to 99.70): BF1
Ion 42.00 (41.70 to 42.70): BF1
Ion 71.00 (70.70 to 71.70): BF1



#10

Phenol

Concen: 2.464 ng

RT: 6.55 min Scan# 759

Delta R.T. -0.01 min

Lab File: BF117923.D

Acq: 21 Nov 2019 20:10


Tgt Ion: 94 Resp: 23833

Ion Ratio Lower Upper



94 100

65 26.3 8.9 48.9

66 35.0 18.7 58.7



Abundance Ion 94.00 (93.70 to 94.70): BF1
Ion 65.00 (64.70 to 65.70): BF1
Ion 66.00 (65.70 to 66.70): BF1









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.20 min Scan# 1040
Delta R.T. -0.01 min
Lab File: BF117923.D
Acq: 21 Nov 2019 20:10

Instrument :
BNA_F
ClientSampleId :
SU-02-112019

Tot Ion	Ratio	Resp	Lower	Upper
136	100			
137	11.1	9.0	13.4	
54	6.9	5.4	8.2	
68	5.1	4.3	6.5	




Manual Integrations
APPROVED


mohammad
11/22/2019 1:21:12 PM



#23
Nitrobenzene-d5
Concen: 75.356 ng
RT: 7.48 min Scan# 917
Delta R.T. -0.01 min
Lab File: BF117923.D
Acq: 21 Nov 2019 20:10

Tgt Ion	Ratio	Resp	Lower	Upper
82	100			
128	50.4	38.9	58.3	
54	48.0	39.8	59.6	






#31
Naphthalene
Concen: 10.751 ng
RT: 8.23 min Scan# 1044
Delta R.T. -0.01 min
Lab File: BF117923.D
Acq: 21 Nov 2019 20:10

Instrument :
BNA_F
ClientSampleId :
SU-02-112019

Tot Ion	Ratio	Lower	Upper
128	100		
129	10.8	9.5	14.3
127	13.1	11.3	16.9

Manual Integrations
APPROVED

mohammad
11/22/2019 1:21:12 PM






Abundance

Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#37
2-Methylnaphthalene
Concen: 8.666 ng
RT: 8.92 min Scan# 1162
Delta R.T. -0.01 min
Lab File: BF117923.D
Acq: 21 Nov 2019 20:10




Tgt Ion	Ratio	Lower	Upper
142	100		
141	90.1	71.1	106.7
115	33.9	29.2	43.8


Abundance

Ion 142.00 (141.70 to 142.70): E

Ion 141.00 (140.70 to 141.70): E

Ion 115.00 (114.70 to 115.70): E





#38

1-Methylnaphthalene

Concen: 8.120 ng

RT: 9.02 min Scan# 1178

Delta R.T. -0.01 min

Lab File: BF117923.D

Acq: 21 Nov 2019 20:10

Instrument :

BNA_F


ClientSampleId :

SU-02-112019




Tot Ion:	142	Resp:	135239
Ion Ratio	Lower	Upper	
142	100		
141	92.8	73.5	110.3
116	3.8	3.2	4.8

Manual Integrations
APPROVED

mohammad
11/22/2019 1:21:12 PM



Abundance Ion 142.00 (141.70 to 142.70): E
Ion 141.00 (140.70 to 141.70): E
Ion 116.00 (115.70 to 116.70): E



#39

Acenaphthene-d10

Concen: 20.000 ng


RT: 9.97 min Scan# 1340

Delta R.T. -0.01 min



Lab File: BF117923.D


Acq: 21 Nov 2019 20:10

Tgt Ion:	164	Resp:	276199
Ion Ratio	Lower	Upper	
164	100		
162	101.1	81.3	121.9
160	44.9	36.6	55.0



Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E





#42

2,4,6-Tribromophenol

Concen: 106.018 ng

RT: 10.76 min Scan# 1474

Delta R.T. -0.01 min

Lab File: BF117923.D

Acq: 21 Nov 2019 20:10

Instrument :

BNA_F


ClientSampleId :

SU-02-112019




Tot Ion:	330	Resp:	310006
Ion Ratio	Lower	Upper	
330	100		
332	97.9	76.7	115.1
141	35.4	27.7	41.5

Manual Integrations
APPROVED

mohammad
11/22/2019 1:21:12 PM



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E



#45

2-Fluorobiphenyl

Concen: 89.218 ng


RT: 9.28 min Scan# 1223

Delta R.T. -0.01 min



Lab File: BF117923.D


Acq: 21 Nov 2019 20:10

Tgt Ion:	172	Resp:	1373543
Ion Ratio	Lower	Upper	
172	100		
171	36.8	30.9	46.3
170	24.4	21.0	31.4



Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#46
1,1'-Biphenyl
Concen: 2.009 ng
RT: 9.38 min Scan# 1240
Delta R.T. -0.01 min
Lab File: BF117923.D
Acq: 21 Nov 2019 20:10

Instrument :
BNA_F
ClientSampleId :
SU-02-112019




Tot Ion:	154	Resp:	41175
Ion Ratio	Lower	Upper	
154	100		
153	41.9	21.9	61.9
76	12.5	0.0	33.6

Manual Integrations
APPROVED

mohammad
11/22/2019 1:21:12 PM




Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 76.00 (75.70 to 76.70): BF1






#49
Acenaphthylene
Concen: 5.861 ng
RT: 9.83 min Scan# 1316
Delta R.T. -0.01 min
Lab File: BF117923.D
Acq: 21 Nov 2019 20:10

Tgt Ion:	152	Resp:	144036
Ion Ratio	Lower	Upper	
152	100		
151	19.9	17.0	25.6
153	14.5	11.0	16.6



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E









#50
Dimethylphthalate
Concen: 12.919 ng
RT: 9.67 min Scan# 1290
Delta R.T. -0.02 min
Lab File: BF117923.D
Acq: 21 Nov 2019 20:10

Instrument :
BNA_F
ClientSampleId :
SU-02-112019

Tot Ion	Ion	Ratio	Resp	Lower	Upper
163	100		250210		
194	3.8		3.3	4.9	
164	9.9		8.7	13.1	




Manual Integrations
APPROVED


mohammad
11/22/2019 1:21:12 PM



#52
Acenaphthene
Concen: 10.603 ng
RT: 10.00 min Scan# 1345
Delta R.T. -0.01 min
Lab File: BF117923.D
Acq: 21 Nov 2019 20:10

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
154	100		166942		
153	112.5		90.1	135.1	
152	53.5		44.0	66.0	









#55
Dibenzenofuran
Concen: 9.393 ng
RT: 10.17 min Scan# 1374
Delta R.T. -0.01 min
Lab File: BF117923.D
Acq: 21 Nov 2019 20:10

Instrument :
BNA_F
ClientSampleId :
SU-02-112019

Tot Ion	Ratio	Lower	Upper
168	100		
139	40.4	33.0	49.6
169	16.2	11.5	17.3




Manual Integrations
APPROVED


mohammad
11/22/2019 1:21:12 PM



#58
Fluorene
Concen: 20.495 ng
RT: 10.52 min Scan# 1433
Delta R.T. -0.01 min
Lab File: BF117923.D
Acq: 21 Nov 2019 20:10

Tgt Ion	Ratio	Lower	Upper
166	100		
165	99.3	78.4	117.6
167	14.1	11.0	16.4





#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.46 min Scan# 1594

Delta R.T. -0.01 min

Lab File: BF117923.D

Acq: 21 Nov 2019 20:10

Instrument :

BNA_F


ClientSampleId :

SU-02-112019

Tot Ion:	188	Resp:	504781
Ion Ratio	Lower	Upper	
188	100		
94	8.0	6.6	10.0
80	8.4	7.2	10.8

Manual Integrations
APPROVED




mohammad
11/22/2019 1:21:12 PM



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BF1

Ion 80.00 (79.70 to 80.70): BF1



#71

Phenanthrene

Concen: 82.596 ng


RT: 11.49 min Scan# 1599

Delta R.T. -0.00 min

Lab File: BF117923.D

Acq: 21 Nov 2019 20:10



Tgt Ion:	178	Resp:	2096175
Ion Ratio	Lower	Upper	
178	100		
176	21.3	16.6	25.0
179	16.7	13.2	19.8




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#72

Anthracene

Concen: 24.898 ng

RT: 11.54 min Scan# 1607

Delta R.T. -0.01 min

Lab File: BF117923.D

Acq: 21 Nov 2019 20:10

Instrument :

BNA_F





ClientSampleId :

SU-02-112019

Tot Ion:	178	Resp:	626495
Ion Ratio	Lower	Upper	
178	100		
176	19.0	16.1	24.1
179	15.8	13.0	19.4

Manual Integrations
APPROVED

mohammad
11/22/2019 1:21:12 PM



#73

Carbazole

Concen: 6.866 ng




RT: 11.69 min Scan# 1632


Delta R.T. -0.01 min

Lab File: BF117923.D

Acq: 21 Nov 2019 20:10

Tgt Ion:	167	Resp:	150979
Ion Ratio	Lower	Upper	
167	100		
166	21.2	18.0	27.0
139	13.7	11.7	17.5





#75

Fluoranthene

Concen: 109.813 ng

RT: 12.69 min Scan# 1802

Delta R.T. 0.00 min

Lab File: BF117923.D

Acq: 21 Nov 2019 20:10

Instrument :

BNA_F





ClientSampleId :

SU-02-112019

Tot Ion:	202	Resp:	2287374
Ion	Ratio	Lower	Upper
202	100		
101	11.9	0.0	31.1
203	19.1	0.0	38.2

Manual Integrations
APPROVED

mohammad
11/22/2019 1:21:12 PM



#76

Chrysene-d12

Concen: 20.000 ng




RT: 14.12 min Scan# 2045


Delta R.T. -0.00 min

Lab File: BF117923.D

Acq: 21 Nov 2019 20:10

Tgt Ion:	240	Resp:	358326
Ion	Ratio	Lower	Upper
240	100		
120	12.0	9.5	14.3
236	27.3	21.3	31.9





#78

Pvrene

Concen: 69.498 ng

RT: 12.92 min Scan# 1842

Delta R.T. 0.01 min

Lab File: BF117923.D

Acq: 21 Nov 2019 20:10

Instrument :

BNA_F





ClientSampleId :

SU-02-112019

Tot Ion:	202	Resp:	2012530
Ion	Ratio	Lower	Upper
202	100		
200	22.7	17.8	26.8
203	18.6	14.6	22.0

Manual Integrations
APPROVED

mohammad
11/22/2019 1:21:12 PM



#79

Terphenyl-d14

Concen: 60.274 ng




RT: 13.05 min Scan# 1864


Delta R.T. -0.01 min

Lab File: BF117923.D

Acq: 21 Nov 2019 20:10

Tgt Ion:	244	Resp:	1181740
Ion	Ratio	Lower	Upper
244	100		
212	7.2	6.3	9.5
122	11.0	9.4	14.2









#81
Benzo(a)anthracene
Concen: 60.702 ng
RT: 14.10 min Scan# 2043
Delta R.T. -0.00 min
Lab File: BF117923.D
Acq: 21 Nov 2019 20:10

Instrument :
BNA_F
ClientSampleId :
SU-02-112019

Tot Ion	Ratio	Lower	Upper
228	100		
226	30.6	22.2	33.2
229	21.3	15.9	23.9




Manual Integrations
APPROVED


mohammad
11/22/2019 1:21:12 PM



#83
Chrysene
Concen: 51.608 ng
RT: 14.14 min Scan# 2050
Delta R.T. -0.00 min
Lab File: BF117923.D
Acq: 21 Nov 2019 20:10

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.2	25.0	37.4
229	22.2	16.6	25.0









#86
Indeno(1,2,3-cd)pyrene
Concen: 26.791 ng/ml
RT: 17.17 min Scan# 2564
Delta R.T. 0.01 min
Lab File: BF117923.D
Acq: 21 Nov 2019 20:10

Instrument :
BNA_F
ClientSampleId :
SU-02-112019

Tot Ion	Ratio	Lower	Upper
276	100		
138	4.5	20.9	31.3#
277	5.3	20.3	30.5#




Manual Integrations
APPROVED


mohammad
11/22/2019 1:21:12 PM



#87
Perylene-d12
Concen: 20.000 ng
RT: 15.63 min Scan# 2302
Delta R.T. 0.01 min
Lab File: BF117923.D
Acq: 21 Nov 2019 20:10

Tgt Ion	Ratio	Lower	Upper
264	100		
260	24.5	19.4	29.0
265	24.9	17.1	25.7









#88
Benzo(b)fluoranthene
Concen: 67.146 ng m
RT: 15.19 min Scan# 2227
Delta R.T. 0.01 min
Lab File: BF117923.D
Acq: 21 Nov 2019 20:10

Instrument :
BNA_F
ClientSampleId :
SU-02-112019

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.9	18.0	27.0
125	11.1	8.6	13.0




Manual Integrations
APPROVED


mohammad
11/22/2019 1:21:12 PM



#89
Benzo(k)fluoranthene
Concen: 19.444 ng m
RT: 15.21 min Scan# 2231
Delta R.T. -0.00 min
Lab File: BF117923.D
Acq: 21 Nov 2019 20:10

Tgt Ion	Ratio	Lower	Upper
252	100		
253	24.6	17.5	26.3
125	10.6	7.9	11.9









#90
Benzo(a)pyrene
Concen: 52.578 ng
RT: 15.57 min Scan# 2292
Delta R.T. 0.01 min
Lab File: BF117923.D
Acq: 21 Nov 2019 20:10

Instrument :
BNA_F
ClientSampleId :
SU-02-112019

Tot Ion	Ratio	Lower	Upper
252	100		
253	23.5	17.8	26.6
125	11.4	9.2	13.8




Manual Integrations
APPROVED




mohammad
11/22/2019 1:21:12 PM



#91
Dibenzo(a,h)anthracene
Concen: 6.963 ng
RT: 17.17 min Scan# 2565
Delta R.T. -0.00 min
Lab File: BF117923.D
Acq: 21 Nov 2019 20:10

Tgt Ion	Ratio	Lower	Upper
278	100		
139	18.6	13.8	20.8
279	27.2	19.6	29.4





#92
 Benzo(a,h,i)perylene
 Concen: 24.490 ng
 RT: 17.63 min Scan# 2643
 Delta R.T. 0.02 min
 Lab File: BF117923.D
 Acq: 21 Nov 2019 20:10

Instrument :
 BNA_F
 ClientSampleId :
 SU-02-112019

Tot Ion	Ratio	Lower	Upper
276	100		
277	24.7	18.9	28.3
138	23.0	17.4	26.2

Manual Integrations
 APPROVED

mohammad
 11/22/2019 1:21:12 PM

