

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112122\
 Data File : BF131309.D
 Acq On : 21 Nov 2022 11:20
 Operator : CG\JU
 Sample : SSTDICC2.5
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC2.5

Quant Time: Nov 22 00:38:56 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 22 00:36:43 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.963	152	119292	20.000	ng	0.00
21) Naphthalene-d8	8.257	136	432438	20.000	ng	0.00
39) Acenaphthene-d10	10.022	164	257841	20.000	ng	0.00
64) Phenanthrene-d10	11.516	188	487007	20.000	ng	0.00
76) Chrysene-d12	14.174	240	394310	20.000	ng	# 0.00
86) Perylene-d12	15.721	264	314275	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	0.000	112	0d	0.000	ng	
7) Phenol-d6	0.000	99	0d	0.000	ng	
23) Nitrobenzene-d5	0.000	82	0d	0.000	ng	
42) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
45) 2-Fluorobiphenyl	0.000	172	0d	0.000	ng	
79) Terphenyl-d14	0.000	244	0d	0.000	ng	
Target Compounds						
19) n-Nitroso-di-n-propyla...	7.363	70	14673	2.469	ng	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112122\
Data File : BF131309.D
Acq On : 21 Nov 2022 11:20
Operator : CG\JU
Sample : SSTDICC2.5
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
SSTDICC2.5

Quant Time: Nov 22 00:38:56 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 22 00:36:43 2022
Response via : Initial Calibration

