



Data Path : Z:\HPCHEM1\BNA F\Data\BF112217\
 Data File : BF100854.D
 Acq On : 23 Nov 2017 5:41
 Operator : SJ/JU
 Sample : DF7TPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 26 23:59:36 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 17 15:22:08 2017
 Response via : Initial Calibration



TIC: BF100854.D

(76) Benzidine



12.716min (-0.024) 5605.20ng

response 144787

Ion	Exp%	Act%
184.00	100	100
185.00	15.70	15.24
183.00	11.60	11.82
0.00	0.00	0.00

Data Path : Z:\HPCHEM1\BNA F\Data\BF112217\
 Data File : BF100854.D
 Acq On : 23 Nov 2017 5:41
 Operator : SJ/JU
 Sample : DF7TPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 26 23:59:36 2017
 Quant Method : Z:\HPCHEM1\BNA F\METHODS\8270-BF110817.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 17 15:22:08 2017
 Response via : Initial Calibration



TIC: BF100854.D

(69) Pentachlorophenol (C)

11.186min (-0.024) 0.00ng

response 22742

Ion	Exp%	Act%
265.70	100	100
268.00	63.90	60.81
264.00	61.90	61.57
0.00	0.00	0.00