

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112222\
 Data File : BF131330.D
 Acq On : 22 Nov 2022 15:04
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 23 03:24:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 22 01:17:52 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
2	1,4-Dioxane	0.501	0.468	6.6	95	-0.02
3	Pyridine	1.391	1.319	5.2	95	-0.01
4	n-Nitrosodimethylamine	0.820	0.766	6.6	94	0.00
5 S	2-Fluorophenol	1.051	1.034	1.6	99	0.00
6	Aniline	1.757	1.635	6.9	94	0.00
7 S	Phenol-d6	1.341	1.307	2.5	99	0.00
8	2-Chlorophenol	1.191	1.178	1.1	99	0.00
9	Benzaldehyde	0.942	0.817	13.3	98	0.00
10 C	Phenol	1.487	1.473	0.9	100	0.00
11	bis(2-Chloroethyl)ether	1.245	1.176	5.5	96	0.00
12	1,3-Dichlorobenzene	1.419	1.338	5.7	98	0.00
13 C	1,4-Dichlorobenzene	1.442	1.349	6.4	96	0.00
14	1,2-Dichlorobenzene	1.328	1.246	6.2	99	0.00
15	Benzyl Alcohol	1.099	1.110	-1.0	96	0.00
16	2,2'-oxybis(1-Chloropropane	2.382	2.194	7.9	93	0.00
17	2-Methylphenol	1.015	1.012	0.3	99	0.00
18	Hexachloroethane	0.516	0.509	1.4	98	0.00
19 P	n-Nitroso-di-n-propylamine	0.982	0.937	4.6	98	0.00
20	3+4-Methylphenols	1.296	1.294	0.2	99	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
22	Acetophenone	0.509	0.469	7.9	97	0.00
23 S	Nitrobenzene-d5	0.397	0.371	6.5	96	0.00
24	Nitrobenzene	0.413	0.387	6.3	97	0.00
25	Isophorone	0.708	0.666	5.9	97	0.00
26 C	2-Nitrophenol	0.111	0.149	-34.2#	116	0.00
27	2,4-Dimethylphenol	0.274	0.273	0.4	101	0.00
28	bis(2-Chloroethoxy)methane	0.402	0.370	8.0	96	0.00
29 C	2,4-Dichlorophenol	0.298	0.293	1.7	98	0.00
30	1,2,4-Trichlorobenzene	0.369	0.336	8.9	95	0.00
31	Naphthalene	1.066	0.978	8.3	98	0.00
32	Benzoic acid	0.156	0.189	-21.2	133	0.00
33	4-Chloroaniline	0.420	0.391	6.9	99	0.00
34 C	Hexachlorobutadiene	0.239	0.222	7.1	97	0.00
35	Caprolactam	0.078	0.090	-15.4	112	0.00
36 C	4-Chloro-3-methylphenol	0.309	0.316	-2.3	103	0.00
37	2-Methylnaphthalene	0.722	0.667	7.6	98	0.00
38	1-Methylnaphthalene	0.700	0.646	7.7	98	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	100	0.00
40	1,2,4,5-Tetrachlorobenzene	0.649	0.586	9.7	98	0.00
41 P	Hexachlorocyclopentadiene	0.292	0.284	2.7	100	0.00
42 S	2,4,6-Tribromophenol	0.168	0.187	-11.3	112	0.00
43 C	2,4,6-Trichlorophenol	0.368	0.386	-4.9	107	0.00
44	2,4,5-Trichlorophenol	0.414	0.428	-3.4	106	0.00
45 S	2-Fluorobiphenyl	1.375	1.220	11.3	99	0.00
46	1,1'-Biphenyl	1.526	1.401	8.2	101	0.00
47	2-Chloronaphthalene	1.224	1.110	9.3	99	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112222\
 Data File : BF131330.D
 Acq On : 22 Nov 2022 15:04
 Operator : CG\JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_F
 LabSampleId :
 SSTDCCC040

Quant Time: Nov 23 03:24:55 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 22 01:17:52 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.372	0.396	-6.5	108	0.00
49	Acenaphthylene	1.866	1.715	8.1	100	0.00
50	Dimethylphthalate	1.393	1.364	2.1	106	0.00
51	2,6-Dinitrotoluene	0.279	0.287	-2.9	106	0.00
52 C	Acenaphthene	1.169	1.097	6.2	103	0.00
53	3-Nitroaniline	0.286	0.297	-3.8	109	0.00
54 P	2,4-Dinitrophenol	0.094	0.112	-19.1	130	0.00
55	Dibenzofuran	1.681	1.562	7.1	103	0.00
56 P	4-Nitrophenol	0.199	0.224	-12.6	113	0.00
57	2,4-Dinitrotoluene	0.351	0.372	-6.0	109	0.00
58	Fluorene	1.312	1.227	6.5	105	0.00
59	2,3,4,6-Tetrachlorophenol	0.345	0.372	-7.8	110	0.00
60	Diethylphthalate	1.316	1.329	-1.0	108	0.00
61	4-Chlorophenyl-phenylether	0.673	0.622	7.6	102	0.00
62	4-Nitroaniline	0.285	0.291	-2.1	110	0.00
63	Azobenzene	1.415	1.324	6.4	102	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
65	4,6-Dinitro-2-methylphenol	0.079	0.089	-12.7	128	0.00
66 c	n-Nitrosodiphenylamine	0.637	0.584	8.3	105	0.00
67	4-Bromophenyl-phenylether	0.245	0.223	9.0	104	0.00
68	Hexachlorobenzene	0.260	0.237	8.8	105	0.00
69	Atrazine	0.196	0.197	-0.5	109	0.00
70 C	Pentachlorophenol	0.133	0.144	-8.3	118	0.00
71	Phenanthrene	1.081	0.987	8.7	107	0.00
72	Anthracene	1.062	0.969	8.8	107	0.00
73	Carbazole	0.906	0.832	8.2	106	0.00
74	Di-n-butylphthalate	1.068	1.082	-1.3	110	0.00
75 C	Fluoranthene	1.139	1.037	9.0	106	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	107	0.00
77	Benzydine	0.369	0.388	-5.1	114	0.00
78	Pyrene	1.765	1.707	3.3	106	0.00
79 S	Terphenyl-d14	1.223	1.166	4.7	107	0.00
80	Butylbenzylphthalate	0.508	0.615	-21.1	124	0.00
81	Benzo(a)anthracene	1.371	1.264	7.8	106	0.00
82	3,3'-Dichlorobenzidine	0.364	0.373	-2.5	113	0.00
83	Chrysene	1.345	1.225	8.9	106	0.00
84	Bis(2-ethylhexyl)phthalate	0.686	0.765	-11.5	114	0.00
85 c	Di-n-octyl phthalate	1.035	1.105	-6.8	117	0.00
86 I	Perylene-d12	1.000	1.000	0.0	101	0.00
87	Indeno(1,2,3-cd)pyrene	1.343	1.342	0.1	101	0.00
88	Benzo(b)fluoranthene	1.329	1.250	5.9	103	0.00
89	Benzo(k)fluoranthene	1.365	1.292	5.3	101	0.00
90 C	Benzo(a)pyrene	1.090	1.023	6.1	99	0.00
91	Dibenzo(a,h)anthracene	1.109	1.106	0.3	100	0.00
92	Benzo(g,h,i)perylene	1.107	1.119	-1.1	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112222\
Data File : BF131330.D
Acq On : 22 Nov 2022 15:04
Operator : CG\JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_F
LabSampleId :
SSTDCCC040

Quant Time: Nov 23 03:24:55 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112122.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 22 01:17:52 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1