

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112224\
 Data File : BF140565.D
 Acq On : 22 Nov 2024 09:46
 Operator : RC/JU
 Sample : PB165129BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 PB165129BL

Quant Time: Nov 22 10:52:21 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 21 15:23:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.869	152	101281	20.000	ng	0.00
21) Naphthalene-d8	8.151	136	382231	20.000	ng	0.00
39) Acenaphthene-d10	9.904	164	221417	20.000	ng	-0.01
64) Phenanthrene-d10	11.398	188	431432	20.000	ng	0.00
76) Chrysene-d12	14.045	240	281525	20.000	ng	0.00
86) Perylene-d12	15.545	264	210949	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.510	112	850387	143.259	ng	0.01
7) Phenol-d6	6.510	99	1102646	140.508	ng	0.00
23) Nitrobenzene-d5	7.434	82	734960	98.353	ng	0.00
42) 2,4,6-Tribromophenol	10.704	330	349202	147.463	ng	0.00
45) 2-Fluorobiphenyl	9.228	172	1432294	96.381	ng	0.00
79) Terphenyl-d14	12.992	244	1684688	93.181	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF112224\
Data File : BF140565.D
Acq On : 22 Nov 2024 09:46
Operator : RC/JU
Sample : PB165129BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_F
ClientSampleId :
PB165129BL

Quant Time: Nov 22 10:52:21 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF112124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 21 15:23:48 2024
Response via : Initial Calibration

