


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF112316\
 Data File : BF091266.D
 Acq On : 23 Nov 2016 9:02
 Operator : UM/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF112316.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Nov 23 13:22:53 2016



AutoFind: Scans 872, 873, 874; Background Corrected with Scan 869

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	45.4	25282	PASS
68	69	0.00	2	1.9	370	PASS
69	198	0.00	100	34.7	19311	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	47.9	26682	PASS
197	198	0.00	2	0.9	476	PASS
198	198	100	100	100.0	55684	PASS
199	198	5	9	6.6	3696	PASS
275	198	10	60	24.2	13486	PASS
365	198	1	100	3.1	1745	PASS
441	442	0.01	24	15.9	7946	PASS
442	198	50	100	89.9	50045	PASS
443	442	15	24	19.0	9526	PASS