


Data Path : Z:\HPCHEM1\BNA_F\DATA\BF112415\
 Data File : BF083241.D
 Acq On : 24 Nov 2015 17:17
 Operator : UM/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_F\METHODS\8270-BF112115.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Sun Nov 22 09:22:46 2015



AutoFind: Scans 824, 825, 826; Background Corrected with Scan 820

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	44.1	72663	PASS
68	69	0.00	2	1.4	965	PASS
69	198	0.00	100	41.0	67447	PASS
70	69	0.00	2	0.5	343	PASS
127	198	10	80	46.6	76714	PASS
197	198	0.00	2	0.3	425	PASS
198	198	100	100	100.0	164679	PASS
199	198	5	9	6.3	10332	PASS
275	198	10	60	24.4	40147	PASS
365	198	1	100	2.5	4087	PASS
441	442	0.01	24	15.9	22004	PASS
442	198	50	100	83.8	138013	PASS
443	442	15	24	19.8	27314	PASS