

Data Path : Z:\HPCHEM1\BNA F\DATA\BF112515\
 Data File : BF083297.D
 Acq On : 26 Nov 2015 2:43
 Operator : UM/IZ
 Sample : G4426-08RE
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SB-8RE

Manual Integrations
 APPROVED

mmdadoda
 11/26/2015 6:45:21 PM

Quant Time: Nov 26 03:34:14 2015
 Quant Method : Z:\HPCHEM1\BNA F\Methods\8270-BF112115.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Nov 22 09:22:46 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.59	152	177522	20.00	ng	-0.07
21) Naphthalene-d8	7.88	136	620368	20.00	ng	-0.07
38) Acenaphthene-d10	9.63	164	259344	20.00	ng	-0.07
63) Phenanthrene-d10	11.11	188	453879	20.00	ng	-0.07
75) Chrysene-d12	13.75	240	368676	20.00	ng	-0.06
86) Perylene-d12	15.10	264	318837	20.00	ng	-0.07
System Monitoring Compounds						
5) 2-Fluorophenol	5.20	112	1560257	132.88	ng	-0.08
7) Phenol-d6	6.27	99	2014904	132.62	ng	-0.07
23) Nitrobenzene-d5	7.16	82	1233359	90.84	ng	-0.07
41) 2,4,6-Tribromophenol	10.42	330	304302	114.69	ng	-0.07
44) 2-Fluorobiphenyl	8.96	172	1371283	73.73	ng	-0.07
78) Terphenyl-d14	12.70	244	1014034	64.77	ng	-0.07
Target Compounds						Qvalue
10) Phenol	6.28	94	37868	2.06	ng	95
31) Naphthalene	7.91	128	622462	18.59	ng	99
37) 2-Methylnaphthalene	8.59	142	179900	8.28	ng	96
45) 1,1'-Biphenyl	9.06	154	56106	2.54	ng	97
48) Acenaphthylene	9.48	152	84023	3.13	ng	# 96
49) Dimethylphthalate	9.36	163	464232	23.25	ng	100
51) Acenaphthene	9.66	154	368539	22.54	ng	98
54) Dibenzofuran	9.84	168	392690	16.98	ng	92
57) Fluorene	10.17	166	465157	24.34	ng	100
70) Phenanthrene	11.14	178	3850203	147.24	ng	95
71) Anthracene	11.19	178	1334198	50.01	ng	100
72) Carbazole	11.35	167	638266	26.91	ng	99
74) Fluoranthene	12.33	202	4487067	167.66	ng	94
77) Pyrene	12.56	202	4162652	165.27	ng	95
80) Benzo(a)anthracene	13.74	228	2437295	106.89	ng	97
82) Chrysene	13.77	228	1908935m	90.28	ng	
85) Indeno(1,2,3-cd)pyrene	16.33	276	782695	40.70	ng	97
87) Benzo(b)fluoranthene	14.74	252	1693523m	77.74	ng	
88) Benzo(k)fluoranthene	14.74	252	948389m	58.43	ng	
89) Benzo(a)pyrene	15.05	252	1395237	77.57	ng	# 94
90) Dibenzo(a,h)anthracene	16.33	278	193742m	11.86	ng	
91) Benzo(g,h,i)perylene	16.70	276	635509	39.85	ng	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\HPCHEM1\BNA F\DATA\BF112515\
Data File : BF083297.D
Acq On : 26 Nov 2015 2:43
Operator : UM/IZ
Sample : G4426-08RE
Misc :
ALS Vial : 24 Sample Multiplier: 1


Instrument :
BNA_F
ClientSampleId :
SB-8RE

Manual Integrations
APPROVED

mmdadoda
11/26/2015 6:45:21 PM

Quant Time: Nov 26 03:34:14 2015
Quant Method : Z:\HPCHEM1\BNA F\Methods\8270-BF112115.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Nov 22 09:22:46 2015
Response via : Initial Calibration





#1





1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 6.59 min Scan# 403
Delta R.T. -0.07 min
Lab File: BF083297.D
Acq: 26 Nov 2015 2:43

Instrument :
BNA_F
ClientSampleId :
SB-8RE

Tot Ion	Ratio	Lower	Upper
152	100		
150	154.4	125.8	188.8
115	61.1	46.9	70.3

Manual Integrations
APPROVED




mmdadoda
11/26/2015 6:45:21 PM




#5

2-Fluorophenol
Concen: 132.88 ng
RT: 5.20 min Scan# 281
Delta R.T. -0.08 min
Lab File: BF083297.D
Acq: 26 Nov 2015 2:43

Tgt Ion	Ratio	Lower	Upper
112	100		
64	69.9	58.4	87.6
63	35.9	30.0	45.0





#7

Phenol-d6

Concen: 132.62 ng

RT: 6.27 min Scan# 375

Delta R.T. -0.07 min

Lab File: BF083297.D

Acq: 26 Nov 2015 2:43

Instrument :

BNA_F

ClientSampleId :

SB-8RE

Tot Ion: 99 Resp: 2014904

Ion Ratio Lower Upper


99 100

42 22.3 17.7 26.5




71 36.0 30.2 45.4

Manual Integrations
APPROVED

mmdadoda
11/26/2015 6:45:21 PM



Abundance Ion 99.00 (98.70 to 99.70): BF0
Ion 42.00 (41.70 to 42.70): BF0
Ion 71.00 (70.70 to 71.70): BF0



#10

Phenol

Concen: 2.06 ng

RT: 6.28 min Scan# 376

Delta R.T. -0.07 min

Lab File: BF083297.D

Acq: 26 Nov 2015 2:43


Tgt Ion: 94 Resp: 37868

Ion Ratio Lower Upper



94 100


65 38.7 21.1 61.1

66 67.4 43.1 83.1



Abundance Ion 94.00 (93.70 to 94.70): BF0
Ion 65.00 (64.70 to 65.70): BF0
Ion 66.00 (65.70 to 66.70): BF0









#21
Naphthalene-d8
Concen: 20.00 ng
RT: 7.88 min Scan# 516
Delta R.T. -0.07 min
Lab File: BF083297.D
Acq: 26 Nov 2015 2:43

Instrument :
BNA_F
ClientSampleId :
SB-8RE

Tot Ion	Ion	Ratio	Resp	Lower	Upper
136	100		620368		
137	10.8		8.7	13.1	
54	7.9		5.8	8.6	
68	5.8		3.8	5.8#	




Manual Integrations
APPROVED


mmdadoda
11/26/2015 6:45:21 PM



#23
Nitrobenzene-d5
Concen: 90.84 ng
RT: 7.16 min Scan# 453
Delta R.T. -0.07 min
Lab File: BF083297.D
Acq: 26 Nov 2015 2:43

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
82	100		1233359		
128	37.0		28.5	42.7	
54	55.4		44.6	67.0	





#31

Naphthalene

Concen: 18.59 ng

RT: 7.91 min Scan# 518

Delta R.T. -0.07 min

Lab File: BF083297.D

Acq: 26 Nov 2015 2:43

Instrument :

BNA_F





ClientSampleId :

SB-8RE

Tot Ion:	128	Resp:	622462
Ion Ratio	Lower	Upper	
128	100		
129	11.1	9.0	13.4
127	12.9	10.6	15.8

Manual Integrations
APPROVED

mmdadoda
11/26/2015 6:45:21 PM



#37

2-Methylnaphthalene

Concen: 8.28 ng




RT: 8.59 min Scan# 578


Delta R.T. -0.07 min

Lab File: BF083297.D

Acq: 26 Nov 2015 2:43

Tgt Ion:	142	Resp:	179900
Ion Ratio	Lower	Upper	
142	100		
141	88.8	70.3	105.5
115	34.9	33.4	50.2






#38
Acenaphthene-d10
Concen: 20.00 ng
RT: 9.63 min Scan# 669
Delta R.T. -0.07 min
Lab File: BF083297.D
Acq: 26 Nov 2015 2:43

Instrument :
BNA_F
ClientSampleId :
SB-8RE




Tot Ion	Ratio	Resp	Lower	Upper
164	100			
162	103.7	81.2	121.8	
160	45.6	35.6	53.4	

Manual Integrations
APPROVED

mmdadoda
11/26/2015 6:45:21 PM




Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E






#41
2,4,6-Tribromophenol
Concen: 114.69 ng
RT: 10.42 min Scan# 738
Delta R.T. -0.07 min
Lab File: BF083297.D
Acq: 26 Nov 2015 2:43

Tgt Ion	Ratio	Resp	Lower	Upper
330	100			
332	97.4	77.7	116.5	
141	41.6	29.7	44.5	



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E






#44
2-Fluorobiphenyl
Concen: 73.73 ng
RT: 8.96 min Scan# 610
Delta R.T. -0.07 min
Lab File: BF083297.D
Acq: 26 Nov 2015 2:43

Instrument :
BNA_F
ClientSampleId :
SB-8RE




Tot Ion	Ratio	Lower	Upper
172	100		
171	36.0	29.5	44.3
170	24.4	19.9	29.9

Manual Integrations
APPROVED

mmdadoda
11/26/2015 6:45:21 PM




Abundance Ion 172.00 (171.70 to 172.70): E
Ion 171.00 (170.70 to 171.70): E
Ion 170.00 (169.70 to 170.70): E






#45
1,1'-Biphenyl
Concen: 2.54 ng
RT: 9.06 min Scan# 619
Delta R.T. -0.07 min
Lab File: BF083297.D
Acq: 26 Nov 2015 2:43

Tgt Ion	Ratio	Lower	Upper
154	100		
153	38.5	20.5	60.5
76	12.0	0.0	31.2



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 76.00 (75.70 to 76.70): BF0





#48

Acenaphthylene

Concen: 3.13 ng

RT: 9.48 min Scan# 656

Delta R.T. -0.08 min

Lab File: BF083297.D

Acq: 26 Nov 2015 2:43

Instrument :

BNA_F


ClientSampleId :

SB-8RE




Tot Ion:	152	Resp:	84023
Ion Ratio	Lower	Upper	
152	100		
151	19.8	16.3	24.5
153	17.0	10.8	16.2#

Manual Integrations
APPROVED

mmdadoda
11/26/2015 6:45:21 PM



Abundance Ion 152.00 (151.70 to 152.70): E
Ion 151.00 (150.70 to 151.70): E
Ion 153.00 (152.70 to 153.70): E



#49

Dimethylphthalate

Concen: 23.25 ng


RT: 9.36 min Scan# 645

Delta R.T. -0.09 min



Lab File: BF083297.D


Acq: 26 Nov 2015 2:43

Tgt Ion:	163	Resp:	464232
Ion Ratio	Lower	Upper	
163	100		
194	3.0	2.6	4.0
164	10.7	8.6	12.8



Abundance Ion 163.00 (162.70 to 163.70): E
Ion 194.00 (193.70 to 194.70): E
Ion 164.00 (163.70 to 164.70): E





#51

Acenaphthene

Concen: 22.54 ng

RT: 9.66 min Scan# 671

Delta R.T. -0.08 min

Lab File: BF083297.D

Acq: 26 Nov 2015 2:43

Instrument :

BNA_F


ClientSampleId :

SB-8RE




Tot Ion:	154	Resp:	368539
Ion Ratio	Lower	Upper	
154	100		
153	111.1	89.6	134.4
152	52.0	43.3	64.9

Manual Integrations
APPROVED

mmdadoda
11/26/2015 6:45:21 PM



Abundance Ion 154.00 (153.70 to 154.70): E
Ion 153.00 (152.70 to 153.70): E
Ion 152.00 (151.70 to 152.70): E



#54

Dibenzofuran

Concen: 16.98 ng


RT: 9.84 min Scan# 687

Delta R.T. -0.07 min



Lab File: BF083297.D


Acq: 26 Nov 2015 2:43

Tgt Ion:	168	Resp:	392690
Ion Ratio	Lower	Upper	
168	100		
139	31.2	30.2	45.2
169	14.3	10.9	16.3



Abundance Ion 168.00 (167.70 to 168.70): E
Ion 139.00 (138.70 to 139.70): E
Ion 169.00 (168.70 to 169.70): E









#57
Fluorene
 Concen: 24.34 ng
 RT: 10.17 min Scan# 716
 Delta R.T. -0.07 min
 Lab File: BF083297.D
 Acq: 26 Nov 2015 2:43

Instrument :
 BNA_F
 ClientSampleID :
 SB-8RE

Tot Ion	Ion	Ratio	Resp	Lower	Upper
166	100		465157		
165	97.6	78.0	117.0		
167	15.1	11.3	16.9		




Manual Integrations
 APPROVED


mmdadoda
 11/26/2015 6:45:21 PM



#63
Phenanthrene-d10
 Concen: 20.00 ng
 RT: 11.11 min Scan# 798
 Delta R.T. -0.07 min
 Lab File: BF083297.D
 Acq: 26 Nov 2015 2:43

Tgt Ion	Ion	Ratio	Resp	Lower	Upper
188	100		453879		
94	10.2	6.0	9.0#		
80	11.5	6.7	10.1#		





#70

Phenanthrene

Concen: 147.24 ng

RT: 11.14 min Scan# 801

Delta R.T. -0.06 min

Lab File: BF083297.D

Acq: 26 Nov 2015 2:43

Instrument :

BNA_F

ClientSampleId :

SB-8RE

Tot Ion:178 Resp: 3850203

Ion Ratio Lower Upper


178 100

176 22.7 16.2 24.2

179 18.0 12.8 19.2

Manual Integrations
APPROVED




mmdadoda
11/26/2015 6:45:21 PM



Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E



#71

Anthracene

Concen: 50.01 ng

RT: 11.19 min Scan# 805

Delta R.T. -0.06 min

Lab File: BF083297.D

Acq: 26 Nov 2015 2:43


Tgt Ion:178 Resp: 1334198

Ion Ratio Lower Upper

178 100

176 19.6 15.6 23.4



179 16.2 13.2 19.8




Abundance Ion 178.00 (177.70 to 178.70): E

Ion 176.00 (175.70 to 176.70): E

Ion 179.00 (178.70 to 179.70): E





#72

Carbazole

Concen: 26.91 ng

RT: 11.35 min Scan# 819

Delta R.T. -0.07 min

Lab File: BF083297.D

Acq: 26 Nov 2015 2:43

Instrument :

BNA_F





ClientSampleId :

SB-8RE

Tot Ion:	167	Resp:	638266
Ion Ratio	Lower	Upper	
167	100		
166	22.4	18.0	27.0
139	12.1	10.5	15.7

Manual Integrations
APPROVED

mmdadoda
11/26/2015 6:45:21 PM



#74

Fluoranthene

Concen: 167.66 ng




RT: 12.33 min Scan# 905


Delta R.T. -0.05 min

Lab File: BF083297.D

Acq: 26 Nov 2015 2:43

Tgt Ion:	202	Resp:	4487067
Ion Ratio	Lower	Upper	
202	100		
101	15.4	0.0	32.3
203	20.5	0.0	38.4









#75
Chrysene-d12
Concen: 20.00 ng
RT: 13.75 min Scan# 1029
Delta R.T. -0.06 min
Lab File: BF083297.D
Acq: 26 Nov 2015 2:43

Instrument :
BNA_F
ClientSampleId :
SB-8RE

Tot Ion	Ratio	Lower	Upper
240	100		
120	7.8	5.4	8.2
236	26.2	20.6	31.0




Manual Integrations
APPROVED


mmdadoda
11/26/2015 6:45:21 PM



#77
Pyrene
Concen: 165.27 ng
RT: 12.56 min Scan# 925
Delta R.T. -0.05 min
Lab File: BF083297.D
Acq: 26 Nov 2015 2:43

Tgt Ion	Ratio	Lower	Upper
202	100		
200	24.2	17.8	26.6
203	21.1	15.0	22.4





#78

Terphenyl-d14

Concen: 64.77 ng

RT: 12.70 min Scan# 937

Delta R.T. -0.07 min

Lab File: BF083297.D

Acq: 26 Nov 2015 2:43

Instrument :

BNA_F

ClientSampleID :

SB-8RE

Tot Ion:244 Resp: 1014034

Ion Ratio Lower Upper

244 100


212 7.3 5.4 8.2

122 9.6 4.3 6.5#

Manual Integrations
APPROVED

mmdadoda




11/26/2015 6:45:21 PM



Abundance Ion 244.00 (243.70 to 244.70): E

Ion 212.00 (211.70 to 212.70): E

Ion 122.00 (121.70 to 122.70): E



#80

Benzo(a)anthracene

Concen: 106.89 ng

RT: 13.74 min Scan# 1028

Delta R.T. -0.06 min

Lab File: BF083297.D

Acq: 26 Nov 2015 2:43


Tgt Ion:228 Resp: 2437295

Ion Ratio Lower Upper

228 100

226 29.9 22.6 33.8



229 22.1 16.6 24.8




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E









#82
Chrysene
Concen: 90.28 ng/ml
RT: 13.77 min Scan# 1031
Delta R.T. -0.06 min
Lab File: BF083297.D
Acq: 26 Nov 2015 2:43

Instrument :
BNA_F
ClientSampleId :
SB-8RE

Tot Ion	Ratio	Lower	Upper
228	100		
226	31.1	25.5	38.3
229	24.0	16.6	25.0




Manual Integrations
APPROVED


mmdadoda
11/26/2015 6:45:21 PM



#85
Indeno(1,2,3-cd)pyrene
Concen: 40.70 ng/ml
RT: 16.33 min Scan# 1255
Delta R.T. -0.11 min
Lab File: BF083297.D
Acq: 26 Nov 2015 2:43

Tgt Ion	Ratio	Lower	Upper
276	100		
138	22.8	18.0	27.0
277	27.7	20.2	30.4





#86

Pervlene-d12

Concen: 20.00 ng

RT: 15.10 min Scan# 1147

Delta R.T. -0.07 min

Lab File: BF083297.D

Acq: 26 Nov 2015 2:43

Instrument :

BNA_F





ClientSampleId :

SB-8RE

Tot Ion:	264	Resp:	318837
Ion Ratio		Lower	Upper
264	100		
260	24.3	19.1	28.7
265	26.5	18.2	27.4

Manual Integrations
APPROVED

mmdadoda
11/26/2015 6:45:21 PM



#87

Benzo(b)fluoranthene

Concen: 77.74 ng m




RT: 14.74 min Scan# 1116


Delta R.T. -0.06 min

Lab File: BF083297.D

Acq: 26 Nov 2015 2:43

Tgt Ion:	252	Resp:	1693523
Ion Ratio		Lower	Upper
252	100		
253	24.0	17.7	26.5
125	12.2	5.6	8.4#





#88

Benzo(k)fluoranthene

Concen: 58.43 ng m

RT: 14.74 min Scan# 1116

Delta R.T. -0.08 min

Lab File: BF083297.D

Acq: 26 Nov 2015 2:43

Instrument :

BNA_F





ClientSampleId :

SB-8RE

Tot Ion	Ratio	Lower	Upper
252	100		
253	24.0	18.2	27.2
125	12.2	6.8	10.2

Manual Integrations
APPROVED

mmdadoda
11/26/2015 6:45:21 PM



#89

Benzo(a)pyrene

Concen: 77.57 ng




RT: 15.05 min Scan# 1143


Delta R.T. -0.07 min

Lab File: BF083297.D

Acq: 26 Nov 2015 2:43

Tgt Ion	Ratio	Lower	Upper
252	100		
253	23.9	17.5	26.3
125	10.9	5.4	8.0









#90
Dibenzo(a,h)anthracene
Concen: 11.86 ng m
RT: 16.33 min Scan# 1255
Delta R.T. -0.13 min
Lab File: BF083297.D
Acq: 26 Nov 2015 2:43

Instrument :
BNA_F
ClientSampleId :
SB-8RE

Tot Ion	Ratio	Lower	Upper
278	100		
139	17.3	10.7	16.1#
279	29.4	19.3	28.9#

Manual Integrations
APPROVED

mmdadoda
11/26/2015 6:45:21 PM



#91
Benzo(g,h,i)perylene
Concen: 39.85 ng
RT: 16.70 min Scan# 1287
Delta R.T. -0.13 min
Lab File: BF083297.D
Acq: 26 Nov 2015 2:43

Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.9	18.7	28.1
138	23.4	13.9	20.9#

